

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN032119\
 Data File : BN004843.D
 Acq On : 21 Mar 2019 11:08
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 22 02:09:02 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 20 16:06:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	131	0.00
2	1,4-Dioxane	0.455	0.375	17.6	107	0.00
3	n-Nitrosodimethylamine	0.269	0.236	12.3	125	0.00
4 S	2-Fluorophenol	1.087	1.020	6.2	121	0.00
5 S	Phenol-d6	1.354	1.276	5.8	124	0.00
6	bis(2-Chloroethyl)ether	1.221	1.198	1.9	132	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	135	-0.01
8 S	Nitrobenzene-d5	0.265	0.236	10.9	121	-0.01
9	Naphthalene	1.122	1.122	0.0	133	-0.01
10	Hexachlorobutadiene	0.214	0.214	0.0	134	-0.01
11 SURR	2-Methylnaphthalene-d10	0.698	0.682	2.3	132	0.00
12	2-Methylnaphthalene	0.833	0.819	1.7	131	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	131	-0.01
14 S	2,4,6-Tribromophenol	0.276	0.223	19.2	112	0.00
15 S	2-Fluorobiphenyl	1.653	1.667	-0.8	129	0.00
16	Acenaphthylene	2.060	1.934	6.1	129	-0.01
17	Acenaphthene	1.349	1.348	0.1	129	0.00
18	Fluorene	1.883	1.872	0.6	127	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	129	0.00
20	4-Bromophenyl-phenylether	0.261	0.252	3.4	126	0.00
21	Hexachlorobenzene	0.299	0.299	0.0	128	0.00
22	Pentachlorophenol	0.058	0.065	-12.1	166#	0.00
23	Phenanthrene	1.292	1.293	-0.1	128	0.00
24	Anthracene	1.154	1.102	4.5	126	0.00
25 SURR	Fluoranthene-d10	1.339	1.319	1.5	128	0.00
26	Fluoranthene	1.660	1.661	-0.1	128	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	135	0.00
28	Pvrene	1.220	1.174	3.8	128	0.00
29 S	Terphenyl-d14	0.819	0.776	5.3	127	0.00
30	Benzo(a)anthracene	1.312	1.254	4.4	131	-0.01
31	Chrysene	1.382	1.398	-1.2	133	-0.01
32	Bis(2-ethylhexyl)phthalate	0.524	0.515	1.7	144	0.00
33	Indeno(1,2,3-cd)pyrene	1.999	2.075	-3.8	136	0.00
34 I	Perylene-d12	1.000	1.000	0.0	140	0.00
35	Benzo(b)fluoranthene	1.447	1.450	-0.2	138	0.00
36	Benzo(k)fluoranthene	1.424	1.397	1.9	139	0.00
37 C	Benzo(a)pyrene	1.327	1.308	1.4	139	0.00
38	Dibenzo(a,h)anthracene	1.537	1.503	2.2	135	0.00
39	Benzo(a,h,i)perylene	1.566	1.534	2.0	135	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN032119\
Data File : BN004843.D
Acq On : 21 Mar 2019 11:08
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 22 02:09:02 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 20 16:06:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			