

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN032119\
 Data File : BN004853.D
 Acq On : 21 Mar 2019 17:01
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 22 02:30:32 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 20 16:06:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	138	0.00
2	1,4-Dioxane	0.400	0.387	3.3	132	0.00
3	n-Nitrosodimethylamine	0.400	0.350	12.5	130	0.00
4 S	2-Fluorophenol	0.400	0.385	3.8	131	0.00
5 S	Phenol-d6	0.400	0.395	1.3	138	0.00
6	bis(2-Chloroethyl)ether	0.400	0.396	1.0	140	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	145	-0.01
8 S	Nitrobenzene-d5	0.400	0.375	6.3	137	0.00
9	Naphthalene	0.400	0.402	-0.5	144	-0.01
10	Hexachlorobutadiene	0.400	0.397	0.8	142	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.399	0.3	145	0.00
12	2-Methylnaphthalene	0.400	0.402	-0.5	144	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	143	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.345	13.8	132	0.00
15 S	2-Fluorobiphenyl	0.400	0.407	-1.7	144	0.00
16	Acenaphthylene	0.400	0.395	1.3	149	-0.01
17	Acenaphthene	0.400	0.407	-1.7	145	0.00
18	Fluorene	0.400	0.400	0.0	141	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	140	0.00
20	4-Bromophenyl-phenylether	0.400	0.392	2.0	138	0.00
21	Hexachlorobenzene	0.400	0.405	-1.3	140	0.00
22	Pentachlorophenol	0.400	0.204	49.0#	83	0.01
23	Phenanthrene	0.400	0.401	-0.3	139	0.00
24	Anthracene	0.400	0.389	2.8	139	0.00
25 SURR	Fluoranthene-d10	0.400	0.383	4.3	135	0.00
26	Fluoranthene	0.400	0.391	2.3	136	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	135	0.00
28	Pvrene	0.400	0.404	-1.0	134	0.00
29 S	Terphenyl-d14	0.400	0.395	1.3	131	0.00
30	Benzo(a)anthracene	0.400	0.390	2.5	133	0.00
31	Chrysene	0.400	0.408	-2.0	133	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.380	5.0	139	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.389	2.8	127	0.00
34 I	Perylene-d12	0.400	0.400	0.0	135	0.00
35	Benzo(b)fluoranthene	0.400	0.404	-1.0	134	0.00
36	Benzo(k)fluoranthene	0.400	0.402	-0.5	138	0.00
37 C	Benzo(a)pyrene	0.400	0.400	0.0	136	0.00
38	Dibenzo(a,h)anthracene	0.400	0.377	5.8	126	0.00
39	Benzo(a,h,i)perylene	0.400	0.369	7.8	123	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN032119\
Data File : BN004853.D
Acq On : 21 Mar 2019 17:01
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 22 02:30:32 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 20 16:06:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				