

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN032119\
 Data File : BN004872.D
 Acq On : 22 Mar 2019 04:13
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDC00.4

Quant Time: Mar 22 04:58:16 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 20 16:06:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	150	0.00
2	1,4-Dioxane	0.455	0.407	10.5	133	0.00
3	n-Nitrosodimethylamine	0.269	0.241	10.4	145	0.00
4 S	2-Fluorophenol	1.087	1.063	2.2	144	0.00
5 S	Phenol-d6	1.354	1.406	-3.8	157#	0.00
6	bis(2-Chloroethyl)ether	1.221	1.239	-1.5	156#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	164#	-0.01
8 S	Nitrobenzene-d5	0.265	0.246	7.2	154#	-0.01
9	Naphthalene	1.122	1.133	-1.0	163#	-0.01
10	Hexachlorobutadiene	0.214	0.212	0.9	161#	-0.01
11 SURR	2-Methylnaphthalene-d10	0.698	0.702	-0.6	165#	0.00
12	2-Methylnaphthalene	0.833	0.849	-1.9	165#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	161#	-0.01
14 S	2,4,6-Tribromophenol	0.276	0.234	15.2	145	0.00
15 S	2-Fluorobiphenyl	1.653	1.698	-2.7	162#	0.00
16	Acenaphthylene	2.060	2.088	-1.4	171#	-0.01
17	Acenaphthene	1.349	1.379	-2.2	163#	0.00
18	Fluorene	1.883	1.871	0.6	157#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	152#	0.00
20	4-Bromophenyl-phenylether	0.261	0.259	0.8	151#	0.00
21	Hexachlorobenzene	0.299	0.299	0.0	150	0.00
22	Pentachlorophenol	0.058	0.031	46.6#	94	0.04
23	Phenanthrene	1.292	1.318	-2.0	153#	0.00
24	Anthracene	1.154	1.152	0.2	155#	0.00
25 SURR	Fluoranthene-d10	1.339	1.290	3.7	147	0.00
26	Fluoranthene	1.660	1.637	1.4	149	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	142	0.00
28	Pvrene	1.220	1.278	-4.8	147	0.00
29 S	Terphenyl-d14	0.819	0.825	-0.7	142	0.00
30	Benzo(a)anthracene	1.312	1.347	-2.7	148	-0.01
31	Chrysene	1.382	1.448	-4.8	145	-0.01
32	Bis(2-ethylhexyl)phthalate	0.524	0.589	-12.4	174#	0.00
33	Indeno(1,2,3-cd)pyrene	1.999	1.883	5.8	130	0.00
34 I	Perylene-d12	1.000	1.000	0.0	144	0.00
35	Benzo(b)fluoranthene	1.447	1.469	-1.5	144	0.00
36	Benzo(k)fluoranthene	1.424	1.437	-0.9	147	0.00
37 C	Benzo(a)pyrene	1.327	1.336	-0.7	146	0.00
38	Dibenzo(a,h)anthracene	1.537	1.393	9.4	129	0.00
39	Benzo(a,h,i)perylene	1.566	1.369	12.6	124	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN032119\
Data File : BN004872.D
Acq On : 22 Mar 2019 04:13
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 22 04:58:16 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 20 16:06:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			