

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN032122\  
 Data File : BN019044.D  
 Acq On : 22 Mar 2022 05:18  
 Operator : CG/JU  
 Sample : N2000-22  
 Misc :  
 ALS Vial : 26 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 DUP05-20220315

Quant Time: Mar 22 07:27:32 2022  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\8270-SIM-BN031722.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Mar 17 13:18:23 2022  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc  | Units | Dev(Min) |
|-------------------------------|--------|------|----------|-------|-------|----------|
| Internal Standards            |        |      |          |       |       |          |
| 1) 1,4-Dichlorobenzene-d4     | 7.854  | 152  | 5843     | 0.400 | ng    | 0.00     |
| 7) Naphthalene-d8             | 10.658 | 136  | 16009    | 0.400 | ng    | # 0.00   |
| 13) Acenaphthene-d10          | 14.485 | 164  | 9148     | 0.400 | ng    | -0.01    |
| 19) Phenanthrene-d10          | 17.236 | 188  | 13744    | 0.400 | ng    | # 0.00   |
| 29) Chrysene-d12              | 21.422 | 240  | 8053     | 0.400 | ng    | # 0.00   |
| 35) Perylene-d12              | 23.785 | 264  | 7389     | 0.400 | ng    | # 0.00   |
| System Monitoring Compounds   |        |      |          |       |       |          |
| 4) 2-Fluorophenol             | 5.406  | 112  | 1571     | 0.120 | ng    | 0.00     |
| 5) Phenol-d6                  | 7.024  | 99   | 731      | 0.053 | ng    | 0.01     |
| 8) Nitrobenzene-d5            | 9.014  | 82   | 2405     | 0.208 | ng    | 0.00     |
| 11) 2-Methylnaphthalene-d10   | 12.253 | 152  | 6361     | 0.240 | ng    | 0.00     |
| 14) 2,4,6-Tribromophenol      | 15.985 | 330  | 714      | 0.162 | ng    | 0.00     |
| 15) 2-Fluorobiphenyl          | 13.116 | 172  | 9271     | 0.253 | ng    | -0.01    |
| 27) Fluoranthene-d10          | 19.261 | 212  | 11749    | 0.288 | ng    | 0.00     |
| 31) Terphenyl-d14             | 19.860 | 244  | 8642     | 0.437 | ng    | 0.00     |
| Target Compounds              |        |      |          |       |       |          |
| 2) 1,4-Dioxane                | 3.247  | 88   | 27514    | 3.810 | ng    | 98       |
| 34) Bis(2-ethylhexyl)phtha... | 21.324 | 149  | 1007     | 0.061 | ng    | # 98     |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN032122\  
Data File : BN019044.D  
Acq On : 22 Mar 2022 05:18  
Operator : CG/JU  
Sample : N2000-22  
Misc :  
ALS Vial : 26 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
DUP05-20220315

Quant Time: Mar 22 07:27:32 2022  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\8270-SIM-BN031722.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Mar 17 13:18:23 2022  
Response via : Initial Calibration

