

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032122\
 Data File : BN019018.D
 Acq On : 21 Mar 2022 11:15
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 22 02:10:06 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 17 13:18:23 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	135	0.00
2	1,4-Dioxane	0.400	0.399	0.3	128	0.00
3	n-Nitrosodimethylamine	0.400	0.351	12.3	124	0.00
4 S	2-Fluorophenol	0.400	0.355	11.3	120	0.00
5 S	Phenol-d6	0.400	0.331	17.3	120	0.00
6	bis(2-Chloroethyl)ether	0.400	0.391	2.3	134	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	133	0.00
8 S	Nitrobenzene-d5	0.400	0.326	18.5	122	-0.01
9	Naphthalene	0.400	0.390	2.5	133	-0.01
10	Hexachlorobutadiene	0.400	0.379	5.3	127	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.371	7.3	126	0.00
12	2-Methylnaphthalene	0.400	0.371	7.3	127	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	120	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.272	32.0#	91	-0.01
15 S	2-Fluorobiphenyl	0.400	0.392	2.0	122	-0.01
16	Acenaphthylene	0.400	0.343	14.2	112	0.00
17	Acenaphthene	0.400	0.375	6.3	116	-0.01
18	Fluorene	0.400	0.369	7.8	112	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	113	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.299	25.3#	90	0.00
21	4-Bromophenyl-phenylether	0.400	0.372	7.0	109	0.00
22	Hexachlorobenzene	0.400	0.390	2.5	111	0.00
23	Atrazine	0.400	0.287	28.3#	91	-0.01
24	Pentachlorophenol	0.400	0.226	43.5#	74	0.00
25	Phenanthrene	0.400	0.378	5.5	112	0.00
26	Anthracene	0.400	0.340	15.0	104	0.00
27 SURR	Fluoranthene-d10	0.400	0.364	9.0	105	0.00
28	Fluoranthene	0.400	0.368	8.0	106	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	101	0.00
30	Pyrene	0.400	0.409	-2.2	105	0.00
31 S	Terphenyl-d14	0.400	0.379	5.3	99	0.00
32	Benzo(a)anthracene	0.400	0.332	17.0	94	0.00
33	Chrysene	0.400	0.403	-0.8	101	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.320	20.0	89	0.00
35 I	Perylene-d12	0.400	0.400	0.0	95	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.379	5.3	95	0.00
37	Benzo(b)fluoranthene	0.400	0.370	7.5	97	0.00
38	Benzo(k)fluoranthene	0.400	0.414	-3.5	99	0.00
39 C	Benzo(a)pyrene	0.400	0.373	6.8	91	0.00
40	Dibenzo(a,h)anthracene	0.400	0.329	17.8	82	0.00
41	Benzo(g,h,i)perylene	0.400	0.388	3.0	102	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4