

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN032219\
 Data File : BN004912.D
 Acq On : 23 Mar 2019 07:23
 Operator : JU/SJ
 Sample : K1817-07
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 EXT-027-SW015-030719

Manual Integrations
 APPROVED

Sohil
 3/25/2019 8:57:03 AM

Quant Time: Mar 24 00:49:16 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 20 16:06:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.69	152	4530	0.40	ng	0.00
7) Naphthalene-d8	10.48	136	19067	0.40	ng	-0.01
13) Acenaphthene-d10	14.33	164	10139	0.40	ng	-0.01
19) Phenanthrene-d10	17.08	188	21543	0.40	ng	-0.01
27) Chrysene-d12	21.28	240	18276	0.40	ng	-0.01
34) Perylene-d12	23.53	264	18773	0.40	ng	-0.02

System Monitoring Compounds						
4) 2-Fluorophenol	5.28	112	4997	0.41	ng	0.00
5) Phenol-d6	6.86	99	6392	0.42	ng	-0.02
8) Nitrobenzene-d5	8.86	82	5030	0.40	ng	-0.01
11) 2-Methylnaphthalene-d10	12.07	152	9726	0.29	ng	-0.01
14) 2,4,6-Tribromophenol	15.84	330	2128	0.30	ng	-0.01
15) 2-Fluorobiphenyl	12.95	172	13345	0.32	ng	-0.01
25) Fluoranthene-d10	19.12	212	14960	0.21	ng	-0.01
29) Terphenyl-d14	19.72	244	10544	0.28	ng	0.00

Target Compounds				Qvalue		
9) Naphthalene	10.52	128	15036	0.281	ng	99
12) 2-Methylnaphthalene	12.14	142	13022	0.328	ng	100
16) Acenaphthylene	14.05	152	11924	0.228	ng	# 94
17) Acenaphthene	14.39	154	70807	2.070	ng	99
18) Fluorene	15.38	166	64343	1.348	ng	# 98
23) Phenanthrene	17.12	178	932764	13.409	ng	99
24) Anthracene	17.22	178	238804	3.844	ng	# 93
26) Fluoranthene	19.15	202	960576	10.745	ng	99
28) Pyrene	19.51	202	765314	13.734	ng	# 93
30) Benzo(a)anthracene	21.26	228	418585	6.984	ng	97
31) Chrysene	21.31	228	324376	5.138	ng	96
32) Bis(2-ethylhexyl)phthalate	21.17	149	40006	1.671	ng	99
33) Indeno(1,2,3-cd)pyrene	25.79	276	161083	1.763	ng	99
35) Benzo(b)fluoranthene	22.85	252	379084	5.582	ng	99
36) Benzo(k)fluoranthene	22.89	252	126321m	1.891	ng	
37) Benzo(a)pyrene	23.43	252	276929	4.445	ng	98
38) Dibenzo(a,h)anthracene	25.79	278	49282	0.683	ng	# 88
39) Benzo(a,h,i)perylene	26.49	276	140742	1.914	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN032219\
Data File : BN004912.D
Acq On : 23 Mar 2019 07:23
Operator : JU/SJ
Sample : K1817-07
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
EXT-027-SW015-030719

Manual Integrations
APPROVED

Sohil
3/25/2019 8:57:03 AM

Quant Time: Mar 24 00:49:16 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 20 16:06:20 2019
Response via : Initial Calibration

