

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN032219\
 Data File : BN004916.D
 Acq On : 23 Mar 2019 09:43
 Operator : JU/SJ
 Sample : K2013-20
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PST-026-SW036-031819

Manual Integrations
 APPROVED

Sohil
 3/25/2019 8:57:14 AM

Quant Time: Mar 25 06:49:08 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 20 16:06:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.69	152	4331	0.40	ng	0.00
7) Naphthalene-d8	10.48	136	18381	0.40	ng	-0.01
13) Acenaphthene-d10	14.33	164	9962	0.40	ng	-0.01
19) Phenanthrene-d10	17.08	188	23036	0.40	ng	-0.01
27) Chrysene-d12	21.28	240	20933	0.40	ng	-0.01
34) Perylene-d12	23.53	264	19082	0.40	ng	-0.01

System Monitoring Compounds						
4) 2-Fluorophenol	5.28	112	4914	0.42	ng	0.00
5) Phenol-d6	6.86	99	6300	0.43	ng	-0.02
8) Nitrobenzene-d5	8.86	82	4842	0.40	ng	-0.01
11) 2-Methylnaphthalene-d10	12.07	152	10452	0.33	ng	-0.01
14) 2,4,6-Tribromophenol	15.84	330	2343	0.34	ng	-0.01
15) 2-Fluorobiphenyl	12.95	172	15195	0.37	ng	-0.01
25) Fluoranthene-d10	19.12	212	24426	0.32	ng	-0.01
29) Terphenyl-d14	19.72	244	17970	0.42	ng	-0.01

Target Compounds				Qvalue		
16) Acenaphthylene	14.05	152	1049	0.020	ng	96
23) Phenanthrene	17.12	178	2862	0.038	ng	96
26) Fluoranthene	19.15	202	4279	0.045	ng	# 94
28) Pyrene	19.52	202	4650	0.073	ng	# 96
30) Benzo(a)anthracene	21.27	228	2030	0.030	ng	# 71
31) Chrysene	21.32	228	2145	0.030	ng	# 77
32) Bis(2-ethylhexyl)phthalate	21.18	149	5495	0.200	ng	98
35) Benzo(b)fluoranthene	22.86	252	2485m	0.036	ng	
37) Benzo(a)pyrene	23.43	252	1631	0.026	ng	# 1
39) Benzo(a,h,i)perylene	26.50	276	1521	0.020	ng	# 42

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN032219\
Data File : BN004916.D
Acq On : 23 Mar 2019 09:43
Operator : JU/SJ
Sample : K2013-20
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PST-026-SW036-031819

Manual Integrations
APPROVED

Sohil
3/25/2019 8:57:14 AM

Quant Time: Mar 25 06:49:08 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 20 16:06:20 2019
Response via : Initial Calibration

