

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN032219\
 Data File : BN004890.D
 Acq On : 22 Mar 2019 19:01
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 23 02:52:00 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 20 16:06:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	170	0.00
2	1,4-Dioxane	0.400	0.432	-8.0	182	0.00
3	n-Nitrosodimethylamine	0.400	0.338	15.5	156	0.00
4 S	2-Fluorophenol	0.400	0.401	-0.3	168	0.00
5 S	Phenol-d6	0.400	0.382	4.5	164	0.00
6	bis(2-Chloroethyl)ether	0.400	0.404	-1.0	177	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	161	-0.01
8 S	Nitrobenzene-d5	0.400	0.403	-0.8	164	-0.01
9	Naphthalene	0.400	0.413	-3.2	165	-0.01
10	Hexachlorobutadiene	0.400	0.403	-0.8	161	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.395	1.3	159	0.00
12	2-Methylnaphthalene	0.400	0.392	2.0	157	-0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	149	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.327	18.3	130	-0.01
15 S	2-Fluorobiphenyl	0.400	0.423	-5.7	155	-0.01
16	Acenaphthylene	0.400	0.393	1.8	154	-0.01
17	Acenaphthene	0.400	0.404	-1.0	149	-0.01
18	Fluorene	0.400	0.395	1.3	144	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	146	-0.01
20	4-Bromophenyl-phenylether	0.400	0.394	1.5	145	-0.01
21	Hexachlorobenzene	0.400	0.384	4.0	139	-0.01
22	Pentachlorophenol	0.400	0.436	-9.0	185	-0.01
23	Phenanthrene	0.400	0.393	1.8	143	-0.01
24	Anthracene	0.400	0.382	4.5	143	-0.01
25 SURR	Fluoranthene-d10	0.400	0.403	-0.8	149	-0.01
26	Fluoranthene	0.400	0.390	2.5	142	-0.01
27 I	Chrysene-d12	0.400	0.400	0.0	142	0.00
28	Pvrene	0.400	0.404	-1.0	141	-0.01
29 S	Terphenyl-d14	0.400	0.417	-4.2	146	0.00
30	Benzo(a)anthracene	0.400	0.394	1.5	142	-0.01
31	Chrysene	0.400	0.392	2.0	134	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.386	3.5	148	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.355	11.3	121	0.00
34 I	Perylene-d12	0.400	0.400	0.0	138	0.00
35	Benzo(b)fluoranthene	0.400	0.389	2.8	132	0.00
36	Benzo(k)fluoranthene	0.400	0.372	7.0	130	0.00
37 C	Benzo(a)pyrene	0.400	0.382	4.5	132	0.00
38	Dibenzo(a,h)anthracene	0.400	0.355	11.3	121	0.00
39	Benzo(a,h,i)perylene	0.400	0.349	12.8	118	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN032219\
Data File : BN004890.D
Acq On : 22 Mar 2019 19:01
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 23 02:52:00 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 20 16:06:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				