

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN032219\
 Data File : BN004890.D
 Acq On : 22 Mar 2019 19:01
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 23 02:52:00 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 20 16:06:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	170#	0.00
2	1,4-Dioxane	0.455	0.491	-7.9	182#	0.00
3	n-Nitrosodimethylamine	0.269	0.227	15.6	156#	0.00
4 S	2-Fluorophenol	1.087	1.091	-0.4	168#	0.00
5 S	Phenol-d6	1.354	1.293	4.5	164#	0.00
6	bis(2-Chloroethyl)ether	1.221	1.235	-1.1	177#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	161#	-0.01
8 S	Nitrobenzene-d5	0.265	0.267	-0.8	164#	-0.01
9	Naphthalene	1.122	1.159	-3.3	165#	-0.01
10	Hexachlorobutadiene	0.214	0.216	-0.9	161#	-0.01
11 SURR	2-Methylnaphthalene-d10	0.698	0.689	1.3	159#	0.00
12	2-Methylnaphthalene	0.833	0.817	1.9	157#	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	149	-0.01
14 S	2,4,6-Tribromophenol	0.276	0.226	18.1	130	-0.01
15 S	2-Fluorobiphenyl	1.653	1.749	-5.8	155#	-0.01
16	Acenaphthylene	2.060	2.024	1.7	154#	-0.01
17	Acenaphthene	1.349	1.364	-1.1	149	-0.01
18	Fluorene	1.883	1.861	1.2	144	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	146	-0.01
20	4-Bromophenyl-phenylether	0.261	0.257	1.5	145	-0.01
21	Hexachlorobenzene	0.299	0.287	4.0	139	-0.01
22	Pentachlorophenol	0.058	0.064	-10.3	185#	-0.01
23	Phenanthrene	1.292	1.269	1.8	143	-0.01
24	Anthracene	1.154	1.102	4.5	143	-0.01
25 SURR	Fluoranthene-d10	1.339	1.351	-0.9	149	-0.01
26	Fluoranthene	1.660	1.616	2.7	142	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	142	0.00
28	Pvrene	1.220	1.231	-0.9	141	-0.01
29 S	Terphenyl-d14	0.819	0.853	-4.2	146	0.00
30	Benzo(a)anthracene	1.312	1.292	1.5	142	-0.01
31	Chrysene	1.382	1.353	2.1	134	-0.01
32	Bis(2-ethylhexyl)phthalate	0.524	0.505	3.6	148	0.00
33	Indeno(1,2,3-cd)pyrene	1.999	1.773	11.3	121	0.00
34 I	Perylene-d12	1.000	1.000	0.0	138	0.00
35	Benzo(b)fluoranthene	1.447	1.406	2.8	132	0.00
36	Benzo(k)fluoranthene	1.424	1.324	7.0	130	0.00
37 C	Benzo(a)pyrene	1.327	1.269	4.4	132	0.00
38	Dibenzo(a,h)anthracene	1.537	1.365	11.2	121	0.00
39	Benzo(a,h,i)perylene	1.566	1.367	12.7	118	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN032219\
Data File : BN004890.D
Acq On : 22 Mar 2019 19:01
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 23 02:52:00 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 20 16:06:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			