

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN032219\
 Data File : BN004900.D
 Acq On : 23 Mar 2019 00:20
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 25 06:36:47 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 20 16:06:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	185#	0.00
2	1,4-Dioxane	0.455	0.453	0.4	182#	0.00
3	n-Nitrosodimethylamine	0.269	0.259	3.7	192#	0.00
4 S	2-Fluorophenol	1.087	1.122	-3.2	187#	0.00
5 S	Phenol-d6	1.354	1.399	-3.3	192#	0.00
6	bis(2-Chloroethyl)ether	1.221	1.268	-3.8	197#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	186#	-0.01
8 S	Nitrobenzene-d5	0.265	0.274	-3.4	194#	-0.01
9	Naphthalene	1.122	1.164	-3.7	190#	-0.01
10	Hexachlorobutadiene	0.214	0.212	0.9	183#	-0.01
11 SURR	2-Methylnaphthalene-d10	0.698	0.705	-1.0	188#	0.00
12	2-Methylnaphthalene	0.833	0.838	-0.6	185#	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	176#	-0.01
14 S	2,4,6-Tribromophenol	0.276	0.235	14.9	160#	-0.01
15 S	2-Fluorobiphenyl	1.653	1.701	-2.9	178#	-0.01
16	Acenaphthylene	2.060	2.085	-1.2	187#	-0.01
17	Acenaphthene	1.349	1.367	-1.3	177#	-0.01
18	Fluorene	1.883	1.828	2.9	168#	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	167#	-0.01
20	4-Bromophenyl-phenylether	0.261	0.250	4.2	161#	-0.01
21	Hexachlorobenzene	0.299	0.283	5.4	156#	-0.01
22	Pentachlorophenol	0.058	0.071	-22.4	235#	-0.01
23	Phenanthrene	1.292	1.272	1.5	163#	-0.01
24	Anthracene	1.154	1.111	3.7	165#	-0.01
25 SURR	Fluoranthene-d10	1.339	1.322	1.3	166#	-0.01
26	Fluoranthene	1.660	1.569	5.5	157#	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	157#	-0.01
28	Pvrene	1.220	1.242	-1.8	157#	-0.01
29 S	Terphenyl-d14	0.819	0.826	-0.9	157#	-0.01
30	Benzo(a)anthracene	1.312	1.302	0.8	158#	-0.01
31	Chrysene	1.382	1.361	1.5	150	-0.01
32	Bis(2-ethylhexyl)phthalate	0.524	0.514	1.9	167#	-0.01
33	Indeno(1,2,3-cd)pyrene	1.999	1.672	16.4	127	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	151#	-0.02
35	Benzo(b)fluoranthene	1.447	1.425	1.5	146	-0.01
36	Benzo(k)fluoranthene	1.424	1.362	4.4	146	-0.01
37 C	Benzo(a)pyrene	1.327	1.276	3.8	146	-0.01
38	Dibenzo(a,h)anthracene	1.537	1.311	14.7	127	-0.02
39	Benzo(a,h,i)perylene	1.566	1.277	18.5	121	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN032219\
Data File : BN004900.D
Acq On : 23 Mar 2019 00:20
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 25 06:36:47 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 20 16:06:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			