

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN032219\
 Data File : BN004917.D
 Acq On : 23 Mar 2019 10:19
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDC00.4

Quant Time: Mar 24 00:55:42 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 20 16:06:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	282#	0.00
2	1,4-Dioxane	0.455	0.436	4.2	268#	0.00
3	n-Nitrosodimethylamine	0.269	0.262	2.6	297#	0.00
4 S	2-Fluorophenol	1.087	1.211	-11.4	308#	0.00
5 S	Phenol-d6	1.354	1.603	-18.4	336#	0.00
6	bis(2-Chloroethyl)ether	1.221	1.301	-6.6	308#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	289#	-0.01
8 S	Nitrobenzene-d5	0.265	0.292	-10.2	321#	-0.01
9	Naphthalene	1.122	1.156	-3.0	294#	-0.03
10	Hexachlorobutadiene	0.214	0.211	1.4	282#	-0.01
11 SURR	2-Methylnaphthalene-d10	0.698	0.686	1.7	284#	-0.01
12	2-Methylnaphthalene	0.833	0.814	2.3	279#	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	254#	-0.01
14 S	2,4,6-Tribromophenol	0.276	0.258	6.5	253#	-0.01
15 S	2-Fluorobiphenyl	1.653	1.740	-5.3	262#	-0.01
16	Acenaphthylene	2.060	2.233	-8.4	289#	-0.01
17	Acenaphthene	1.349	1.375	-1.9	256#	-0.01
18	Fluorene	1.883	1.793	4.8	237#	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	226#	-0.01
20	4-Bromophenyl-phenylether	0.261	0.261	0.0	227#	-0.01
21	Hexachlorobenzene	0.299	0.286	4.3	214#	-0.01
22	Pentachlorophenol	0.058	0.101	-74.1#	450#	-0.01
23	Phenanthrene	1.292	1.282	0.8	222#	-0.01
24	Anthracene	1.154	1.179	-2.2	236#	-0.01
25 SURR	Fluoranthene-d10	1.339	1.310	2.2	223#	0.00
26	Fluoranthene	1.660	1.542	7.1	209#	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	169#	-0.01
28	Pvrene	1.220	1.520	-24.6	207#	0.00
29 S	Terphenyl-d14	0.819	1.009	-23.2	206#	0.00
30	Benzo(a)anthracene	1.312	1.388	-5.8	181#	-0.01
31	Chrysene	1.382	1.369	0.9	162#	-0.01
32	Bis(2-ethylhexyl)phthalate	0.524	0.921	-75.8#	322#	0.00
33	Indeno(1,2,3-cd)pyrene	1.999	1.368	31.6#	112	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	134	-0.02
35	Benzo(b)fluoranthene	1.447	1.452	-0.3	132	-0.01
36	Benzo(k)fluoranthene	1.424	1.359	4.6	130	-0.01
37 C	Benzo(a)pyrene	1.327	1.276	3.8	130	-0.01
38	Dibenzo(a,h)anthracene	1.537	1.288	16.2	111	-0.01
39	Benzo(a,h,i)perylene	1.566	1.284	18.0	108	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN032219\
Data File : BN004917.D
Acq On : 23 Mar 2019 10:19
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 24 00:55:42 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 20 16:06:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			