


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032219\
 Data File : BN004901.D
 Acq On : 23 Mar 2019 00:55
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032019.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Mar 20 16:06:20 2019



AutoFind: Scans 2467, 2468, 2469; Background Corrected with Scan 2462

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	26.1	3842	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	30.3	4454	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	44.1	6489	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	14709	PASS
199	198	5	9	6.8	1007	PASS
275	198	10	60	31.4	4621	PASS
365	198	1	100	3.2	472	PASS
441	198	0.01	100	12.8	1881	PASS
442	198	50	100	80.7	11872	PASS
443	442	15	24	19.4	2304	PASS