

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032221\
 Data File : BN014026.D
 Acq On : 22 Mar 2021 23:05
 Operator : CG/JU
 Sample : PB135271BSD
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB135271BSD

Quant Time: Mar 23 00:21:48 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 19 14:40:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.700	152	5692	0.40	ng	0.00
7) Naphthalene-d8	10.479	136	20534	0.40	ng	# 0.00
13) Acenaphthene-d10	14.321	164	10652	0.40	ng	-0.01
19) Phenanthrene-d10	17.067	188	23005	0.40	ng	# 0.00
29) Chrysene-d12	21.237	240	23389	0.40	ng	-0.01
36) Perylene-d12	23.449	264	23916	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.300	112	5690	0.43	ng	0.00
5) Phenol-d6	6.863	99	7079	0.43	ng	0.00
8) Nitrobenzene-d5	8.845	82	5154	0.38	ng	0.00
11) 2-Methylnaphthalene-d10	12.066	152	12826	0.40	ng	0.00
14) 2,4,6-Tribromophenol	15.818	330	1236	0.36	ng	0.00
15) 2-Fluorobiphenyl	12.950	172	16870	0.42	ng	0.00
27) Fluoranthene-d10	19.091	212	26676	0.42	ng	0.00
31) Terphenyl-d14	19.692	244	21456	0.42	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.247	88	3118	0.42	ng	96
3) n-Nitrosodimethylamine	3.585	42	1518	0.30	ng	# 98
6) bis(2-Chloroethyl)ether	7.128	93	6094	0.42	ng	96
9) Naphthalene	10.518	128	20941	0.41	ng	99
10) Hexachlorobutadiene	10.809	225	3903	0.41	ng	# 100
12) 2-Methylnaphthalene	12.137	142	13858	0.40	ng	99
16) Acenaphthylene	14.042	152	16110	0.39	ng	99
17) Acenaphthene	14.384	154	12539	0.42	ng	100
18) Fluorene	15.374	166	15453	0.41	ng	100
20) 4,6-Dinitro-2-methylph...	15.450	198	928	0.36	ng	# 86
21) 4-Bromophenyl-phenylether	16.264	248	4853	0.41	ng	# 90
22) Hexachlorobenzene	16.373	284	5690	0.42	ng	99
23) Atrazine	16.531	200	2891	0.36	ng	98
24) Pentachlorophenol	16.714	266	1126	0.35	ng	94
25) Phenanthrene	17.103	178	25666	0.41	ng	100
26) Anthracene	17.189	178	20785	0.40	ng	100
28) Fluoranthene	19.121	202	28809	0.41	ng	99
30) Pyrene	19.483	202	29384	0.41	ng	100
32) Benzo(a)anthracene	21.226	228	26167	0.39	ng	99
33) Chrysene	21.279	228	31103	0.43	ng	96
34) Bis(2-ethylhexyl)phtha...	21.162	149	8020	0.32	ng	99
35) Indeno(1,2,3-cd)pyrene	25.656	276	38711	0.49	ng	99
37) Benzo(b)fluoranthene	22.785	252	34827	0.40	ng	# 97
38) Benzo(k)fluoranthene	22.785	252	34827	0.41	ng	97
39) Benzo(a)pyrene	23.353	252	31110	0.42	ng	# 95
40) Dibenzo(a,h)anthracene	25.674	278	32505	0.41	ng	98
41) Benzo(g,h,i)perylene	26.334	276	34095	0.41	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032221\
Data File : BN014026.D
Acq On : 22 Mar 2021 23:05
Operator : CG/JU
Sample : PB135271BSD
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB135271BSD

Quant Time: Mar 23 00:21:48 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 19 14:40:09 2021
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.700 min Scan# 611
 Delta R.T. 0.000 min
 Lab File: BN014026.D
 Acq: 22 Mar 2021 23:05




Instrument :
 BNA_N
 ClientSampleId :
 PB135271BSD


Tgt Ion:152 Resp: 5692
 Ion Ratio Lower Upper
 152 100
 150 155.7 121.8 182.8
 115 60.0 46.6 69.8



#2
 1,4-Dioxane
 Concen: 0.42 ng
 RT: 3.247 min Scan# 58
 Delta R.T. -0.000 min
 Lab File: BN014026.D
 Acq: 22 Mar 2021 23:05

Tgt Ion: 88 Resp: 3118
 Ion Ratio Lower Upper
 88 100
 58 77.3 64.6 96.8
 43 31.5 26.6 40.0









#3
n-Nitrosodimethylamine
Concen: 0.30 ng
RT: 3.585 min Scan# 100
Delta R.T. 0.008 min
Lab File: BN014026.D
Acq: 22 Mar 2021 23:05




Instrument :
BNA_N
ClientSampleId :
PB135271BSD


Tgt Ion: 42 Resp: 1518
Ion Ratio Lower Upper
42 100
74 117.3 95.0 142.6
44 3.8 5.8 8.8#



#4
2-Fluorophenol
Concen: 0.43 ng
RT: 5.300 min Scan# 313
Delta R.T. -0.000 min
Lab File: BN014026.D
Acq: 22 Mar 2021 23:05

Tgt Ion: 112 Resp: 5690
Ion Ratio Lower Upper
112 100
64 59.8 44.3 66.5
63 31.8 23.0 34.6









#5
Phenol-d6
Concen: 0.43 ng
RT: 6.863 min Scan# 507
Delta R.T. -0.008 min
Lab File: BN014026.D
Acq: 22 Mar 2021 23:05




Instrument :
BNA_N
ClientSampleId :
PB135271BSD


Tgt Ion: 99 Resp: 7079
Ion Ratio Lower Upper
99 100
42 22.8 15.2 22.8#
71 33.0 26.1 39.1



#6
bis(2-Chloroethyl)ether
Concen: 0.42 ng
RT: 7.128 min Scan# 540
Delta R.T. -0.008 min
Lab File: BN014026.D
Acq: 22 Mar 2021 23:05

Tgt Ion: 93 Resp: 6094
Ion Ratio Lower Upper
93 100
63 80.7 60.8 91.2
95 33.1 26.0 39.0









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.479 min Scan# 858
Delta R.T. -0.000 min
Lab File: BN014026.D
Acq: 22 Mar 2021 23:05




Instrument :
BNA_N
ClientSampleId :
PB135271BSD


Tgt Ion:	136	Resp:	20534
Ion Ratio	Lower	Upper	
136	100		
137	10.9	9.0	13.6
54	5.7	6.4	9.6#
68	4.5	4.5	6.7



#8
Nitrobenzene-d5
Concen: 0.38 ng
RT: 8.845 min Scan# 729
Delta R.T. -0.000 min
Lab File: BN014026.D
Acq: 22 Mar 2021 23:05

Tgt Ion:	82	Resp:	5154
Ion Ratio	Lower	Upper	
82	100		
128	41.2	39.2	58.8
54	50.3	40.1	60.1









#9
Naphthalene
Concen: 0.41 ng
RT: 10.518 min Scan# 861
Delta R.T. -0.013 min
Lab File: BN014026.D
Acq: 22 Mar 2021 23:05




Instrument :
BNA_N
ClientSampleId :
PB135271BSD


Tgt Ion:	128	Resp:	20941
Ion Ratio	Lower	Upper	
128	100		
129	11.3	9.0	13.4
127	13.4	10.4	15.6



#10
Hexachlorobutadiene
Concen: 0.41 ng
RT: 10.809 min Scan# 884
Delta R.T. -0.013 min
Lab File: BN014026.D
Acq: 22 Mar 2021 23:05

Tgt Ion:	225	Resp:	3903
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	64.0	51.2	76.8









#11
2-Methylnaphthalene-d10
Concen: 0.40 ng
RT: 12.066 min Scan# 1065
Delta R.T. -0.005 min
Lab File: BN014026.D
Acq: 22 Mar 2021 23:05




Instrument :
BNA_N
ClientSampleId :
PB135271BSD


Tgt Ion:152 Resp: 12826
Ion Ratio Lower Upper
152 100
151 20.3 16.2 24.2



#12
2-Methylnaphthalene
Concen: 0.40 ng
RT: 12.137 min Scan# 1081
Delta R.T. -0.005 min
Lab File: BN014026.D
Acq: 22 Mar 2021 23:05

Tgt Ion:142 Resp: 13858
Ion Ratio Lower Upper
142 100
141 88.9 70.5 105.7
115 36.0 27.4 41.0









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.321 min Scan# 1307
Delta R.T. -0.013 min
Lab File: BN014026.D
Acq: 22 Mar 2021 23:05




Instrument :
BNA_N
ClientSampleId :
PB135271BSD


Tgt Ion:164 Resp: 10652
Ion Ratio Lower Upper
164 100
162 104.1 79.6 119.4
160 48.4 36.0 54.0



#14
2,4,6-Tribromophenol
Concen: 0.36 ng
RT: 15.818 min Scan# 1425
Delta R.T. -0.000 min
Lab File: BN014026.D
Acq: 22 Mar 2021 23:05

Tgt Ion:330 Resp: 1236
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 38.1 26.8 40.2









#15
2-Fluorobiphenyl
Concen: 0.42 ng
RT: 12.950 min Scan# 1199
Delta R.T. -0.000 min
Lab File: BN014026.D
Acq: 22 Mar 2021 23:05




Instrument :
BNA_N
ClientSampleId :
PB135271BSD


Tgt Ion:	172	Resp:	16870
Ion	Ratio	Lower	Upper
172	100		
171	33.5	28.6	42.8
170	22.0	18.9	28.3



#16
Acenaphthylene
Concen: 0.39 ng
RT: 14.042 min Scan# 1285
Delta R.T. -0.000 min
Lab File: BN014026.D
Acq: 22 Mar 2021 23:05

Tgt Ion:	152	Resp:	16110
Ion	Ratio	Lower	Upper
152	100		
151	19.1	15.7	23.5
153	13.0	10.5	15.7









#17
Acenaphthene
Concen: 0.42 ng
RT: 14.384 min Scan# 1312
Delta R.T. -0.000 min
Lab File: BN014026.D
Acq: 22 Mar 2021 23:05




Instrument :
BNA_N
ClientSampleId :
PB135271BSD


Tgt Ion:154 Resp: 12539
Ion Ratio Lower Upper
154 100
153 110.4 88.4 132.6
152 53.9 43.0 64.4



#18
Fluorene
Concen: 0.41 ng
RT: 15.374 min Scan# 1390
Delta R.T. -0.000 min
Lab File: BN014026.D
Acq: 22 Mar 2021 23:05

Tgt Ion:166 Resp: 15453
Ion Ratio Lower Upper
166 100
165 97.8 77.8 116.8
167 13.3 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.067 min Scan# 1527
Delta R.T. -0.001 min
Lab File: BN014026.D
Acq: 22 Mar 2021 23:05




Instrument :
BNA_N
ClientSampleId :
PB135271BSD


Tgt Ion:188 Resp: 23005
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 6.9 9.0 13.6#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.36 ng
RT: 15.450 min Scan# 1396
Delta R.T. -0.000 min
Lab File: BN014026.D
Acq: 22 Mar 2021 23:05

Tgt Ion:198 Resp: 928
Ion Ratio Lower Upper
198 100
51 93.5 61.7 92.5#
105 44.8 33.0 49.6









#21
4-Bromophenyl-phenylether
Concen: 0.41 ng
RT: 16.264 min Scan# 1461
Delta R.T. -0.001 min
Lab File: BN014026.D
Acq: 22 Mar 2021 23:05




Instrument :
BNA_N
ClientSampleId :
PB135271BSD


Tgt Ion:248 Resp: 4853
Ion Ratio Lower Upper
248 100
250 97.2 75.5 113.3
141 52.1 53.8 80.8#



#22
Hexachlorobenzene
Concen: 0.42 ng
RT: 16.373 min Scan# 1470
Delta R.T. -0.013 min
Lab File: BN014026.D
Acq: 22 Mar 2021 23:05

Tgt Ion:284 Resp: 5690
Ion Ratio Lower Upper
284 100
142 34.7 27.9 41.9
249 30.7 24.2 36.2









#23
Atrazine
Concen: 0.36 ng
RT: 16.531 min Scan# 1483
Delta R.T. -0.001 min
Lab File: BN014026.D
Acq: 22 Mar 2021 23:05




Instrument :
BNA_N
ClientSampleId :
PB135271BSD


Tgt Ion:	200	Resp:	2891
Ion Ratio	Lower	Upper	
200	100		
173	25.3	18.6	28.0
215	50.3	41.0	61.6



#24
Pentachlorophenol
Concen: 0.35 ng
RT: 16.714 min Scan# 1498
Delta R.T. -0.013 min
Lab File: BN014026.D
Acq: 22 Mar 2021 23:05

Tgt Ion:	266	Resp:	1126
Ion Ratio	Lower	Upper	
266	100		
264	59.1	51.6	77.4
268	63.4	54.6	82.0









#25
Phenanthrene
Concen: 0.41 ng
RT: 17.103 min Scan# 1530
Delta R.T. -0.001 min
Lab File: BN014026.D
Acq: 22 Mar 2021 23:05




Instrument :
BNA_N
ClientSampleId :
PB135271BSD


Tgt Ion:	178	Resp:	25666
Ion Ratio	Lower	Upper	
178	100		
176	18.9	15.0	22.4
179	15.5	12.3	18.5



#26
Anthracene
Concen: 0.40 ng
RT: 17.189 min Scan# 1537
Delta R.T. -0.013 min
Lab File: BN014026.D
Acq: 22 Mar 2021 23:05

Tgt Ion:	178	Resp:	20785
Ion Ratio	Lower	Upper	
178	100		
176	18.1	14.6	21.8
179	15.5	12.2	18.2









#27
Fluoranthene-d10
Concen: 0.42 ng
RT: 19.091 min Scan# 1823
Delta R.T. -0.008 min
Lab File: BN014026.D
Acq: 22 Mar 2021 23:05




Instrument :
BNA_N
ClientSampleId :
PB135271BSD


Tgt Ion:212 Resp: 26676
Ion Ratio Lower Upper
212 100
106 12.8 9.7 14.5
104 7.1 5.5 8.3



#28
Fluoranthene
Concen: 0.41 ng
RT: 19.121 min Scan# 1829
Delta R.T. -0.008 min
Lab File: BN014026.D
Acq: 22 Mar 2021 23:05

Tgt Ion:202 Resp: 28809
Ion Ratio Lower Upper
202 100
101 11.1 8.1 12.1
203 17.1 13.9 20.9









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.237 min Scan# 2123
Delta R.T. -0.013 min
Lab File: BN014026.D
Acq: 22 Mar 2021 23:05




Instrument :
BNA_N
ClientSampleId :
PB135271BSD


Tgt Ion:240	Resp:	23389
Ion Ratio	Lower	Upper
240	100	
120	12.1	10.2 15.4
236	26.9	21.3 31.9



#30
Pyrene
Concen: 0.41 ng
RT: 19.483 min Scan# 1902
Delta R.T. -0.008 min
Lab File: BN014026.D
Acq: 22 Mar 2021 23:05

Tgt Ion:202	Resp:	29384
Ion Ratio	Lower	Upper
202	100	
200	20.4	16.3 24.5
203	17.6	14.0 21.0









#31
Terphenyl-d14
Concen: 0.42 ng
RT: 19.692 min Scan# 1944
Delta R.T. -0.008 min
Lab File: BN014026.D
Acq: 22 Mar 2021 23:05




Instrument :
BNA_N
ClientSampleId :
PB135271BSD


Tgt Ion:244 Resp: 21456
Ion Ratio Lower Upper
244 100
212 7.7 6.1 9.1
122 11.9 9.8 14.8



#32
Benzo(a)anthracene
Concen: 0.39 ng
RT: 21.226 min Scan# 2122
Delta R.T. -0.003 min
Lab File: BN014026.D
Acq: 22 Mar 2021 23:05

Tgt Ion:228 Resp: 26167
Ion Ratio Lower Upper
228 100
226 26.2 20.5 30.7
229 19.9 16.2 24.4





#33

Chrysene

Concen: 0.43 ng

RT: 21.279 min Scan# 2127

Delta R.T. -0.003 min

Lab File: BN014026.D

Acq: 22 Mar 2021 23:05

Tgt Ion:228 Resp: 31103

Ion Ratio Lower Upper

228 100

226 28.3 24.3 36.5





229 20.2 15.1 22.7

Instrument :

BNA_N

ClientSampleId :

PB135271BSD



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.32 ng

RT: 21.162 min Scan# 2116

Delta R.T. -0.003 min

Lab File: BN014026.D

Acq: 22 Mar 2021 23:05




Tgt Ion:149 Resp: 8020


Ion Ratio Lower Upper

149 100

167 27.1 22.1 33.1

279 4.1 3.5 5.3









#35
Indeno(1,2,3-cd)pyrene
Concen: 0.49 ng
RT: 25.656 min Scan# 3183
Delta R.T. -0.017 min
Lab File: BN014026.D
Acq: 22 Mar 2021 23:05




Instrument :
BNA_N
ClientSampleId :
PB135271BSD


Tgt Ion:276 Resp: 38711
Ion Ratio Lower Upper
276 100
138 23.5 19.8 29.6
277 24.8 19.9 29.9



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.449 min Scan# 2538
Delta R.T. -0.010 min
Lab File: BN014026.D
Acq: 22 Mar 2021 23:05

Tgt Ion:264 Resp: 23916
Ion Ratio Lower Upper
264 100
260 24.4 19.3 28.9
265 34.0 66.2 99.4#









#37
Benzo(b)fluoranthene
Concen: 0.40 ng
RT: 22.785 min Scan# 2344
Delta R.T. -0.010 min
Lab File: BN014026.D
Acq: 22 Mar 2021 23:05




Instrument :
BNA_N
ClientSampleId :
PB135271BSD


Tgt Ion:252 Resp: 34827
Ion Ratio Lower Upper
252 100
253 24.9 20.4 30.6
125 17.6 11.6 17.4#



#38
Benzo(k)fluoranthene
Concen: 0.41 ng
RT: 22.785 min Scan# 2344
Delta R.T. -0.054 min
Lab File: BN014026.D
Acq: 22 Mar 2021 23:05

Tgt Ion:252 Resp: 34827
Ion Ratio Lower Upper
252 100
253 24.9 20.5 30.7
125 17.6 12.3 18.5





#39

Benzo(a)pyrene

Concen: 0.42 ng





RT: 23.353 min Scan# 2510

Delta R.T. -0.010 min

Lab File: BN014026.D

Acq: 22 Mar 2021 23:05

Tgt Ion:252	Resp:	31110
Ion Ratio	Lower	Upper
252	100	
253	27.3	21.8 32.6
125	23.6	14.6 21.8#



#40

Dibenzo(a,h)anthracene

Concen: 0.41 ng




RT: 25.674 min Scan# 3188


Delta R.T. -0.013 min

Lab File: BN014026.D

Acq: 22 Mar 2021 23:05

Tgt Ion:278	Resp:	32505
Ion Ratio	Lower	Upper
278	100	
139	16.3	13.1 19.7
279	24.9	21.4 32.0





#41
Benzo(g,h,i)perylene
Concen: 0.41 ng
RT: 26.334 min Scan# 3381
Delta R.T. -0.017 min
Lab File: BN014026.D
Acq: 22 Mar 2021 23:05

Instrument :
BNA_N
ClientSampleId :
PB135271BSD

Tgt Ion:	276	Resp:	34095
Ion	Ratio	Lower	Upper
276	100		
277	24.0	20.0	30.0
138	21.6	17.8	26.8

