

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032221\
 Data File : BN014029.D
 Acq On : 23 Mar 2021 02:04
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Mar 23 02:38:01 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 19 14:40:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.700	152	5984	0.40	ng	0.00
7) Naphthalene-d8	10.480	136	21756	0.40	ng	# 0.00
13) Acenaphthene-d10	14.321	164	11884	0.40	ng	-0.01
19) Phenanthrene-d10	17.068	188	25833	0.40	ng	# 0.00
29) Chrysene-d12	21.239	240	25745	0.40	ng	#-0.01
36) Perylene-d12	23.448	264	24848	0.40	ng	#-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.300	112	5989	0.43	ng	0.00
5) Phenol-d6	6.871	99	7541	0.43	ng	0.00
8) Nitrobenzene-d5	8.845	82	5423	0.38	ng	0.00
11) 2-Methylnaphthalene-d10	12.066	152	13942	0.41	ng	0.00
14) 2,4,6-Tribromophenol	15.818	330	1396	0.36	ng	0.00
15) 2-Fluorobiphenyl	12.951	172	18607	0.42	ng	0.00
27) Fluoranthene-d10	19.094	212	29802	0.42	ng	0.00
31) Terphenyl-d14	19.695	244	23995	0.42	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.247	88	3094	0.39	ng	99
3) n-Nitrosodimethylamine	3.585	42	1569	0.30	ng	# 95
6) bis(2-Chloroethyl)ether	7.129	93	6406	0.42	ng	95
9) Naphthalene	10.518	128	22485	0.42	ng	99
10) Hexachlorobutadiene	10.809	225	4198	0.41	ng	# 99
12) 2-Methylnaphthalene	12.138	142	15019	0.41	ng	99
16) Acenaphthylene	14.042	152	17673	0.39	ng	100
17) Acenaphthene	14.384	154	13777	0.41	ng	99
18) Fluorene	15.374	166	17133	0.41	ng	100
20) 4,6-Dinitro-2-methylph...	15.450	198	1024	0.35	ng	94
21) 4-Bromophenyl-phenylether	16.264	248	5428	0.41	ng	# 86
22) Hexachlorobenzene	16.374	284	6363	0.42	ng	99
23) Atrazine	16.532	200	3282	0.36	ng	97
24) Pentachlorophenol	16.715	266	1209	0.33	ng	92
25) Phenanthrene	17.104	178	28855	0.41	ng	100
26) Anthracene	17.189	178	23103	0.40	ng	99
28) Fluoranthene	19.124	202	32469	0.41	ng	99
30) Pyrene	19.486	202	32643	0.41	ng	100
32) Benzo(a)anthracene	21.218	228	28570	0.39	ng	97
33) Chrysene	21.271	228	33603	0.42	ng	100
34) Bis(2-ethylhexyl)phtha...	21.154	149	8622	0.32	ng	99
35) Indeno(1,2,3-cd)pyrene	25.656	276	39471	0.46	ng	99
37) Benzo(b)fluoranthene	22.788	252	36765	0.41	ng	97
38) Benzo(k)fluoranthene	22.788	252	36765	0.41	ng	98
39) Benzo(a)pyrene	23.349	252	32620	0.42	ng	# 93
40) Dibenzo(a,h)anthracene	25.670	278	33027	0.40	ng	98
41) Benzo(g,h,i)perylene	26.334	276	34495	0.40	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032221\
Data File : BN014029.D
Acq On : 23 Mar 2021 02:04
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Quant Time: Mar 23 02:38:01 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 19 14:40:09 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.700 min Scan# 611
 Delta R.T. 0.000 min
 Lab File: BN014029.D
 Acq: 23 Mar 2021 02:04




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion: 152 Resp: 5984
 Ion Ratio Lower Upper
 152 100
 150 155.9 121.8 182.8
 115 59.6 46.6 69.8



#2
 1,4-Dioxane
 Concen: 0.39 ng
 RT: 3.247 min Scan# 58
 Delta R.T. 0.000 min
 Lab File: BN014029.D
 Acq: 23 Mar 2021 02:04

Tgt Ion: 88 Resp: 3094
 Ion Ratio Lower Upper
 88 100
 58 80.6 64.6 96.8
 43 31.9 26.6 40.0









#3
n-Nitrosodimethylamine
Concen: 0.30 ng
RT: 3.585 min Scan# 100
Delta R.T. 0.008 min
Lab File: BN014029.D
Acq: 23 Mar 2021 02:04




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 42 Resp: 1569
Ion Ratio Lower Upper
42 100
74 114.5 95.0 142.6
44 0.0 5.8 8.8#



#4
2-Fluorophenol
Concen: 0.43 ng
RT: 5.300 min Scan# 313
Delta R.T. 0.000 min
Lab File: BN014029.D
Acq: 23 Mar 2021 02:04

Tgt Ion: 112 Resp: 5989
Ion Ratio Lower Upper
112 100
64 59.6 44.3 66.5
63 31.9 23.0 34.6









#5
Phenol-d6
Concen: 0.43 ng
RT: 6.871 min Scan# 508
Delta R.T. 0.000 min
Lab File: BN014029.D
Acq: 23 Mar 2021 02:04




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 99 Resp: 7541
Ion Ratio Lower Upper
99 100
42 22.0 15.2 22.8
71 33.0 26.1 39.1



#6
bis(2-Chloroethyl)ether
Concen: 0.42 ng
RT: 7.129 min Scan# 540
Delta R.T. -0.008 min
Lab File: BN014029.D
Acq: 23 Mar 2021 02:04

Tgt Ion: 93 Resp: 6406
Ion Ratio Lower Upper
93 100
63 81.3 60.8 91.2
95 33.3 26.0 39.0









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.480 min Scan# 858
Delta R.T. 0.000 min
Lab File: BN014029.D
Acq: 23 Mar 2021 02:04




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	136	Resp:	21756
Ion Ratio	Lower	Upper	
136	100		
137	11.4	9.0	13.6
54	5.4	6.4	9.6#
68	4.4	4.5	6.7#



#8
Nitrobenzene-d5
Concen: 0.38 ng
RT: 8.845 min Scan# 729
Delta R.T. 0.000 min
Lab File: BN014029.D
Acq: 23 Mar 2021 02:04

Tgt Ion:	82	Resp:	5423
Ion Ratio	Lower	Upper	
82	100		
128	42.0	39.2	58.8
54	50.8	40.1	60.1









#9
Naphthalene
Concen: 0.42 ng
RT: 10.518 min Scan# 861
Delta R.T. -0.013 min
Lab File: BN014029.D
Acq: 23 Mar 2021 02:04




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	128	Resp:	22485
Ion Ratio	Lower	Upper	
128	100		
129	11.2	9.0	13.4
127	13.4	10.4	15.6



#10
Hexachlorobutadiene
Concen: 0.41 ng
RT: 10.809 min Scan# 884
Delta R.T. -0.013 min
Lab File: BN014029.D
Acq: 23 Mar 2021 02:04

Tgt Ion:	225	Resp:	4198
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	64.5	51.2	76.8









#11
2-Methylnaphthalene-d10
Concen: 0.41 ng
RT: 12.066 min Scan# 1065
Delta R.T. -0.004 min
Lab File: BN014029.D
Acq: 23 Mar 2021 02:04




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 13942
Ion Ratio Lower Upper
152 100
151 20.2 16.2 24.2



#12
2-Methylnaphthalene
Concen: 0.41 ng
RT: 12.138 min Scan# 1081
Delta R.T. -0.004 min
Lab File: BN014029.D
Acq: 23 Mar 2021 02:04

Tgt Ion:142 Resp: 15019
Ion Ratio Lower Upper
142 100
141 87.9 70.5 105.7
115 35.1 27.4 41.0









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.321 min Scan# 1307
Delta R.T. -0.013 min
Lab File: BN014029.D
Acq: 23 Mar 2021 02:04




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:164 Resp: 11884
Ion Ratio Lower Upper
164 100
162 101.2 79.6 119.4
160 46.8 36.0 54.0



#14
2,4,6-Tribromophenol
Concen: 0.36 ng
RT: 15.818 min Scan# 1425
Delta R.T. 0.000 min
Lab File: BN014029.D
Acq: 23 Mar 2021 02:04

Tgt Ion:330 Resp: 1396
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 36.3 26.8 40.2









#15
2-Fluorobiphenyl
Concen: 0.42 ng
RT: 12.951 min Scan# 1199
Delta R.T. 0.000 min
Lab File: BN014029.D
Acq: 23 Mar 2021 02:04




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	172	Resp:	18607
Ion Ratio	Lower	Upper	
172	100		
171	33.1	28.6	42.8
170	21.6	18.9	28.3



#16
Acenaphthylene
Concen: 0.39 ng
RT: 14.042 min Scan# 1285
Delta R.T. 0.000 min
Lab File: BN014029.D
Acq: 23 Mar 2021 02:04

Tgt Ion:	152	Resp:	17673
Ion Ratio	Lower	Upper	
152	100		
151	19.3	15.7	23.5
153	13.0	10.5	15.7









#17
Acenaphthene
Concen: 0.41 ng
RT: 14.384 min Scan# 1312
Delta R.T. 0.000 min
Lab File: BN014029.D
Acq: 23 Mar 2021 02:04




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:154 Resp: 13777
Ion Ratio Lower Upper
154 100
153 109.8 88.4 132.6
152 53.8 43.0 64.4



#18
Fluorene
Concen: 0.41 ng
RT: 15.374 min Scan# 1390
Delta R.T. 0.000 min
Lab File: BN014029.D
Acq: 23 Mar 2021 02:04

Tgt Ion:166 Resp: 17133
Ion Ratio Lower Upper
166 100
165 97.3 77.8 116.8
167 13.3 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.068 min Scan# 1527
Delta R.T. 0.000 min
Lab File: BN014029.D
Acq: 23 Mar 2021 02:04




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:188 Resp: 25833
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 6.4 9.0 13.6#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.35 ng
RT: 15.450 min Scan# 1396
Delta R.T. 0.000 min
Lab File: BN014029.D
Acq: 23 Mar 2021 02:04

Tgt Ion:198 Resp: 1024
Ion Ratio Lower Upper
198 100
51 84.1 61.7 92.5
105 39.5 33.0 49.6









#21
4-Bromophenyl-phenylether
Concen: 0.41 ng
RT: 16.264 min Scan# 1461
Delta R.T. 0.000 min
Lab File: BN014029.D
Acq: 23 Mar 2021 02:04




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:248 Resp: 5428
Ion Ratio Lower Upper
248 100
250 98.7 75.5 113.3
141 45.5 53.8 80.8#



#22
Hexachlorobenzene
Concen: 0.42 ng
RT: 16.374 min Scan# 1470
Delta R.T. -0.012 min
Lab File: BN014029.D
Acq: 23 Mar 2021 02:04

Tgt Ion:284 Resp: 6363
Ion Ratio Lower Upper
284 100
142 35.5 27.9 41.9
249 30.7 24.2 36.2









#23
Atrazine
Concen: 0.36 ng
RT: 16.532 min Scan# 1483
Delta R.T. 0.000 min
Lab File: BN014029.D
Acq: 23 Mar 2021 02:04




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:200 Resp: 3282
Ion Ratio Lower Upper
200 100
173 23.6 18.6 28.0
215 48.9 41.0 61.6



#24
Pentachlorophenol
Concen: 0.33 ng
RT: 16.715 min Scan# 1498
Delta R.T. -0.012 min
Lab File: BN014029.D
Acq: 23 Mar 2021 02:04

Tgt Ion:266 Resp: 1209
Ion Ratio Lower Upper
266 100
264 59.5 51.6 77.4
268 61.0 54.6 82.0









#25
Phenanthrene
Concen: 0.41 ng
RT: 17.104 min Scan# 1530
Delta R.T. 0.000 min
Lab File: BN014029.D
Acq: 23 Mar 2021 02:04




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:178 Resp: 28855
Ion Ratio Lower Upper
178 100
176 18.9 15.0 22.4
179 15.5 12.3 18.5



#26
Anthracene
Concen: 0.40 ng
RT: 17.189 min Scan# 1537
Delta R.T. -0.012 min
Lab File: BN014029.D
Acq: 23 Mar 2021 02:04

Tgt Ion:178 Resp: 23103
Ion Ratio Lower Upper
178 100
176 18.7 14.6 21.8
179 15.4 12.2 18.2









#27
Fluoranthene-d10
Concen: 0.42 ng
RT: 19.094 min Scan# 1822
Delta R.T. -0.005 min
Lab File: BN014029.D
Acq: 23 Mar 2021 02:04




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:212 Resp: 29802
Ion Ratio Lower Upper
212 100
106 12.6 9.7 14.5
104 7.1 5.5 8.3



#28
Fluoranthene
Concen: 0.41 ng
RT: 19.124 min Scan# 1828
Delta R.T. -0.005 min
Lab File: BN014029.D
Acq: 23 Mar 2021 02:04

Tgt Ion:202 Resp: 32469
Ion Ratio Lower Upper
202 100
101 11.0 8.1 12.1
203 17.1 13.9 20.9









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.239 min Scan# 2122
Delta R.T. -0.011 min
Lab File: BN014029.D
Acq: 23 Mar 2021 02:04




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:240 Resp: 25745
Ion Ratio Lower Upper
240 100
120 8.4 10.2 15.4#
236 25.9 21.3 31.9



#30
Pyrene
Concen: 0.41 ng
RT: 19.486 min Scan# 1901
Delta R.T. -0.005 min
Lab File: BN014029.D
Acq: 23 Mar 2021 02:04

Tgt Ion:202 Resp: 32643
Ion Ratio Lower Upper
202 100
200 20.4 16.3 24.5
203 17.6 14.0 21.0









#31
Terphenyl-d14
Concen: 0.42 ng
RT: 19.695 min Scan# 1943
Delta R.T. -0.005 min
Lab File: BN014029.D
Acq: 23 Mar 2021 02:04




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:244	Resp:	23995
Ion Ratio	Lower	Upper
244	100	
212	7.4	6.1 9.1
122	10.4	9.8 14.8



#32
Benzo(a)anthracene
Concen: 0.39 ng
RT: 21.218 min Scan# 2120
Delta R.T. -0.011 min
Lab File: BN014029.D
Acq: 23 Mar 2021 02:04

Tgt Ion:228	Resp:	28570
Ion Ratio	Lower	Upper
228	100	
226	27.6	20.5 30.7
229	19.2	16.2 24.4









#33
Chrysene
Concen: 0.42 ng
RT: 21.271 min Scan# 2125
Delta R.T. -0.011 min
Lab File: BN014029.D
Acq: 23 Mar 2021 02:04




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	228	Resp:	33603
Ion Ratio	Lower	Upper	
228	100		
226	30.3	24.3	36.5
229	19.2	15.1	22.7



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.32 ng
RT: 21.154 min Scan# 2114
Delta R.T. -0.011 min
Lab File: BN014029.D
Acq: 23 Mar 2021 02:04

Tgt Ion:	149	Resp:	8622
Ion Ratio	Lower	Upper	
149	100		
167	27.1	22.1	33.1
279	3.9	3.5	5.3









#35
Indeno(1,2,3-cd)pyrene
Concen: 0.46 ng
RT: 25.656 min Scan# 3181
Delta R.T. -0.017 min
Lab File: BN014029.D
Acq: 23 Mar 2021 02:04




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:276 Resp: 39471
Ion Ratio Lower Upper
276 100
138 23.5 19.8 29.6
277 24.9 19.9 29.9



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.448 min Scan# 2536
Delta R.T. -0.010 min
Lab File: BN014029.D
Acq: 23 Mar 2021 02:04

Tgt Ion:264 Resp: 24848
Ion Ratio Lower Upper
264 100
260 24.4 19.3 28.9
265 33.4 66.2 99.4#









#37
Benzo(b)fluoranthene
Concen: 0.41 ng
RT: 22.788 min Scan# 2343
Delta R.T. -0.007 min
Lab File: BN014029.D
Acq: 23 Mar 2021 02:04




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:252 Resp: 36765
Ion Ratio Lower Upper
252 100
253 24.9 20.4 30.6
125 17.1 11.6 17.4



#38
Benzo(k)fluoranthene
Concen: 0.41 ng
RT: 22.788 min Scan# 2343
Delta R.T. -0.051 min
Lab File: BN014029.D
Acq: 23 Mar 2021 02:04

Tgt Ion:252 Resp: 36765
Ion Ratio Lower Upper
252 100
253 24.9 20.5 30.7
125 17.1 12.3 18.5





#39

Benzo(a)pyrene

Concen: 0.42 ng

RT: 23.349 min Scan# 2507

Delta R.T. -0.014 min

Lab File: BN014029.D

Acq: 23 Mar 2021 02:04





Tgt Ion:252	Resp:	32620
Ion Ratio	Lower	Upper
252	100	
253	27.1	21.8 32.6
125	25.5	14.6 21.8#

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4



#40

Dibenzo(a,h)anthracene

Concen: 0.40 ng




RT: 25.670 min Scan# 3185


Delta R.T. -0.017 min

Lab File: BN014029.D

Acq: 23 Mar 2021 02:04

Tgt Ion:278	Resp:	33027
Ion Ratio	Lower	Upper
278	100	
139	16.7	13.1 19.7
279	25.4	21.4 32.0





#41
Benzo(g,h,i)perylene
Concen: 0.40 ng
RT: 26.334 min Scan# 3379
Delta R.T. -0.017 min
Lab File: BN014029.D
Acq: 23 Mar 2021 02:04

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:276 Resp: 34495
Ion Ratio Lower Upper
276 100
277 24.1 20.0 30.0
138 21.2 17.8 26.8

