

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032221\
 Data File : BN014013.D
 Acq On : 22 Mar 2021 15:19
 Operator : CG/JU
 Sample : M1736-02
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE120D1-20210317

Quant Time: Mar 22 16:13:30 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 19 14:40:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.700	152	4988	0.40	ng	0.00
7) Naphthalene-d8	10.480	136	18598	0.40	ng	0.00
13) Acenaphthene-d10	14.334	164	9979	0.40	ng	0.00
19) Phenanthrene-d10	17.067	188	22571	0.40	ng	0.00
29) Chrysene-d12	21.247	240	22028	0.40	ng	# 0.00
36) Perylene-d12	23.463	264	20708	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.300	112	1513	0.13	ng	0.00
5) Phenol-d6	6.871	99	1059	0.07	ng	0.00
8) Nitrobenzene-d5	8.845	82	3941	0.32	ng	0.00
11) 2-Methylnaphthalene-d10	12.066	152	9707	0.33	ng	0.00
14) 2,4,6-Tribromophenol	15.818	330	1011	0.31	ng	0.00
15) 2-Fluorobiphenyl	12.951	172	13667	0.37	ng	0.00
27) Fluoranthene-d10	19.101	212	23902	0.38	ng	0.00
31) Terphenyl-d14	19.697	244	20903	0.43	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.239	88	45374	6.90	ng	97
25) Phenanthrene	17.104	178	1565	0.03	ng	96
34) Bis(2-ethylhexyl)phtha...	21.162	149	1205	0.05	ng	# 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032221\
Data File : BN014013.D
Acq On : 22 Mar 2021 15:19
Operator : CG/JU
Sample : M1736-02
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE120D1-20210317

Quant Time: Mar 22 16:13:30 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 19 14:40:09 2021
Response via : Initial Calibration

