

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032221\
 Data File : BN014014.D
 Acq On : 22 Mar 2021 15:55
 Operator : CG/JU
 Sample : M1736-03
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE120D3-20210317

Quant Time: Mar 22 16:35:50 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 19 14:40:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.700	152	5282	0.40	ng	0.00
7) Naphthalene-d8	10.479	136	20090	0.40	ng	0.00
13) Acenaphthene-d10	14.321	164	11178	0.40	ng	-0.01
19) Phenanthrene-d10	17.067	188	25540	0.40	ng	# 0.00
29) Chrysene-d12	21.250	240	25975	0.40	ng	# 0.00
36) Perylene-d12	23.458	264	23257	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.300	112	2006	0.17	ng	0.00
5) Phenol-d6	6.871	99	1439	0.09	ng	0.00
8) Nitrobenzene-d5	8.845	82	4998	0.38	ng	0.00
11) 2-Methylnaphthalene-d10	12.066	152	12302	0.39	ng	0.00
14) 2,4,6-Tribromophenol	15.818	330	1366	0.37	ng	0.00
15) 2-Fluorobiphenyl	12.950	172	18064	0.43	ng	0.00
27) Fluoranthene-d10	19.099	212	30747	0.43	ng	0.00
31) Terphenyl-d14	19.700	244	30556	0.54	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.239	88	1188	0.17	ng	# 65
34) Bis(2-ethylhexyl)phtha...	21.165	149	1989	0.07	ng	# 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032221\
Data File : BN014014.D
Acq On : 22 Mar 2021 15:55
Operator : CG/JU
Sample : M1736-03
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE120D3-20210317

Quant Time: Mar 22 16:35:50 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 19 14:40:09 2021
Response via : Initial Calibration

