

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032221\
 Data File : BN014015.D
 Acq On : 22 Mar 2021 16:31
 Operator : CG/JU
 Sample : M1736-04
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE120D2-20210317

Quant Time: Mar 22 17:32:39 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 19 14:40:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.700	152	5211	0.40	ng	0.00
7) Naphthalene-d8	10.480	136	19599	0.40	ng	0.00
13) Acenaphthene-d10	14.321	164	10891	0.40	ng	-0.01
19) Phenanthrene-d10	17.067	188	25021	0.40	ng	0.00
29) Chrysene-d12	21.247	240	25083	0.40	ng	# 0.00
36) Perylene-d12	23.462	264	22936	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.300	112	1472	0.12	ng	0.00
5) Phenol-d6	6.871	99	1065	0.07	ng	0.00
8) Nitrobenzene-d5	8.845	82	3519	0.27	ng	0.00
11) 2-Methylnaphthalene-d10	12.066	152	8543	0.28	ng	0.00
14) 2,4,6-Tribromophenol	15.818	330	1132	0.32	ng	0.00
15) 2-Fluorobiphenyl	12.950	172	11849	0.29	ng	0.00
27) Fluoranthene-d10	19.096	212	28644	0.41	ng	0.00
31) Terphenyl-d14	19.697	244	28078	0.51	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.239	88	28857	4.20	ng	Qvalue 97
25) Phenanthrene	17.103	178	1707	0.02	ng	98
28) Fluoranthene	19.131	202	1528	0.02	ng	# 93
34) Bis(2-ethylhexyl)phtha...	21.162	149	1812	0.07	ng	100
35) Indeno(1,2,3-cd)pyrene	25.674	276	1714	0.02	ng	# 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032221\
Data File : BN014015.D
Acq On : 22 Mar 2021 16:31
Operator : CG/JU
Sample : M1736-04
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE120D2-20210317

Quant Time: Mar 22 17:32:39 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 19 14:40:09 2021
Response via : Initial Calibration

