

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032221\
 Data File : BN014021.D
 Acq On : 22 Mar 2021 20:06
 Operator : CG/JU
 Sample : M1735-06
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE134D3-20210315

Quant Time: Mar 23 00:21:06 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 19 14:40:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.700	152	5296	0.40	ng	0.00
7) Naphthalene-d8	10.480	136	20106	0.40	ng	0.00
13) Acenaphthene-d10	14.321	164	11155	0.40	ng	-0.01
19) Phenanthrene-d10	17.068	188	24965	0.40	ng	# 0.00
29) Chrysene-d12	21.250	240	25039	0.40	ng	# 0.00
36) Perylene-d12	23.458	264	23766	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.300	112	1595	0.13	ng	0.00
5) Phenol-d6	6.871	99	1162	0.08	ng	0.00
8) Nitrobenzene-d5	8.845	82	3672	0.28	ng	0.00
11) 2-Methylnaphthalene-d10	12.066	152	8318	0.27	ng	0.00
14) 2,4,6-Tribromophenol	15.818	330	888	0.24	ng	0.00
15) 2-Fluorobiphenyl	12.951	172	11126	0.27	ng	0.00
27) Fluoranthene-d10	19.099	212	25894	0.37	ng	0.00
31) Terphenyl-d14	19.700	244	21472	0.39	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.239	88	36467	5.22	ng	96
34) Bis(2-ethylhexyl)phtha...	21.165	149	2047	0.08	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032221\
Data File : BN014021.D
Acq On : 22 Mar 2021 20:06
Operator : CG/JU
Sample : M1735-06
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE134D3-20210315

Quant Time: Mar 23 00:21:06 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 19 14:40:09 2021
Response via : Initial Calibration

