

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032221\
 Data File : BN014022.D
 Acq On : 22 Mar 2021 20:42
 Operator : CG/JU
 Sample : M1735-02
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE134D1-20210315

Quant Time: Mar 23 00:21:13 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 19 14:40:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.700	152	5468	0.40	ng	0.00
7) Naphthalene-d8	10.479	136	20961	0.40	ng	# 0.00
13) Acenaphthene-d10	14.321	164	12212	0.40	ng	-0.01
19) Phenanthrene-d10	17.067	188	29290	0.40	ng	# 0.00
29) Chrysene-d12	21.239	240	30095	0.40	ng	-0.01
36) Perylene-d12	23.451	264	27146	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.300	112	1379	0.11	ng	0.00
5) Phenol-d6	6.863	99	992	0.06	ng	0.00
8) Nitrobenzene-d5	8.845	82	3557	0.26	ng	0.00
11) 2-Methylnaphthalene-d10	12.066	152	8554	0.26	ng	0.00
14) 2,4,6-Tribromophenol	15.818	330	1298	0.33	ng	0.00
15) 2-Fluorobiphenyl	12.950	172	11753	0.26	ng	0.00
27) Fluoranthene-d10	19.094	212	29815	0.37	ng	0.00
31) Terphenyl-d14	19.695	244	31016	0.47	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.239	88	2885	0.40	ng	# 77
12) 2-Methylnaphthalene	12.137	142	1421	0.04	ng	# 92
25) Phenanthrene	17.104	178	2162	0.03	ng	# 80
30) Pyrene	19.486	202	3560	0.04	ng	# 92
34) Bis(2-ethylhexyl)phtha...	21.165	149	2380	0.07	ng	# 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032221\
Data File : BN014022.D
Acq On : 22 Mar 2021 20:42
Operator : CG/JU
Sample : M1735-02
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE134D1-20210315

Quant Time: Mar 23 00:21:13 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 19 14:40:09 2021
Response via : Initial Calibration

