




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032221\
Data File : BN014009.D
Acq On : 22 Mar 2021 12:56
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 22 13:30:17 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 05 13:12:06 2021
Response via : Initial Calibration



TIC: BN014009.D\data.ms

(77) Benzidine




19.316min (-0.029) 0.00 ng

response 1943679

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.04
183.00	13.20	11.48
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032221\
 Data File : BN014009.D
 Acq On : 22 Mar 2021 12:56
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 22 13:30:17 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 05 13:12:06 2021
 Response via : Initial Calibration



TIC: BN014009.D\data.ms

(70) Pentachlorophenol (C)

16.722min (-0.029) 22536.71 ng

response 259634

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	61.57
264.00	61.60	59.59
0.00	0.00	0.00