

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032221\
 Data File : BN014006.D
 Acq On : 22 Mar 2021 10:18
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 22 10:50:53 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 19 14:40:09 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	97	0.00
2	1,4-Dioxane	0.400	0.392	2.0	93	0.00
3	n-Nitrosodimethylamine	0.400	0.319	20.3	87	0.00
4 S	2-Fluorophenol	0.400	0.424	-6.0	97	0.00
5 S	Phenol-d6	0.400	0.420	-5.0	96	0.00
6	bis(2-Chloroethyl)ether	0.400	0.424	-6.0	97	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	100	0.00
8 S	Nitrobenzene-d5	0.400	0.377	5.8	93	0.00
9	Naphthalene	0.400	0.414	-3.5	100	0.01
10	Hexachlorobutadiene	0.400	0.416	-4.0	100	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.411	-2.7	100	0.00
12	2-Methylnaphthalene	0.400	0.410	-2.5	99	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	101	0.00
14 S	2,4,6-Tribromophenol	0.400	0.329	17.8	86	0.00
15 S	2-Fluorobiphenyl	0.400	0.410	-2.5	99	0.00
16	Acenaphthylene	0.400	0.385	3.8	100	0.00
17	Acenaphthene	0.400	0.402	-0.5	99	0.00
18	Fluorene	0.400	0.409	-2.2	100	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	100	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.338	15.5	93	0.00
21	4-Bromophenyl-phenylether	0.400	0.393	1.8	95	0.00
22	Hexachlorobenzene	0.400	0.419	-4.7	101	0.00
23	Atrazine	0.400	0.362	9.5	94	0.00
24	Pentachlorophenol	0.400	0.294	26.5#	84	0.01
25	Phenanthrene	0.400	0.410	-2.5	100	0.00
26	Anthracene	0.400	0.394	1.5	100	0.01
27 SURR	Fluoranthene-d10	0.400	0.414	-3.5	103	0.01
28	Fluoranthene	0.400	0.411	-2.7	103	0.01
29 I	Chrysene-d12	0.400	0.400	0.0	104	0.01
30	Pyrene	0.400	0.410	-2.5	102	0.00
31 S	Terphenyl-d14	0.400	0.422	-5.5	104	0.01
32	Benzo(a)anthracene	0.400	0.390	2.5	98	0.01
33	Chrysene	0.400	0.414	-3.5	103	0.01
34	Bis(2-ethylhexyl)phthalate	0.400	0.312	22.0	85	0.01
35	Indeno(1,2,3-cd)pyrene	0.400	0.449	-12.2	112	0.02
36 I	Perylene-d12	0.400	0.400	0.0	112	0.01
37	Benzo(b)fluoranthene	0.400	0.409	-2.2	112	0.01
38	Benzo(k)fluoranthene	0.400	0.400	0.0	108	0.01
39 C	Benzo(a)pyrene	0.400	0.418	-4.5	115	0.01
40	Dibenzo(a,h)anthracene	0.400	0.407	-1.7	111	0.02
41	Benzo(g,h,i)perylene	0.400	0.409	-2.2	112	0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4