

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032221\
 Data File : BN014034.D
 Acq On : 23 Mar 2021 09:12
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 23 10:06:14 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 19 14:40:09 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	137	0.00
2	1,4-Dioxane	0.400	0.465	-16.3	157	0.00
3	n-Nitrosodimethylamine	0.400	0.459	-14.8	177	-0.03
4 S	2-Fluorophenol	0.400	0.428	-7.0	138	0.00
5 S	Phenol-d6	0.400	0.426	-6.5	139	0.00
6	bis(2-Chloroethyl)ether	0.400	0.434	-8.5	141	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	140	-0.01
8 S	Nitrobenzene-d5	0.400	0.376	6.0	130	-0.01
9	Naphthalene	0.400	0.413	-3.2	139	0.00
10	Hexachlorobutadiene	0.400	0.409	-2.2	137	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.411	-2.7	140	0.00
12	2-Methylnaphthalene	0.400	0.411	-2.7	140	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	138	0.00
14 S	2,4,6-Tribromophenol	0.400	0.395	1.3	142	-0.01
15 S	2-Fluorobiphenyl	0.400	0.419	-4.7	139	-0.01
16	Acenaphthylene	0.400	0.391	2.3	139	0.00
17	Acenaphthene	0.400	0.415	-3.7	140	0.00
18	Fluorene	0.400	0.407	-1.7	137	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	136	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.376	6.0	140	0.00
21	4-Bromophenyl-phenylether	0.400	0.404	-1.0	133	0.00
22	Hexachlorobenzene	0.400	0.406	-1.5	134	0.00
23	Atrazine	0.400	0.364	9.0	129	0.00
24	Pentachlorophenol	0.400	0.392	2.0	153	-0.01
25	Phenanthrene	0.400	0.414	-3.5	138	0.00
26	Anthracene	0.400	0.397	0.8	137	0.00
27 SURR	Fluoranthene-d10	0.400	0.407	-1.7	138	0.00
28	Fluoranthene	0.400	0.407	-1.7	138	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	137	0.00
30	Pyrene	0.400	0.414	-3.5	136	0.00
31 S	Terphenyl-d14	0.400	0.427	-6.7	139	0.00
32	Benzo(a)anthracene	0.400	0.392	2.0	130	0.01
33	Chrysene	0.400	0.429	-7.2	141	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.338	15.5	122	0.01
35	Indeno(1,2,3-cd)pyrene	0.400	0.455	-13.7	150	0.00
36 I	Perylene-d12	0.400	0.400	0.0	147	0.00
37	Benzo(b)fluoranthene	0.400	0.404	-1.0	146	0.00
38	Benzo(k)fluoranthene	0.400	0.415	-3.7	148	0.00
39 C	Benzo(a)pyrene	0.400	0.425	-6.2	154	0.00
40	Dibenzo(a,h)anthracene	0.400	0.411	-2.7	149	0.00
41	Benzo(g,h,i)perylene	0.400	0.407	-1.7	147	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4