

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN032420\
 Data File : BN010329.D
 Acq On : 24 Mar 2020 16:52
 Operator : CG/JU
 Sample : L1997-04
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MW203D1-20200315

Quant Time: Mar 24 17:23:27 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN031920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 19 15:26:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.62	152	4918	0.40	ng	0.00
7) Naphthalene-d8	10.38	136	17160	0.40	ng	0.00
13) Acenaphthene-d10	14.24	164	9959	0.40	ng	-0.01
19) Phenanthrene-d10	16.99	188	23634	0.40	ng	0.00
27) Chrysene-d12	21.19	240	23598	0.40	ng	0.00
34) Perylene-d12	23.35	264	24095	0.40	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.25	112	1219	0.13	ng	0.00
5) Phenol-d6	6.80	99	907	0.07	ng	0.00
8) Nitrobenzene-d5	8.74	82	6144	0.56	ng	0.00
11) 2-Methylnaphthalene-d10	11.97	152	8057	0.29	ng	0.00
14) 2,4,6-Tribromophenol	15.73	330	1108	0.29	ng	0.00
15) 2-Fluorobiphenyl	12.85	172	23258	0.61	ng	-0.01
25) Fluoranthene-d10	19.02	212	22008	0.31	ng	0.00
29) Terphenyl-d14	19.63	244	28409	0.56	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.25	88	6877	1.489	ng	97
32) Bis(2-ethylhexyl)phthalate	21.13	149	1756	0.069	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN032420\
Data File : BN010329.D
Acq On : 24 Mar 2020 16:52
Operator : CG/JU
Sample : L1997-04
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW203D1-20200315

Quant Time: Mar 24 17:23:27 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN031920.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 19 15:26:45 2020
Response via : Initial Calibration

