

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN032420\
 Data File : BN010333.D
 Acq On : 24 Mar 2020 19:16
 Operator : CG/JU
 Sample : L1997-08
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE109D1-20200316

Quant Time: Mar 25 01:50:14 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN031920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 19 15:26:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.62	152	4831	0.40	ng	0.00
7) Naphthalene-d8	10.38	136	16750	0.40	ng	0.00
13) Acenaphthene-d10	14.24	164	9860	0.40	ng	-0.01
19) Phenanthrene-d10	16.99	188	23293	0.40	ng	0.00
27) Chrysene-d12	21.19	240	24049	0.40	ng	0.00
34) Perylene-d12	23.35	264	24445	0.40	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.25	112	1354	0.14	ng	0.00
5) Phenol-d6	6.80	99	1014	0.09	ng	0.00
8) Nitrobenzene-d5	8.74	82	5920	0.55	ng	0.00
11) 2-Methylnaphthalene-d10	11.97	152	7864	0.29	ng	0.00
14) 2,4,6-Tribromophenol	15.73	330	1217	0.32	ng	0.00
15) 2-Fluorobiphenyl	12.85	172	22151	0.59	ng	-0.01
25) Fluoranthene-d10	19.02	212	21360	0.30	ng	0.00
29) Terphenyl-d14	19.63	244	27775	0.53	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.25	88	11035	2.432	ng	98
32) Bis(2-ethylhexyl)phthalate	21.12	149	1232	0.048	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN032420\
Data File : BN010333.D
Acq On : 24 Mar 2020 19:16
Operator : CG/JU
Sample : L1997-08
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE109D1-20200316

Quant Time: Mar 25 01:50:14 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN031920.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 19 15:26:45 2020
Response via : Initial Calibration

