

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN032420\
 Data File : BN010337.D
 Acq On : 24 Mar 2020 21:40
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 25 04:13:25 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN031920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 19 15:26:45 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.376	0.362	3.7	94	0.00
3	n-Nitrosodimethylamine	0.317	0.297	6.3	99	0.02
4 S	2-Fluorophenol	0.792	0.796	-0.5	101	0.00
5 S	Phenol-d6	0.985	0.998	-1.3	104	0.00
6	bis(2-Chloroethyl)ether	0.946	0.961	-1.6	100	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
8 S	Nitrobenzene-d5	0.256	0.248	3.1	101	0.00
9	Naphthalene	0.997	1.015	-1.8	101	-0.01
10	Hexachlorobutadiene	0.242	0.251	-3.7	104	0.00
11 SURR	2-Methylnaphthalene-d10	0.639	0.714	-11.7	114	0.00
12	2-Methylnaphthalene	0.692	0.693	-0.1	102	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	97	-0.01
14 S	2,4,6-Tribromophenol	0.154	0.136	11.7	96	0.00
15 S	2-Fluorobiphenyl	1.531	1.567	-2.4	98	-0.01
16	Acenaphthylene	1.569	1.477	5.9	98	-0.01
17	Acenaphthene	1.118	1.149	-2.8	101	0.00
18	Fluorene	1.485	1.480	0.3	100	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	97	0.00
20	4-Bromophenyl-phenylether	0.243	0.232	4.5	95	0.00
21	Hexachlorobenzene	0.249	0.267	-7.2	105	0.00
22	Pentachlorophenol	0.085	0.070	17.6	93	0.00
23	Phenanthrene	1.099	1.105	-0.5	99	-0.01
24	Anthracene	0.940	0.876	6.8	96	0.00
25 SURR	Fluoranthene-d10	1.206	1.310	-8.6	111	0.00
26	Fluoranthene	1.347	1.329	1.3	100	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	94	0.00
28	Pvrene	1.245	1.293	-3.9	98	0.00
29 S	Terphenyl-d14	0.864	0.887	-2.7	96	0.00
30	Benzo(a)anthracene	1.274	1.202	5.7	92	-0.01
31	Chrysene	1.366	1.386	-1.5	95	-0.01
32	Bis(2-ethylhexyl)phthalate	0.430	0.321	25.3#	78	-0.01
33	Indeno(1,2,3-cd)pyrene	1.463	1.501	-2.6	89	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	86	-0.01
35	Benzo(b)fluoranthene	1.359	1.373	-1.0	96	-0.01
36	Benzo(k)fluoranthene	1.383	1.460	-5.6	93	-0.01
37 C	Benzo(a)pyrene	1.147	1.122	2.2	93	-0.01
38	Dibenzo(a,h)anthracene	1.240	1.271	-2.5	88	-0.02
39	Benzo(a,h,i)perylene	1.248	1.297	-3.9	95	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN032420\
Data File : BN010337.D
Acq On : 24 Mar 2020 21:40
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 25 04:13:25 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN031920.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 19 15:26:45 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			