

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN032519\
 Data File : BN004954.D
 Acq On : 25 Mar 2019 11:14
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 26 01:27:11 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 20 16:06:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	254	-0.02
2	1,4-Dioxane	0.400	0.444	-11.0	279	-0.02
3	n-Nitrosodimethylamine	0.400	0.535	-33.8#	366	-0.02
4 S	2-Fluorophenol	0.400	0.470	-17.5	293	0.00
5 S	Phenol-d6	0.400	0.460	-15.0	294	0.00
6	bis(2-Chloroethyl)ether	0.400	0.447	-11.7	291	-0.02
7 I	Naphthalene-d8	0.400	0.400	0.0	236	-0.03
8 S	Nitrobenzene-d5	0.400	0.490	-22.5	292	-0.03
9	Naphthalene	0.400	0.432	-8.0	251	-0.03
10	Hexachlorobutadiene	0.400	0.419	-4.7	245	-0.03
11 SURR	2-Methylnaphthalene-d10	0.400	0.391	2.3	231	-0.02
12	2-Methylnaphthalene	0.400	0.393	1.8	230	-0.02
13 I	Acenaphthene-d10	0.400	0.400	0.0	182	-0.02
14 S	2,4,6-Tribromophenol	0.400	0.371	7.3	179	-0.01
15 S	2-Fluorobiphenyl	0.400	0.485	-21.2	217	-0.02
16	Acenaphthylene	0.400	0.458	-14.5	219	-0.02
17	Acenaphthene	0.400	0.432	-8.0	194	-0.01
18	Fluorene	0.400	0.392	2.0	175	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	147	-0.01
20	4-Bromophenyl-phenylether	0.400	0.445	-11.2	165	-0.01
21	Hexachlorobenzene	0.400	0.407	-1.7	148	-0.01
22	Pentachlorophenol	0.400	0.421	-5.2	179	-0.01
23	Phenanthrene	0.400	0.416	-4.0	151	-0.01
24	Anthracene	0.400	0.426	-6.5	160	-0.02
25 SURR	Fluoranthene-d10	0.400	0.381	4.8	141	-0.01
26	Fluoranthene	0.400	0.378	5.5	138	-0.01
27 I	Chrysene-d12	0.400	0.400	0.0	120	-0.01
28	Pvrene	0.400	0.465	-16.3	137	-0.01
29 S	Terphenyl-d14	0.400	0.485	-21.2	143	-0.01
30	Benzo(a)anthracene	0.400	0.436	-9.0	132	-0.01
31	Chrysene	0.400	0.420	-5.0	121	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.729	-82.2#	237	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.382	4.5	110	-0.02
34 I	Perylene-d12	0.400	0.400	0.0	117	-0.02
35	Benzo(b)fluoranthene	0.400	0.412	-3.0	118	-0.01
36	Benzo(k)fluoranthene	0.400	0.399	0.3	118	-0.01
37 C	Benzo(a)pyrene	0.400	0.400	0.0	117	-0.01
38	Dibenzo(a,h)anthracene	0.400	0.378	5.5	109	-0.02
39	Benzo(a,h,i)perylene	0.400	0.373	6.8	107	-0.03

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN032519\
Data File : BN004954.D
Acq On : 25 Mar 2019 11:14
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 26 01:27:11 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 20 16:06:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				