

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN032519\
 Data File : BN004954.D
 Acq On : 25 Mar 2019 11:14
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 26 01:27:11 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 20 16:06:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	254#	-0.02
2	1,4-Dioxane	0.455	0.505	-11.0	279#	-0.02
3	n-Nitrosodimethylamine	0.269	0.359	-33.5#	366#	-0.02
4 S	2-Fluorophenol	1.087	1.277	-17.5	293#	0.00
5 S	Phenol-d6	1.354	1.555	-14.8	294#	0.00
6	bis(2-Chloroethyl)ether	1.221	1.365	-11.8	291#	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	236#	-0.03
8 S	Nitrobenzene-d5	0.265	0.324	-22.3	292#	-0.03
9	Naphthalene	1.122	1.211	-7.9	251#	-0.03
10	Hexachlorobutadiene	0.214	0.224	-4.7	245#	-0.03
11 SURR	2-Methylnaphthalene-d10	0.698	0.682	2.3	231#	-0.02
12	2-Methylnaphthalene	0.833	0.819	1.7	230#	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	182#	-0.02
14 S	2,4,6-Tribromophenol	0.276	0.256	7.2	179#	-0.01
15 S	2-Fluorobiphenyl	1.653	2.006	-21.4	217#	-0.02
16	Acenaphthylene	2.060	2.360	-14.6	219#	-0.02
17	Acenaphthene	1.349	1.457	-8.0	194#	-0.01
18	Fluorene	1.883	1.845	2.0	175#	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	147	-0.01
20	4-Bromophenyl-phenylether	0.261	0.290	-11.1	165#	-0.01
21	Hexachlorobenzene	0.299	0.304	-1.7	148	-0.01
22	Pentachlorophenol	0.058	0.062	-6.9	179#	-0.01
23	Phenanthrene	1.292	1.344	-4.0	151#	-0.01
24	Anthracene	1.154	1.228	-6.4	160#	-0.02
25 SURR	Fluoranthene-d10	1.339	1.277	4.6	141	-0.01
26	Fluoranthene	1.660	1.567	5.6	138	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	120	-0.01
28	Pvrene	1.220	1.417	-16.1	137	-0.01
29 S	Terphenyl-d14	0.819	0.992	-21.1	143	-0.01
30	Benzo(a)anthracene	1.312	1.430	-9.0	132	-0.01
31	Chrysene	1.382	1.449	-4.8	121	-0.01
32	Bis(2-ethylhexyl)phthalate	0.524	0.956	-82.4#	237#	-0.01
33	Indeno(1,2,3-cd)pyrene	1.999	1.908	4.6	110	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	117	-0.02
35	Benzo(b)fluoranthene	1.447	1.490	-3.0	118	-0.01
36	Benzo(k)fluoranthene	1.424	1.420	0.3	118	-0.01
37 C	Benzo(a)pyrene	1.327	1.328	-0.1	117	-0.01
38	Dibenzo(a,h)anthracene	1.537	1.454	5.4	109	-0.02
39	Benzo(a,h,i)perylene	1.566	1.459	6.8	107	-0.03

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN032519\
Data File : BN004954.D
Acq On : 25 Mar 2019 11:14
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 26 01:27:11 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 20 16:06:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			