

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN032519\
 Data File : BN004970.D
 Acq On : 25 Mar 2019 20:39
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDC00.4

Quant Time: Mar 26 02:07:08 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 20 16:06:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	305#	0.00
2	1,4-Dioxane	0.455	0.464	-2.0	308#	-0.02
3	n-Nitrosodimethylamine	0.269	0.316	-17.5	387#	-0.02
4 S	2-Fluorophenol	1.087	1.308	-20.3	360#	0.00
5 S	Phenol-d6	1.354	1.650	-21.9	374#	0.00
6	bis(2-Chloroethyl)ether	1.221	1.326	-8.6	340#	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	297#	-0.03
8 S	Nitrobenzene-d5	0.265	0.332	-25.3#	376#	-0.03
9	Naphthalene	1.122	1.210	-7.8	316#	-0.03
10	Hexachlorobutadiene	0.214	0.226	-5.6	311#	-0.03
11 SURR	2-Methylnaphthalene-d10	0.698	0.691	1.0	294#	-0.01
12	2-Methylnaphthalene	0.833	0.830	0.4	293#	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	233#	-0.01
14 S	2,4,6-Tribromophenol	0.276	0.258	6.5	232#	-0.01
15 S	2-Fluorobiphenyl	1.653	1.981	-19.8	275#	-0.01
16	Acenaphthylene	2.060	2.431	-18.0	289#	-0.01
17	Acenaphthene	1.349	1.463	-8.5	250#	-0.01
18	Fluorene	1.883	1.828	2.9	222#	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	175#	-0.01
20	4-Bromophenyl-phenylether	0.261	0.312	-19.5	211#	-0.01
21	Hexachlorobenzene	0.299	0.314	-5.0	182#	0.00
22	Pentachlorophenol	0.058	0.069	-19.0	238#	-0.01
23	Phenanthrene	1.292	1.329	-2.9	179#	-0.01
24	Anthracene	1.154	1.311	-13.6	203#	-0.01
25 SURR	Fluoranthene-d10	1.339	1.188	11.3	157#	-0.01
26	Fluoranthene	1.660	1.462	11.9	153#	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	124	0.00
28	Pvrene	1.220	1.495	-22.5	149	-0.01
29 S	Terphenyl-d14	0.819	1.034	-26.3#	155#	-0.01
30	Benzo(a)anthracene	1.312	1.466	-11.7	140	-0.01
31	Chrysene	1.382	1.420	-2.7	123	-0.01
32	Bis(2-ethylhexyl)phthalate	0.524	1.027	-96.0#	263#	0.00
33	Indeno(1,2,3-cd)pyrene	1.999	1.982	0.9	118	0.00
34 I	Perylene-d12	1.000	1.000	0.0	125	-0.01
35	Benzo(b)fluoranthene	1.447	1.479	-2.2	125	0.00
36	Benzo(k)fluoranthene	1.424	1.381	3.0	123	0.00
37 C	Benzo(a)pyrene	1.327	1.342	-1.1	127	-0.01
38	Dibenzo(a,h)anthracene	1.537	1.468	4.5	118	-0.01
39	Benzo(a,h,i)perylene	1.566	1.485	5.2	116	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN032519\
Data File : BN004970.D
Acq On : 25 Mar 2019 20:39
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 26 02:07:08 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 20 16:06:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			