

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN032619\
 Data File : BN004988.D
 Acq On : 26 Mar 2019 23:08
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 27 05:47:56 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-BN032619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 27 05:27:09 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
2	1,4-Dioxane	0.422	0.430	-1.9	99	0.00
3	Pyridine	1.178	1.174	0.3	98	0.00
4	n-Nitrosodimethylamine	0.384	0.398	-3.6	98	0.00
5 S	2-Fluorophenol	1.157	1.180	-2.0	99	0.00
6	Aniline	1.954	1.925	1.5	93	0.00
7 S	Phenol-d6	1.534	1.519	1.0	94	0.00
8	2-Chlorophenol	1.469	1.474	-0.3	96	0.00
9	Benzaldehyde	0.882	0.948	-7.5	99	0.00
10 C	Phenol	1.644	1.619	1.5	94	0.00
11	bis(2-Chloroethyl)ether	1.271	1.266	0.4	95	0.00
12	1,3-Dichlorobenzene	1.600	1.626	-1.6	98	0.00
13 C	1,4-Dichlorobenzene	1.620	1.649	-1.8	98	0.00
14	1,2-Dichlorobenzene	1.550	1.571	-1.4	97	0.00
15	Benzyl Alcohol	1.167	1.133	2.9	92	0.00
16	2,2'-oxybis(1-Chloropropane	1.534	1.527	0.5	95	0.00
17	2-Methylphenol	1.146	1.132	1.2	93	0.00
18	Hexachloroethane	0.574	0.581	-1.2	97	0.00
19 P	n-Nitroso-di-n-propylamine	1.006	0.973	3.3	90	0.00
20	3+4-Methylphenols	1.556	1.509	3.0	92	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	91	0.00
22	Acetophenone	0.457	0.463	-1.3	92	0.00
23 S	Nitrobenzene-d5	0.331	0.339	-2.4	93	0.00
24	Nitrobenzene	0.335	0.340	-1.5	92	0.00
25	Isophorone	0.644	0.633	1.7	89	0.00
26 C	2-Nitrophenol	0.194	0.193	0.5	90	0.00
27	2,4-Dimethylphenol	0.270	0.270	0.0	91	0.00
28	bis(2-Chloroethoxy)methane	0.399	0.400	-0.3	91	0.00
29 C	2,4-Dichlorophenol	0.311	0.312	-0.3	91	0.00
30	1,2,4-Trichlorobenzene	0.342	0.353	-3.2	94	0.00
31	Naphthalene	1.043	1.057	-1.3	92	0.00
32	Benzoic acid	0.217	0.192	11.5	81	0.00
33	4-Chloroaniline	0.425	0.419	1.4	89	0.00
34 C	Hexachlorobutadiene	0.222	0.229	-3.2	95	0.00
35	Caprolactam	0.102	0.096	5.9	81	0.00
36 C	4-Chloro-3-methylphenol	0.336	0.326	3.0	86	0.00
37	2-Methylnaphthalene	0.737	0.736	0.1	90	0.00
38	1-Methylnaphthalene	0.721	0.713	1.1	89	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	84	0.00
40	1,2,4,5-Tetrachlorobenzene	0.709	0.742	-4.7	89	0.00
41 P	Hexachlorocyclopentadiene	0.381	0.400	-5.0	92	0.00
42 S	2,4,6-Tribromophenol	0.323	0.314	2.8	81	0.00
43 C	2,4,6-Trichlorophenol	0.440	0.455	-3.4	87	0.00
44	2,4,5-Trichlorophenol	0.480	0.485	-1.0	84	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN032619\
 Data File : BN004988.D
 Acq On : 26 Mar 2019 23:08
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 27 05:47:56 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-BN032619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 27 05:27:09 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.509	1.577	-4.5	88	0.00
46	1,1'-Biphenyl	1.693	1.771	-4.6	88	0.00
47	2-Chloronaphthalene	1.299	1.351	-4.0	89	0.00
48	2-Nitroaniline	0.346	0.351	-1.4	85	0.00
49	Acenaphthylene	2.185	2.242	-2.6	86	0.00
50	Dimethylphthalate	1.643	1.644	-0.1	84	0.00
51	2,6-Dinitrotoluene	0.374	0.375	-0.3	84	0.00
52 C	Acenaphthene	1.229	1.258	-2.4	86	0.00
53	3-Nitroaniline	0.382	0.377	1.3	81	0.00
54 P	2,4-Dinitrophenol	0.207	0.189	8.7	78	0.00
55	Dibenzofuran	1.956	1.995	-2.0	86	0.00
56 P	4-Nitrophenol	0.301	0.285	5.3	79	0.00
57	2,4-Dinitrotoluene	0.508	0.502	1.2	82	0.00
58	Fluorene	1.580	1.597	-1.1	84	0.00
59	2,3,4,6-Tetrachlorophenol	0.445	0.443	0.4	83	0.00
60	Diethylphthalate	1.640	1.610	1.8	81	0.00
61	4-Chlorophenyl-phenylether	0.815	0.824	-1.1	85	0.00
62	4-Nitroaniline	0.384	0.375	2.3	79	0.00
63	Azobenzene	1.365	1.371	-0.4	84	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	79	0.00
65	4,6-Dinitro-2-methylphenol	0.134	0.129	3.7	77	0.00
66 c	n-Nitrosodiphenylamine	0.630	0.651	-3.3	83	0.00
67	4-Bromophenyl-phenylether	0.250	0.256	-2.4	82	0.00
68	Hexachlorobenzene	0.292	0.300	-2.7	82	0.00
69	Atrazine	0.225	0.222	1.3	77	0.00
70 C	Pentachlorophenol	0.154	0.148	3.9	78	0.00
71	Phenanthrene	1.124	1.150	-2.3	81	0.00
72	Anthracene	1.119	1.146	-2.4	81	0.00
73	Carbazole	1.018	1.034	-1.6	80	0.00
74	Di-n-butylphthalate	1.240	1.210	2.4	76	0.00
75 C	Fluoranthene	1.282	1.282	0.0	79	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	86	0.00
77	Benzidine	0.623	0.600	3.7	74	0.00
78	Pyrene	1.335	1.282	4.0	80	0.00
79 S	Terphenyl-d14	1.039	1.001	3.7	79	0.00
80	Butylbenzylphthalate	0.543	0.510	6.1	79	0.00
81	Benzo(a)anthracene	1.246	1.254	-0.6	86	0.00
82	3,3'-Dichlorobenzidine	0.466	0.478	-2.6	87	0.00
83	Chrysene	1.190	1.211	-1.8	88	0.00
84	Bis(2-ethylhexyl)phthalate	0.751	0.718	4.4	81	0.00
85 c	Di-n-octyl phthalate	1.212	1.229	-1.4	88	0.00
86	Indeno(1,2,3-cd)pyrene	1.352	1.628	-20.4	113	0.00
87 I	Perylene-d12	1.000	1.000	0.0	101	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN032619\
Data File : BN004988.D
Acq On : 26 Mar 2019 23:08
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC040

Quant Time: Mar 27 05:47:56 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-BN032619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 27 05:27:09 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.231	1.204	2.2	99	0.00
89	Benzo(k)fluoranthene	1.129	1.139	-0.9	99	0.00
90 C	Benzo(a)pyrene	1.127	1.142	-1.3	102	0.00
91	Dibenzo(a,h)anthracene	1.131	1.224	-8.2	113	0.00
92	Benzo(q,h,i)perylene	1.118	1.216	-8.8	114	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0