


Data Path : Z:\SVOASRV\HPCHEM1\BNA\_N\DATA\BN032619\  
 Data File : BN004971.D  
 Acq On : 26 Mar 2019 11:55  
 Operator : JU/SJ  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\8270-BN032619.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Wed Mar 27 05:27:09 2019



AutoFind: Scans 2466, 2467, 2468; Background Corrected with Scan 2460

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 26.7      | 2402    | PASS             |
| 68          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 69          | 198          | 0.00         | 100          | 29.1      | 2610    | PASS             |
| 70          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 127         | 198          | 10           | 80           | 41.3      | 3710    | PASS             |
| 197         | 198          | 0.00         | 2            | 0.0       | 0       | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 8984    | PASS             |
| 199         | 198          | 5            | 9            | 6.6       | 594     | PASS             |
| 275         | 198          | 10           | 60           | 32.0      | 2874    | PASS             |
| 365         | 198          | 1            | 100          | 3.1       | 277     | PASS             |
| 441         | 198          | 0.01         | 100          | 13.7      | 1234    | PASS             |
| 442         | 198          | 50           | 100          | 87.1      | 7827    | PASS             |
| 443         | 442          | 15           | 24           | 18.6      | 1458    | PASS             |