


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032719\
 Data File : BN004989.D
 Acq On : 27 Mar 2019 08:28
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-BN032619.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Mar 27 05:27:09 2019



AutoFind: Scans 2465, 2466, 2467; Background Corrected with Scan 2460

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	26.2	2757	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	30.4	3195	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	43.5	4570	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	10517	PASS
199	198	5	9	6.8	711	PASS
275	198	10	60	31.1	3276	PASS
365	198	1	100	4.1	426	PASS
441	198	0.01	100	12.9	1357	PASS
442	198	50	100	82.5	8672	PASS
443	442	15	24	19.0	1646	PASS