

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN032720\
 Data File : BN010356.D
 Acq On : 27 Mar 2020 17:45
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 27 18:20:34 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN031920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 19 15:26:45 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	158#	0.00
2	1,4-Dioxane	0.376	0.365	2.9	149	0.00
3	n-Nitrosodimethylamine	0.317	0.212	33.1#	112	0.00
4 S	2-Fluorophenol	0.792	0.809	-2.1	163#	0.00
5 S	Phenol-d6	0.985	1.017	-3.2	167#	0.00
6	bis(2-Chloroethyl)ether	0.946	0.978	-3.4	161#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	158#	-0.01
8 S	Nitrobenzene-d5	0.256	0.263	-2.7	170#	0.00
9	Naphthalene	0.997	1.018	-2.1	160#	-0.01
10	Hexachlorobutadiene	0.242	0.256	-5.8	169#	-0.01
11 SURR	2-Methylnaphthalene-d10	0.639	0.718	-12.4	181#	0.00
12	2-Methylnaphthalene	0.692	0.697	-0.7	162#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	155#	-0.02
14 S	2,4,6-Tribromophenol	0.154	0.139	9.7	158#	0.00
15 S	2-Fluorobiphenyl	1.531	1.567	-2.4	157#	0.00
16	Acenaphthylene	1.569	1.406	10.4	150#	0.00
17	Acenaphthene	1.118	1.122	-0.4	158#	0.00
18	Fluorene	1.485	1.446	2.6	156#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	154#	0.00
20	4-Bromophenyl-phenylether	0.243	0.226	7.0	146	0.00
21	Hexachlorobenzene	0.249	0.261	-4.8	162#	0.00
22	Pentachlorophenol	0.085	0.068	20.0	143	0.00
23	Phenanthrene	1.099	1.083	1.5	154#	0.00
24	Anthracene	0.940	0.864	8.1	151#	0.00
25 SURR	Fluoranthene-d10	1.206	1.301	-7.9	175#	0.00
26	Fluoranthene	1.347	1.311	2.7	156#	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	146	0.00
28	Pvrene	1.245	1.368	-9.9	162#	0.00
29 S	Terphenyl-d14	0.864	0.909	-5.2	153#	0.00
30	Benzo(a)anthracene	1.274	1.235	3.1	148	0.00
31	Chrysene	1.366	1.402	-2.6	149	0.00
32	Bis(2-ethylhexyl)phthalate	0.430	0.367	14.7	139	0.00
33	Indeno(1,2,3-cd)pyrene	1.463	1.682	-15.0	156#	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	142	-0.01
35	Benzo(b)fluoranthene	1.359	1.324	2.6	152#	-0.01
36	Benzo(k)fluoranthene	1.383	1.470	-6.3	155#	-0.02
37 C	Benzo(a)pyrene	1.147	1.123	2.1	153#	0.00
38	Dibenzo(a,h)anthracene	1.240	1.330	-7.3	152#	-0.02
39	Benzo(a,h,i)perylene	1.248	1.269	-1.7	153#	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN032720\
Data File : BN010356.D
Acq On : 27 Mar 2020 17:45
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 27 18:20:34 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN031920.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 19 15:26:45 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			