

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032724\
 Data File : BN030504.D
 Acq On : 27 Mar 2024 14:29
 Operator : MA/JU
 Sample : SSTD3.285
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD3.2214

Quant Time: Mar 27 15:08:06 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN032724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 27 15:02:37 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.724	152	5222	0.400	ng/ul	0.00
4) Naphthalene-d8	10.507	136	16349	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.358	164	9981	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.108	188	22473	0.400	ng/ul	0.00
17) Chrysene-d12	21.298	240	20893	0.400	ng/ul	0.00
23) Perylene-d12	23.543	264	16199	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.218	96	16866	2.948	ng/ul	0.00
6) 2-Methylnaphthalene-d10	0.000	152	0d	0.000	ng/ul	
18) Fluoranthene-d10	0.000	212	0d	0.000	ng/ul	
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.252	88	17500	3.018	ng/ul#	82
14) Pentachlorophenol	16.749	266	14844	3.613	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032724\
Data File : BN030504.D
Acq On : 27 Mar 2024 14:29
Operator : MA/JU
Sample : SST3.285
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD3.2214

Quant Time: Mar 27 15:08:06 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN032724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 27 15:02:37 2024
Response via : Initial Calibration

