

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN032819\
 Data File : BN005023.D
 Acq On : 28 Mar 2019 21:35
 Operator : JU/SJ
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN032819

Quant Time: Mar 29 03:58:27 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 29 03:42:17 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.430	0.426	0.9	101	0.00
3	n-Nitrosodimethylamine	0.262	0.234	10.7	108	0.00
4 S	2-Fluorophenol	0.999	1.022	-2.3	106	0.00
5 S	Phenol-d6	1.183	1.171	1.0	105	0.00
6	bis(2-Chloroethyl)ether	1.117	1.146	-2.6	104	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
8 S	Nitrobenzene-d5	0.232	0.226	2.6	106	0.00
9	Naphthalene	1.006	1.029	-2.3	106	0.00
10	Hexachlorobutadiene	0.187	0.193	-3.2	106	0.00
11 SURR	2-Methylnaphthalene-d10	0.620	0.632	-1.9	106	0.00
12	2-Methylnaphthalene	0.725	0.734	-1.2	105	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	102	0.00
14 S	2,4,6-Tribromophenol	0.224	0.206	8.0	106	0.00
15 S	2-Fluorobiphenyl	1.510	1.542	-2.1	105	0.00
16	Acenaphthylene	1.888	1.825	3.3	106	0.00
17	Acenaphthene	1.199	1.221	-1.8	107	0.00
18	Fluorene	1.609	1.630	-1.3	106	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	101	0.00
20	4-Bromophenyl-phenylether	0.229	0.226	1.3	104	0.00
21	Hexachlorobenzene	0.253	0.256	-1.2	107	0.00
22	Pentachlorophenol	0.056	0.043	23.2	110	0.00
23	Phenanthrene	1.119	1.121	-0.2	105	0.00
24	Anthracene	0.997	0.972	2.5	106	0.00
25 SURR	Fluoranthene-d10	1.142	1.141	0.1	107	0.00
26	Fluoranthene	1.371	1.363	0.6	106	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
28	Pyrene	1.136	1.166	-2.6	105	0.00
29 S	Terphenyl-d14	0.822	0.838	-1.9	105	0.00
30	Benzo(a)anthracene	1.128	1.102	2.3	105	-0.01
31	Chrysene	1.216	1.248	-2.6	104	-0.01
32	Bis(2-ethylhexyl)phthalate	0.472	0.429	9.1	106	0.00
33	Indeno(1,2,3-cd)pyrene	1.367	1.408	-3.0	105	0.00
34 I	Perylene-d12	1.000	1.000	0.0	100	0.00
35	Benzo(b)fluoranthene	1.282	1.321	-3.0	108	0.00
36	Benzo(k)fluoranthene	1.239	1.252	-1.0	102	0.00
37 C	Benzo(a)pyrene	1.140	1.152	-1.1	106	0.00
38	Dibenzo(a,h)anthracene	1.141	1.148	-0.6	105	0.00
39	Benzo(g,h,i)perylene	1.149	1.163	-1.2	105	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN032819\
Data File : BN005023.D
Acq On : 28 Mar 2019 21:35
Operator : JU/SJ
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN032819

Quant Time: Mar 29 03:58:27 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 29 03:42:17 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			