

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN032919\
 Data File : BN005025.D
 Acq On : 29 Mar 2019 08:46
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 29 23:47:35 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 29 03:42:17 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	103	0.00
2	1,4-Dioxane	0.400	0.380	5.0	100	0.00
3	n-Nitrosodimethylamine	0.400	0.341	14.8	108	0.00
4 S	2-Fluorophenol	0.400	0.402	-0.5	107	0.00
5 S	Phenol-d6	0.400	0.390	2.5	107	0.00
6	bis(2-Chloroethyl)ether	0.400	0.400	0.0	106	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	101	0.00
8 S	Nitrobenzene-d5	0.400	0.387	3.3	105	-0.01
9	Naphthalene	0.400	0.401	-0.3	104	0.00
10	Hexachlorobutadiene	0.400	0.403	-0.8	103	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.391	2.3	102	0.00
12	2-Methylnaphthalene	0.400	0.390	2.5	101	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	96	0.00
14 S	2,4,6-Tribromophenol	0.400	0.350	12.5	96	0.00
15 S	2-Fluorobiphenyl	0.400	0.410	-2.5	99	0.00
16	Acenaphthylene	0.400	0.382	4.5	99	0.00
17	Acenaphthene	0.400	0.399	0.3	99	0.00
18	Fluorene	0.400	0.397	0.8	98	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	95	0.00
20	4-Bromophenyl-phenylether	0.400	0.390	2.5	96	0.00
21	Hexachlorobenzene	0.400	0.397	0.8	98	0.00
22	Pentachlorophenol	0.400	0.380	5.0	91	0.00
23	Phenanthrene	0.400	0.394	1.5	97	0.00
24	Anthracene	0.400	0.386	3.5	98	0.00
25 SURR	Fluoranthene-d10	0.400	0.399	0.3	100	0.00
26	Fluoranthene	0.400	0.396	1.0	98	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	100	0.00
28	Pvrene	0.400	0.384	4.0	99	0.00
29 S	Terphenyl-d14	0.400	0.372	7.0	96	0.00
30	Benzo(a)anthracene	0.400	0.381	4.8	102	0.00
31	Chrysene	0.400	0.403	-0.8	102	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.350	12.5	102	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.454	-13.5	116	0.00
34 I	Perylene-d12	0.400	0.400	0.0	107	0.00
35	Benzo(b)fluoranthene	0.400	0.378	5.5	106	0.00
36	Benzo(k)fluoranthene	0.400	0.394	1.5	107	0.00
37 C	Benzo(a)pyrene	0.400	0.390	2.5	109	0.00
38	Dibenzo(a,h)anthracene	0.400	0.413	-3.2	116	0.00
39	Benzo(a,h,i)perylene	0.400	0.426	-6.5	118	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN032919\
Data File : BN005025.D
Acq On : 29 Mar 2019 08:46
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 29 23:47:35 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 29 03:42:17 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				