

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN032919\
 Data File : BN005025.D
 Acq On : 29 Mar 2019 08:46
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 29 23:47:35 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 29 03:42:17 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
2	1,4-Dioxane	0.430	0.408	5.1	100	0.00
3	n-Nitrosodimethylamine	0.262	0.224	14.5	108	0.00
4 S	2-Fluorophenol	0.999	1.003	-0.4	107	0.00
5 S	Phenol-d6	1.183	1.153	2.5	107	0.00
6	bis(2-Chloroethyl)ether	1.117	1.118	-0.1	106	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	101	0.00
8 S	Nitrobenzene-d5	0.232	0.224	3.4	105	-0.01
9	Naphthalene	1.006	1.008	-0.2	104	0.00
10	Hexachlorobutadiene	0.187	0.188	-0.5	103	0.00
11 SURR	2-Methylnaphthalene-d10	0.620	0.606	2.3	102	0.00
12	2-Methylnaphthalene	0.725	0.707	2.5	101	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	96	0.00
14 S	2,4,6-Tribromophenol	0.224	0.197	12.1	96	0.00
15 S	2-Fluorobiphenyl	1.510	1.546	-2.4	99	0.00
16	Acenaphthylene	1.888	1.803	4.5	99	0.00
17	Acenaphthene	1.199	1.195	0.3	99	0.00
18	Fluorene	1.609	1.597	0.7	98	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	95	0.00
20	4-Bromophenyl-phenylether	0.229	0.223	2.6	96	0.00
21	Hexachlorobenzene	0.253	0.251	0.8	98	0.00
22	Pentachlorophenol	0.056	0.038	32.1#	91	0.00
23	Phenanthrene	1.119	1.104	1.3	97	0.00
24	Anthracene	0.997	0.961	3.6	98	0.00
25 SURR	Fluoranthene-d10	1.142	1.139	0.3	100	0.00
26	Fluoranthene	1.371	1.357	1.0	98	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
28	Pvrene	1.136	1.091	4.0	99	0.00
29 S	Terphenyl-d14	0.822	0.765	6.9	96	0.00
30	Benzo(a)anthracene	1.128	1.074	4.8	102	0.00
31	Chrysene	1.216	1.224	-0.7	102	0.00
32	Bis(2-ethylhexyl)phthalate	0.472	0.413	12.5	102	0.00
33	Indeno(1,2,3-cd)pyrene	1.367	1.551	-13.5	116	0.00
34 I	Perylene-d12	1.000	1.000	0.0	107	0.00
35	Benzo(b)fluoranthene	1.282	1.213	5.4	106	0.00
36	Benzo(k)fluoranthene	1.239	1.221	1.5	107	0.00
37 C	Benzo(a)pyrene	1.140	1.112	2.5	109	0.00
38	Dibenzo(a,h)anthracene	1.141	1.179	-3.3	116	0.00
39	Benzo(a,h,i)perylene	1.149	1.224	-6.5	118	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN032919\
Data File : BN005025.D
Acq On : 29 Mar 2019 08:46
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 29 23:47:35 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 29 03:42:17 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			