

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032921\
 Data File : BN014128.D
 Acq On : 29 Mar 2021 11:50
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Mar 29 14:45:52 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 29 12:24:58 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.684	152	6817	0.40	ng	0.00
7) Naphthalene-d8	10.454	136	24408	0.40	ng	# 0.00
13) Acenaphthene-d10	14.308	164	12265	0.40	ng	0.00
19) Phenanthrene-d10	17.043	188	23773	0.40	ng	# 0.00
29) Chrysene-d12	21.226	240	20072	0.40	ng	# 0.00
36) Perylene-d12	23.428	264	18338	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.284	112	7033	0.40	ng	0.00
5) Phenol-d6	6.855	99	8689	0.39	ng	0.00
8) Nitrobenzene-d5	8.832	82	6203	0.37	ng	0.00
11) 2-Methylnaphthalene-d10	12.048	152	15010	0.40	ng	0.00
14) 2,4,6-Tribromophenol	15.805	330	1372	0.35	ng	0.00
15) 2-Fluorobiphenyl	12.938	172	19197	0.42	ng	0.00
27) Fluoranthene-d10	19.081	212	24894	0.40	ng	0.00
31) Terphenyl-d14	19.677	244	19389	0.42	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.231	88	3516	0.40	ng	98
3) n-Nitrosodimethylamine	3.569	42	2423	0.36	ng	# 92
6) bis(2-Chloroethyl)ether	7.120	93	7329	0.41	ng	99
9) Naphthalene	10.505	128	24786	0.41	ng	100
10) Hexachlorobutadiene	10.796	225	4593	0.42	ng	# 99
12) 2-Methylnaphthalene	12.124	142	16030	0.41	ng	100
16) Acenaphthylene	14.029	152	17814	0.38	ng	99
17) Acenaphthene	14.371	154	13459	0.40	ng	99
18) Fluorene	15.361	166	16458	0.40	ng	99
20) 4,6-Dinitro-2-methylph...	15.437	198	984	0.33	ng	98
21) 4-Bromophenyl-phenylether	16.252	248	4940	0.40	ng	97
22) Hexachlorobenzene	16.361	284	5375	0.39	ng	98
23) Atrazine	16.519	200	2906	0.35	ng	98
24) Pentachlorophenol	16.702	266	1344	0.32	ng	97
25) Phenanthrene	17.091	178	25640	0.40	ng	100
26) Anthracene	17.176	178	20573	0.38	ng	100
28) Fluoranthene	19.111	202	26675	0.40	ng	100
30) Pyrene	19.473	202	26727	0.41	ng	100
32) Benzo(a)anthracene	21.205	228	22127	0.40	ng	97
33) Chrysene	21.258	228	25771	0.42	ng	99
34) Bis(2-ethylhexyl)phtha...	21.141	149	6363	0.37	ng	100
35) Indeno(1,2,3-cd)pyrene	25.629	276	28214	0.44	ng	99
37) Benzo(b)fluoranthene	22.767	252	26124	0.40	ng	99
38) Benzo(k)fluoranthene	22.809	252	25532	0.40	ng	99
39) Benzo(a)pyrene	23.329	252	21879	0.39	ng	99
40) Dibenzo(a,h)anthracene	25.639	278	23809	0.42	ng	99
41) Benzo(g,h,i)perylene	26.300	276	25669	0.42	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032921\
Data File : BN014128.D
Acq On : 29 Mar 2021 11:50
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Quant Time: Mar 29 14:45:52 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 29 12:24:58 2021
Response via : Initial Calibration






Abundance Scan 610 (7.692 min): BN014120.D\data.ms (-602) (#1)




1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.684 min Scan# 609
Delta R.T. 0.000 min
Lab File: BN014128.D
Acq: 29 Mar 2021 11:50

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:152 Resp: 6817
Ion Ratio Lower Upper
152 100
150 159.6 121.2 181.8
115 62.1 46.2 69.2







Abundance Scan 56 (3.231 min): BN014120.D\data.ms (-52) (#2)



1,4-Dioxane
Concen: 0.40 ng
RT: 3.231 min Scan# 56
Delta R.T. 0.000 min
Lab File: BN014128.D
Acq: 29 Mar 2021 11:50

Tgt Ion: 88 Resp: 3516
Ion Ratio Lower Upper
88 100
58 81.4 63.1 94.7
43 32.4 26.2 39.2









n-Nitrosodimethylamine
Concen: 0.36 ng
RT: 3.569 min Scan# 98
Delta R.T. 0.000 min
Lab File: BN014128.D
Acq: 29 Mar 2021 11:50




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 42 Resp: 2423
Ion Ratio Lower Upper
42 100
74 109.0 94.6 141.8
44 6.7 4.2 6.2#



2-Fluorophenol
Concen: 0.40 ng
RT: 5.284 min Scan# 311
Delta R.T. 0.000 min
Lab File: BN014128.D
Acq: 29 Mar 2021 11:50

Tgt Ion: 112 Resp: 7033
Ion Ratio Lower Upper
112 100
64 59.8 48.3 72.5
63 31.6 25.4 38.0









Phenol-d6
 Concen: 0.39 ng
 RT: 6.855 min Scan# 506
 Delta R.T. 0.000 min
 Lab File: BN014128.D
 Acq: 29 Mar 2021 11:50




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Tgt Ion:	99	Resp:	8689
Ion	Ratio	Lower	Upper
99	100		
42	21.5	18.2	27.4
71	33.2	26.7	40.1




bis(2-Chloroethyl)ether
 Concen: 0.41 ng
 RT: 7.120 min Scan# 539
 Delta R.T. 0.000 min
 Lab File: BN014128.D
 Acq: 29 Mar 2021 11:50

Tgt Ion:	93	Resp:	7329
Ion	Ratio	Lower	Upper
93	100		
63	82.2	65.1	97.7
95	33.6	26.6	39.8






Abundance Scan 857 (10.467 min): BN014120.D\data.ms (-857#7-)




Naphthalene-d8
Concen: 0.40 ng
RT: 10.454 min Scan# 856
Delta R.T. 0.000 min
Lab File: BN014128.D
Acq: 29 Mar 2021 11:50

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:	136	Resp:	24408
Ion	Ratio	Lower	Upper
136	100		
137	10.9	9.1	13.7
54	10.7	4.3	6.5#
68	7.6	3.4	5.2#






Abundance Scan 728 (8.832 min): BN014120.D\data.ms (-724#8)




Nitrobenzene-d5
Concen: 0.37 ng
RT: 8.832 min Scan# 728
Delta R.T. 0.000 min
Lab File: BN014128.D
Acq: 29 Mar 2021 11:50

Tgt Ion:	82	Resp:	6203
Ion	Ratio	Lower	Upper
82	100		
128	42.7	33.5	50.3
54	49.1	40.3	60.5





Abundance Scan 860 (10.505 min): BN014120.D\data.ms (-855) (-)




Naphthalene
Concen: 0.41 ng
RT: 10.505 min Scan# 860
Delta R.T. 0.000 min
Lab File: BN014128.D
Acq: 29 Mar 2021 11:50

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	128	Resp:	24786
Ion	Ratio	Lower	Upper
128	100		
129	11.4	9.2	13.8
127	13.4	10.8	16.2



Abundance Scan 860 (10.505 min): BN014128.D\data.ms (-852) (-)





Abundance Scan 883 (10.796 min): BN014120.D\data.ms (-875) (-)




Hexachlorobutadiene
Concen: 0.42 ng
RT: 10.796 min Scan# 883
Delta R.T. 0.000 min
Lab File: BN014128.D
Acq: 29 Mar 2021 11:50


Tgt Ion:	225	Resp:	4593
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	63.4	51.6	77.4



Abundance Scan 883 (10.796 min): BN014128.D\data.ms (-875) (-)



Abundance Scan 1062 (12.053 min): BN014120.D\data.ms (-1019) (-)





2-Methylnaphthalene-d10
Concen: 0.40 ng
RT: 12.048 min Scan# 1061
Delta R.T. 0.000 min
Lab File: BN014128.D
Acq: 29 Mar 2021 11:50


Tgt Ion:152 Resp: 15010
Ion Ratio Lower Upper
152 100
151 20.0 16.1 24.1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Abundance Scan 1061 (12.048 min): BN014128.D\data.ms



Abundance Scan 1061 (12.048 min): BN014128.D\data.ms (-1016) (-)





Abundance Scan 1078 (12.124 min): BN014120.D\data.ms (-1012) (-)




2-Methylnaphthalene
Concen: 0.41 ng
RT: 12.124 min Scan# 1078
Delta R.T. 0.000 min
Lab File: BN014128.D
Acq: 29 Mar 2021 11:50

Tgt Ion:142 Resp: 16030
Ion Ratio Lower Upper
142 100
141 89.1 71.1 106.7
115 35.9 29.0 43.6

Abundance Scan 1078 (12.124 min): BN014128.D\data.ms




Abundance Scan 1078 (12.124 min): BN014128.D\data.ms (-1056) (-)





Abundance Scan 1306 (14.308 min): BN014120.D\data.ms (-1290) (-)

Acenaphthene-d10
Concen: 0.40 ng
RT: 14.308 min Scan# 1306
Delta R.T. 0.000 min
Lab File: BN014128.D
Acq: 29 Mar 2021 11:50


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:	164	Resp:	12265
Ion	Ratio	Lower	Upper
164	100		
162	103.5	81.4	122.2
160	47.2	37.5	56.3






Abundance Scan 1306 (14.308 min): BN014128.D\data.ms (-1290) (-)




Abundance Scan 1424 (15.805 min): BN014120.D\data.ms (-1416) (-)

2,4,6-Tribromophenol
Concen: 0.35 ng
RT: 15.805 min Scan# 1424
Delta R.T. 0.000 min
Lab File: BN014128.D
Acq: 29 Mar 2021 11:50


Tgt Ion:	330	Resp:	1372
Ion	Ratio	Lower	Upper
330	100		
332	0.0	0.0	0.0
141	38.3	31.2	46.8



Abundance Scan 1424 (15.805 min): BN014128.D\data.ms (-1416) (-)






Abundance Scan 1198 (12.938 min): BN014120.D\data.ms (-1198) (-)




2-Fluorobiphenyl
Concen: 0.42 ng
RT: 12.938 min Scan# 1198
Delta R.T. 0.000 min
Lab File: BN014128.D
Acq: 29 Mar 2021 11:50

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:	172	Resp:	19197
Ion	Ratio	Lower	Upper
172	100		
171	33.6	26.9	40.3
170	22.1	17.8	26.8






Abundance Scan 1284 (14.029 min): BN014120.D\data.ms (-1215) (-)



Acenaphthylene
Concen: 0.38 ng
RT: 14.029 min Scan# 1284
Delta R.T. 0.000 min
Lab File: BN014128.D
Acq: 29 Mar 2021 11:50


Tgt Ion:	152	Resp:	17814
Ion	Ratio	Lower	Upper
152	100		
151	19.1	15.4	23.2
153	13.2	10.3	15.5






Abundance Scan 1311 (14.371 min): BN014120.D\data.ms (-1305) (-)

Acenaphthene
Concen: 0.40 ng
RT: 14.371 min Scan# 1311
Delta R.T. 0.000 min
Lab File: BN014128.D
Acq: 29 Mar 2021 11:50

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4







Tgt Ion:154 Resp: 13459
Ion Ratio Lower Upper
154 100
153 111.7 89.0 133.4
152 54.5 43.1 64.7




Abundance Scan 1389 (15.361 min): BN014120.D\data.ms (-1314) (-)

Fluorene
Concen: 0.40 ng
RT: 15.361 min Scan# 1389
Delta R.T. 0.000 min
Lab File: BN014128.D
Acq: 29 Mar 2021 11:50

Tgt Ion:166 Resp: 16458
Ion Ratio Lower Upper
166 100
165 98.4 77.9 116.9
167 13.4 10.6 16.0





Abundance Scan 1526 (17.055 min): BN014120.D\data.ms (-1526) (-)




Phenanthrene-d10
Concen: 0.40 ng
RT: 17.043 min Scan# 1525
Delta R.T. 0.000 min
Lab File: BN014128.D
Acq: 29 Mar 2021 11:50

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:188 Resp: 23773
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 12.8 5.2 7.8#



Abundance Scan 1525 (17.043 min): BN014128.D\data.ms (-1509) (-)





Abundance Scan 1395 (15.437 min): BN014120.D\data.ms (-1395) (-)




4,6-Dinitro-2-methylphenol
Concen: 0.33 ng
RT: 15.437 min Scan# 1395
Delta R.T. 0.000 min
Lab File: BN014128.D
Acq: 29 Mar 2021 11:50


Tgt Ion:198 Resp: 984
Ion Ratio Lower Upper
198 100
51 60.7 47.8 71.6
105 42.2 35.4 53.2



Abundance Scan 1395 (15.437 min): BN014128.D\data.ms (-1387) (-)






Abundance Scan 1460 (16.252 min): BN014120.D\data.ms (-1423) (-)




4-Bromophenyl-phenylether
Concen: 0.40 ng
RT: 16.252 min Scan# 1460
Delta R.T. 0.000 min
Lab File: BN014128.D
Acq: 29 Mar 2021 11:50

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:	248	Resp:	4940
Ion	Ratio	Lower	Upper
248	100		
250	98.8	77.0	115.6
141	46.2	39.7	59.5






Abundance Scan 1469 (16.361 min): BN014120.D\data.ms (-1423) (-)




Hexachlorobenzene
Concen: 0.39 ng
RT: 16.361 min Scan# 1469
Delta R.T. 0.000 min
Lab File: BN014128.D
Acq: 29 Mar 2021 11:50

Tgt Ion:	284	Resp:	5375
Ion	Ratio	Lower	Upper
284	100		
142	37.6	29.2	43.8
249	31.5	24.7	37.1





Abundance Scan 1482 (16.519 min): BN014120.D\data.ms (-1425) (-)




Atrazine
Concen: 0.35 ng
RT: 16.519 min Scan# 1482
Delta R.T. 0.000 min
Lab File: BN014128.D
Acq: 29 Mar 2021 11:50

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	200	Resp:	2906
Ion	Ratio	Lower	Upper
200	100		
173	24.2	20.9	31.3
215	50.2	39.5	59.3



Abundance Scan 1482 (16.519 min): BN014128.D\data.ms (-1441) (-)





Abundance Scan 1497 (16.702 min): BN014120.D\data.ms (-1424) (-)




Pentachlorophenol
Concen: 0.32 ng
RT: 16.702 min Scan# 1497
Delta R.T. 0.000 min
Lab File: BN014128.D
Acq: 29 Mar 2021 11:50

Tgt Ion:	266	Resp:	1344
Ion	Ratio	Lower	Upper
266	100		
264	64.6	50.2	75.2
268	62.9	52.5	78.7




Abundance Scan 1497 (16.702 min): BN014128.D\data.ms (-1489) (-)






Abundance Scan 1529 (17.091 min): BN014120.D\data.ms (-1525) (-)

Phenanthrene
Concen: 0.40 ng
RT: 17.091 min Scan# 1529
Delta R.T. 0.000 min
Lab File: BN014128.D
Acq: 29 Mar 2021 11:50

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4







Tgt Ion:	178	Resp:	25640
Ion	Ratio	Lower	Upper
178	100		
176	18.7	15.0	22.6
179	15.5	12.4	18.6




Abundance Scan 1536 (17.177 min): BN014120.D\data.ms (-1535) (-)

Anthracene
Concen: 0.38 ng
RT: 17.176 min Scan# 1536
Delta R.T. 0.000 min
Lab File: BN014128.D
Acq: 29 Mar 2021 11:50

Tgt Ion:	178	Resp:	20573
Ion	Ratio	Lower	Upper
178	100		
176	18.0	14.7	22.1
179	15.4	12.3	18.5






Abundance Scan 1821 (19.081 min): BN014120.D\data.ms (-1821) (-)




Fluoranthene-d10
Concen: 0.40 ng
RT: 19.081 min Scan# 1821
Delta R.T. 0.000 min
Lab File: BN014128.D
Acq: 29 Mar 2021 11:50

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:212 Resp: 24894
Ion Ratio Lower Upper
212 100
106 12.5 10.1 15.1
104 7.1 5.5 8.3






Abundance Scan 1827 (19.111 min): BN014120.D\data.ms (-1827) (-)




Fluoranthene
Concen: 0.40 ng
RT: 19.111 min Scan# 1827
Delta R.T. 0.000 min
Lab File: BN014128.D
Acq: 29 Mar 2021 11:50

Tgt Ion:202 Resp: 26675
Ion Ratio Lower Upper
202 100
101 10.9 8.8 13.2
203 17.2 13.8 20.6






Abundance Scan 2122 (21.226 min): BN014120.D\data.ms (-2129) (-)




Chrysene-d12
Concen: 0.40 ng
RT: 21.226 min Scan# 2122
Delta R.T. 0.000 min
Lab File: BN014128.D
Acq: 29 Mar 2021 11:50

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:240 Resp: 20072
Ion Ratio Lower Upper
240 100
120 7.8 8.4 12.6#
236 25.7 21.0 31.6






Abundance Scan 1900 (19.473 min): BN014120.D\data.ms (-1830) (-)



Pyrene
Concen: 0.41 ng
RT: 19.473 min Scan# 1900
Delta R.T. 0.000 min
Lab File: BN014128.D
Acq: 29 Mar 2021 11:50


Tgt Ion:202 Resp: 26727
Ion Ratio Lower Upper
202 100
200 20.4 16.2 24.4
203 17.8 13.9 20.9





Abundance Scan 1942 (19.682 min): BN014120.D\data.ms (-1934) (-)

Terphenyl-d14
Concen: 0.42 ng
RT: 19.677 min Scan# 1941
Delta R.T. 0.000 min
Lab File: BN014128.D
Acq: 29 Mar 2021 11:50


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:	244	Resp:	19389
Ion	Ratio	Lower	Upper
244	100		
212	7.7	6.0	9.0
122	12.5	8.6	13.0






Abundance Scan 1941 (19.677 min): BN014128.D\data.ms (-1911) (-)




Abundance Scan 2121 (21.215 min): BN014120.D\data.ms (-2135) (-)

Benzo(a)anthracene
Concen: 0.40 ng
RT: 21.205 min Scan# 2120
Delta R.T. 0.000 min
Lab File: BN014128.D
Acq: 29 Mar 2021 11:50


Tgt Ion:	228	Resp:	22127
Ion	Ratio	Lower	Upper
228	100		
226	27.5	20.6	31.0
229	19.4	16.2	24.2



Abundance Scan 2120 (21.205 min): BN014128.D\data.ms (-2101) (-)






Abundance Scan 2125 (21.258 min): BN014120.D\data.ms (-228) (-)




Chrysene
Concen: 0.42 ng
RT: 21.258 min Scan# 2125
Delta R.T. 0.000 min
Lab File: BN014128.D
Acq: 29 Mar 2021 11:50

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:	228	Resp:	25771
Ion	Ratio	Lower	Upper
228	100		
226	29.8	24.4	36.6
229	19.9	15.5	23.3






Abundance Scan 2115 (21.152 min): BN014120.D\data.ms (-149) (-)




Bis(2-ethylhexyl)phthalate
Concen: 0.37 ng
RT: 21.141 min Scan# 2114
Delta R.T. 0.000 min
Lab File: BN014128.D
Acq: 29 Mar 2021 11:50

Tgt Ion:	149	Resp:	6363
Ion	Ratio	Lower	Upper
149	100		
167	26.7	21.4	32.2
279	4.4	3.5	5.3






Abundance Scan 3176 (25.633 min): BN014120.D\data.ms (-3135) (-)




Indeno(1,2,3-cd)pyrene
Concen: 0.44 ng
RT: 25.629 min Scan# 3175
Delta R.T. 0.000 min
Lab File: BN014128.D
Acq: 29 Mar 2021 11:50

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:	276	Resp:	28214
Ion	Ratio	Lower	Upper
276	100		
138	24.0	19.4	29.2
277	24.9	19.8	29.6






Abundance Scan 2533 (23.432 min): BN014120.D\data.ms (-2526) (-)




Perylene-d12
Concen: 0.40 ng
RT: 23.428 min Scan# 2532
Delta R.T. 0.000 min
Lab File: BN014128.D
Acq: 29 Mar 2021 11:50

Tgt Ion:	264	Resp:	18338
Ion	Ratio	Lower	Upper
264	100		
260	24.3	19.8	29.8
265	35.1	30.7	46.1






Abundance Scan 2340 (22.771 min): BN014120.D\data.ms (-225) (-)




Benzo(b)fluoranthene
Concen: 0.40 ng
RT: 22.767 min Scan# 2339
Delta R.T. 0.000 min
Lab File: BN014128.D
Acq: 29 Mar 2021 11:50

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:252 Resp: 26124
Ion Ratio Lower Upper
252 100
253 25.3 20.6 30.8
125 11.6 9.4 14.2






Abundance Scan 2352 (22.812 min): BN014120.D\data.ms (-226) (-)



Benzo(k)fluoranthene
Concen: 0.40 ng
RT: 22.809 min Scan# 2351
Delta R.T. 0.000 min
Lab File: BN014128.D
Acq: 29 Mar 2021 11:50


Tgt Ion:252 Resp: 25532
Ion Ratio Lower Upper
252 100
253 25.2 20.6 30.8
125 11.0 9.5 14.3





Abundance Scan 2505 (23.336 min): BN014120.D\data.ms (-2439) (-)

Benzo(a)pyrene
Concen: 0.39 ng
RT: 23.329 min Scan# 2503
Delta R.T. 0.000 min
Lab File: BN014128.D
Acq: 29 Mar 2021 11:50


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:	252	Resp:	21879
Ion	Ratio	Lower	Upper
252	100		
253	27.4	21.9	32.9
125	14.2	10.8	16.2





Abundance Scan 2503 (23.329 min): BN014128.D\data.ms (-2445) (-)




Abundance Scan 3179 (25.643 min): BN014120.D\data.ms (-3140) (-)

Dibenzo(a,h)anthracene
Concen: 0.42 ng
RT: 25.639 min Scan# 3178
Delta R.T. 0.000 min
Lab File: BN014128.D
Acq: 29 Mar 2021 11:50


Tgt Ion:	278	Resp:	23809
Ion	Ratio	Lower	Upper
278	100		
139	16.1	13.4	20.0
279	26.3	21.3	31.9



Abundance Scan 3178 (25.639 min): BN014128.D\data.ms (-3155) (-)




Abundance Scan 3372 (26.303 min): BN014120.D\data.ms (-3349) (-)




Benzo(g,h,i)perylene
Concen: 0.42 ng
RT: 26.300 min Scan# 3371
Delta R.T. 0.000 min
Lab File: BN014128.D
Acq: 29 Mar 2021 11:50

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Abundance Scan 3371 (26.300 min): BN014128.D\data.ms



Tgt Ion:	276	Resp:	25669
Ion	Ratio	Lower	Upper
276	100		
277	24.1	20.1	30.1
138	21.5	17.0	25.4



Abundance Scan 3371 (26.300 min): BN014128.D\data.ms (-3313) (-)

