

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032921\
 Data File : BN014136.D
 Acq On : 29 Mar 2021 20:43
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 30 04:59:45 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 29 12:24:58 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	93	0.00
2	1,4-Dioxane	0.400	0.413	-3.2	88	0.00
3	n-Nitrosodimethylamine	0.400	0.337	15.8	81	0.00
4 S	2-Fluorophenol	0.400	0.404	-1.0	88	0.00
5 S	Phenol-d6	0.400	0.389	2.8	86	0.00
6	bis(2-Chloroethyl)ether	0.400	0.416	-4.0	91	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	91	0.00
8 S	Nitrobenzene-d5	0.400	0.358	10.5	82	0.00
9	Naphthalene	0.400	0.414	-3.5	91	0.00
10	Hexachlorobutadiene	0.400	0.415	-3.7	91	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.401	-0.3	89	0.00
12	2-Methylnaphthalene	0.400	0.405	-1.3	89	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	87	0.00
14 S	2,4,6-Tribromophenol	0.400	0.338	15.5	79	0.00
15 S	2-Fluorobiphenyl	0.400	0.406	-1.5	88	0.00
16	Acenaphthylene	0.400	0.357	10.8	83	0.00
17	Acenaphthene	0.400	0.387	3.3	85	0.00
18	Fluorene	0.400	0.394	1.5	86	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	90	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.292	27.0#	74	0.00
21	4-Bromophenyl-phenylether	0.400	0.388	3.0	86	0.00
22	Hexachlorobenzene	0.400	0.400	0.0	88	0.00
23	Atrazine	0.400	0.342	14.5	80	0.00
24	Pentachlorophenol	0.400	0.388	3.0	100	0.00
25	Phenanthrene	0.400	0.405	-1.3	89	0.00
26	Anthracene	0.400	0.380	5.0	87	0.00
27 SURR	Fluoranthene-d10	0.400	0.412	-3.0	93	0.00
28	Fluoranthene	0.400	0.412	-3.0	94	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	109	0.00
30	Pyrene	0.400	0.374	6.5	94	0.00
31 S	Terphenyl-d14	0.400	0.389	2.8	98	0.00
32	Benzo(a)anthracene	0.400	0.415	-3.7	112	0.00
33	Chrysene	0.400	0.417	-4.2	107	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.394	1.5	111	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.500	-25.0	132	0.00
36 I	Perylene-d12	0.400	0.400	0.0	121	0.00
37	Benzo(b)fluoranthene	0.400	0.395	1.3	117	0.00
38	Benzo(k)fluoranthene	0.400	0.449	-12.2	134	0.00
39 C	Benzo(a)pyrene	0.400	0.416	-4.0	127	0.00
40	Dibenzo(a,h)anthracene	0.400	0.451	-12.7	136	0.00
41	Benzo(g,h,i)perylene	0.400	0.433	-8.2	128	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4