

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032923\
 Data File : BN024596.D
 Acq On : 28 Mar 2023 21:49
 Operator : CG/JU
 Sample : PB151703BS
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS703

Quant Time: Mar 29 05:04:46 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032923.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 29 03:32:05 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.998	152	354141	20.000	ng/u1	0.00
20) Naphthalene-d8	10.816	136	1596987	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.639	164	1024591	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.386	188	2177398	20.000	ng/u1	0.00
79) Chrysene-d12	21.580	240	1989320	20.000	ng/u1	0.00
88) Perylene-d12	24.068	264	2122482	20.000	ng/u1	0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.369	96	52856	5.787	ng/uL	0.00
4) Pyridine-d5	3.799	84	745700	28.273	ng/u1	0.00
7) Phenol-d5	7.157	99	1080155	32.229	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.328	67	687303	31.757	ng/u1	0.00
11) 2-Chlorophenol-d4	7.528	132	810703	32.628	ng/u1	0.00
15) 4-Methylphenol-d8	8.716	113	864421	32.251	ng/u1	0.00
21) Nitrobenzene-d5	9.175	128	416928	32.988	ng/u1	0.00
24) 2-Nitrophenol-d4	9.898	143	460001	33.625	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.434	165	834394	33.581	ng/u1	0.00
31) 4-Chloroaniline-d4	10.957	131	1130927	29.830	ng/u1	0.00
46) Dimethylphthalate-d6	14.051	166	2484907	31.538	ng/u1	0.00
49) Acenaphthylene-d8	14.339	160	2925683	32.610	ng/u1	0.00
54) 4-Nitrophenol-d4	14.839	143	534645	32.421	ng/u1	0.00
60) Fluorene-d10	15.633	176	2141558	31.678	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.751	200	471242	32.858	ng/u1	0.00
73) Anthracene-d10	17.486	188	3300863	32.231	ng/u1	0.00
81) Pyrene-d10	19.780	212	3729202	32.421	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.909	264	3757837	34.942	ng/u1	0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.405	88	116540	11.841	ng/uL	98
5) Pyridine	3.816	79	808936	29.465	ng/u1	96
6) Benzaldehyde	7.134	77	641533	41.938	ng/u1	96
8) Phenol	7.187	94	1124698	32.329	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.422	93	884057	32.226	ng/u1	98
12) 2-Chlorophenol	7.563	128	838451	32.344	ng/u1	98
13) 2-Methylphenol	8.445	108	842134	32.415	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.534	45	1553985	31.888	ng/u1	100
16) Acetophenone	8.834	105	1353884	32.358	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.822	70	720678	32.729	ng/u1	100
18) 4-Methylphenol	8.781	108	932318	32.324	ng/u1	100
19) Hexachloroethane	9.087	117	329736	32.206	ng/u1	97
22) Nitrobenzene	9.216	77	1065346	32.451	ng/u1	99
23) Isophorone	9.745	82	2019010	33.088	ng/u1	98
25) 2-Nitrophenol	9.928	139	496661	33.394	ng/u1	98
26) 2,4-Dimethylphenol	9.987	107	965740	31.261	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.228	93	1241514	31.704	ng/u1	99
29) 2,4-Dichlorophenol	10.463	162	818472	32.476	ng/u1	95
30) Naphthalene	10.869	128	2802893	31.586	ng/u1	99
32) 4-Chloroaniline	10.981	127	1162475	30.279	ng/u1	100
33) Hexachlorobutadiene	11.151	225	460030	31.695	ng/u1	98
34) Caprolactam	11.763	113	158102	18.139	ng/u1	97
35) 4-Chloro-3-methylphenol	12.098	107	963044	33.424	ng/u1	99
36) 2-Methylnaphthalene	12.475	142	1883536	31.705	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032923\
 Data File : BN024596.D
 Acq On : 28 Mar 2023 21:49
 Operator : CG/JU
 Sample : PB151703BS
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS703

Quant Time: Mar 29 05:04:46 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032923.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 29 03:32:05 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.692	142	1956791	33.243	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.839	216	948483	31.229	ng/ul	98
40) Hexachlorocyclopentadiene	12.822	237	565297	29.577	ng/ul	98
41) 2,4,6-Trichlorophenol	13.075	196	651832	32.213	ng/ul	99
42) 2,4,5-Trichlorophenol	13.151	196	739639	33.258	ng/ul	99
43) 1,1'-Biphenyl	13.480	154	2511960	30.858	ng/ul	99
44) 2-Chloronaphthalene	13.522	162	1991643	31.340	ng/ul	99
45) 2-Nitroaniline	13.728	65	688262	34.858	ng/ul	96
47) Dimethylphthalate	14.098	163	2552942	31.907	ng/ul	99
48) 2,6-Dinitrotoluene	14.222	165	533704	35.215	ng/ul	98
50) Acenaphthylene	14.369	152	3146951	32.038	ng/ul	100
51) 3-Nitroaniline	14.551	138	568291	35.277	ng/ul	95
52) Acenaphthene	14.704	153	2156853	31.510	ng/ul	95
53) 2,4-Dinitrophenol	14.751	184	329173	31.975	ng/ul	99
55) 4-Nitrophenol	14.851	109	413634	32.784	ng/ul	98
56) Dibenzofuran	15.039	168	3011201	31.614	ng/ul	97
57) 2,4-Dinitrotoluene	15.004	165	769633	34.663	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	15.263	232	620165	33.463	ng/ul	98
59) Diethylphthalate	15.457	149	2533423	32.357	ng/ul	99
61) Fluorene	15.692	166	2426915	31.778	ng/ul	97
62) 4-Chlorophenyl-phenyle...	15.680	204	1145569	31.615	ng/ul	99
63) 4-Nitroaniline	15.716	138	609535	38.408	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.769	198	463548	32.685	ng/ul#	95
67) N-Nitrosodiphenylamine	15.892	169	2129613	32.707	ng/ul	94
68) 4-Bromophenyl-phenylether	16.574	248	705650	32.470	ng/ul	99
69) Hexachlorobenzene	16.692	284	817213	31.952	ng/ul	96
70) Atrazine	16.845	200	570796	25.659	ng/ul	96
71) Pentachlorophenol	17.033	266	507554	32.463	ng/ul	96
72) Phenanthrene	17.433	178	3945144	32.334	ng/ul	98
74) Anthracene	17.521	178	3892583	31.756	ng/ul	96
75) 1,2,3,4-Tetrachloroben...	13.445	216	943186	30.299	ng/ul	97
76) Pentachlorobenzene	14.957	250	961040	31.782	ng/ul	96
77) Carbazole	17.792	167	3707335	33.638	ng/ul	99
78) Di-n-butylphthalate	18.351	149	4301107	34.439	ng/ul	100
80) Fluoranthene	19.445	202	4517201	33.116	ng/ul	99
82) Pyrene	19.810	202	4691666	32.627	ng/ul	97
83) Butylbenzylphthalate	20.698	149	1920988	35.085	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.492	252	1512255	32.520	ng/ul	97
85) Benzo(a)anthracene	21.562	228	4477698	31.864	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.480	149	2709391	35.491	ng/ul#	100
87) Chrysene	21.615	228	4325305	32.528	ng/ul	98
89) Di-n-octyl phthalate	22.439	149	4699073	36.547	ng/ul	100
90) Benzo(b)fluoranthene	23.304	252	4710992	33.896	ng/ul	97
91) Benzo(k)fluoranthene	23.356	252	4559795	34.367	ng/ul	97
93) Benzo(a)pyrene	23.962	252	4127688	33.916	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	26.680	276	5190034	33.465	ng/ul	99
95) Dibenzo(a,h)anthracene	26.697	278	4149118	32.258	ng/ul	99
96) Benzo(g,h,i)perylene	27.480	276	4114548	31.930	ng/ul	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032923\
Data File : BN024596.D
Acq On : 28 Mar 2023 21:49
Operator : CG/JU
Sample : PB151703BS
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SLCS703

Quant Time: Mar 29 05:04:46 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032923.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Mar 29 03:32:05 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 11.841 ng/uL
RT: 3.405 min Scan# 20
Delta R.T. -0.000 min
Lab File: BN024596.D
Acq: 28 Mar 2023 21:49




Instrument :
BNA_N
ClientSampleId :
SLCS703


Tgt Ion: 88 Resp: 116540
Ion Ratio Lower Upper
88 100
43 36.8 29.0 43.6
58 86.0 67.5 101.3



#5
Pyridine
Concen: 29.465 ng/uL
RT: 3.816 min Scan# 90
Delta R.T. -0.000 min
Lab File: BN024596.D
Acq: 28 Mar 2023 21:49

Tgt Ion: 79 Resp: 808936
Ion Ratio Lower Upper
79 100
52 90.6 76.0 114.0
51 42.5 35.0 52.4









#6
Benzaldehyde
Concen: 41.938 ng/ul
RT: 7.134 min Scan# 61
Delta R.T. -0.000 min
Lab File: BN024596.D
Acq: 28 Mar 2023 21:49




Instrument :
BNA_N
ClientSampleId :
SLCS703


Tgt Ion: 77 Resp: 641533
Ion Ratio Lower Upper
77 100
105 95.1 74.0 111.0
106 94.4 71.7 107.5



#8
Phenol
Concen: 32.329 ng/ul
RT: 7.187 min Scan# 663
Delta R.T. -0.000 min
Lab File: BN024596.D
Acq: 28 Mar 2023 21:49

Tgt Ion: 94 Resp: 1124698
Ion Ratio Lower Upper
94 100
65 29.8 23.7 35.5
66 38.0 29.9 44.9









#10
 Bis(2-Chloroethyl)ether
 Concen: 32.226 ng/ul
 RT: 7.422 min Scan# 703
 Delta R.T. -0.000 min
 Lab File: BN024596.D
 Acq: 28 Mar 2023 21:49

Instrument :
 BNA_N
 ClientSampleId :
 SLCS703

Tgt Ion: 93 Resp: 884057




Ion	Ratio	Lower	Upper
93	100		
63	82.7	64.9	97.3
95	32.0	25.2	37.8




#12
 2-Chlorophenol
 Concen: 32.344 ng/ul
 RT: 7.563 min Scan# 727
 Delta R.T. -0.000 min
 Lab File: BN024596.D
 Acq: 28 Mar 2023 21:49

Tgt Ion:128 Resp: 838451

Ion	Ratio	Lower	Upper
128	100		
64	53.3	41.6	62.4
130	33.0	25.6	38.4





#13

2-Methylphenol

Concen: 32.415 ng/ul

RT: 8.445 min Scan# 81

Delta R.T. -0.000 min

Lab File: BN024596.D

Acq: 28 Mar 2023 21:49

Instrument :

BNA_N

ClientSampleId :





SLCS703

Tgt Ion:108 Resp: 842134

Ion Ratio Lower Upper

108 100

107 86.7 70.9 106.3



#14

2,2'-oxybis(1-Chloropropane)

Concen: 31.888 ng/ul

RT: 8.534 min Scan# 892

Delta R.T. -0.000 min

Lab File: BN024596.D

Acq: 28 Mar 2023 21:49




Tgt Ion: 45 Resp: 1553985


Ion Ratio Lower Upper

45 100

77 11.4 9.0 13.4

79 8.7 6.9 10.3









#16
Acetophenone
Concen: 32.358 ng/ul
RT: 8.834 min Scan# 941
Delta R.T. -0.000 min
Lab File: BN024596.D
Acq: 28 Mar 2023 21:49




Instrument :
BNA_N
ClientSampleId :
SLCS703


Tgt Ion:105 Resp: 1353884
Ion Ratio Lower Upper
105 100
77 77.5 64.3 96.5
51 32.4 27.3 40.9



#17
N-Nitroso-di-n-propylamine
Concen: 32.729 ng/ul
RT: 8.822 min Scan# 941
Delta R.T. 0.006 min
Lab File: BN024596.D
Acq: 28 Mar 2023 21:49

Tgt Ion: 70 Resp: 720678
Ion Ratio Lower Upper
70 100
42 72.0 57.5 86.3
101 9.4 7.8 11.8
130 20.1 15.7 23.5









#18
4-Methylphenol
Concen: 32.324 ng/ul
RT: 8.781 min Scan# 91
Delta R.T. 0.006 min
Lab File: BN024596.D
Acq: 28 Mar 2023 21:49




Instrument :
BNA_N
ClientSampleId :
SLCS703


Tgt Ion:108 Resp: 932318
Ion Ratio Lower Upper
108 100
107 109.5 87.7 131.5



#19
Hexachloroethane
Concen: 32.206 ng/ul
RT: 9.087 min Scan# 986
Delta R.T. -0.000 min
Lab File: BN024596.D
Acq: 28 Mar 2023 21:49

Tgt Ion:117 Resp: 329736
Ion Ratio Lower Upper
117 100
201 94.8 71.9 107.9
199 57.0 45.2 67.8









#22
Nitrobenzene
Concen: 32.451 ng/ul
RT: 9.216 min Scan# 1008
Delta R.T. -0.000 min
Lab File: BN024596.D
Acq: 28 Mar 2023 21:49




Instrument :
BNA_N
ClientSampleId :
SLCS703


Tgt Ion: 77 Resp: 1065346
Ion Ratio Lower Upper
77 100
123 43.4 34.3 51.5
65 13.8 10.8 16.2



#23
Isophorone
Concen: 33.088 ng/ul
RT: 9.745 min Scan# 1098
Delta R.T. 0.006 min
Lab File: BN024596.D
Acq: 28 Mar 2023 21:49

Tgt Ion: 82 Resp: 2019010
Ion Ratio Lower Upper
82 100
95 7.0 5.4 8.0
138 16.8 12.6 18.8









#25
2-Nitrophenol
Concen: 33.394 ng/ul
RT: 9.928 min Scan# 1129
Delta R.T. -0.000 min
Lab File: BN024596.D
Acq: 28 Mar 2023 21:49




Instrument :
BNA_N
ClientSampleId :
SLCS703


Tgt Ion:139 Resp: 496661
Ion Ratio Lower Upper
139 100
65 56.8 46.3 69.5
109 29.5 24.6 37.0



#26
2,4-Dimethylphenol
Concen: 31.261 ng/ul
RT: 9.987 min Scan# 1139
Delta R.T. -0.000 min
Lab File: BN024596.D
Acq: 28 Mar 2023 21:49

Tgt Ion:107 Resp: 965740
Ion Ratio Lower Upper
107 100
121 50.7 40.7 61.1
122 85.9 71.2 106.8









#27
 Bis(2-Chloroethoxy)methane
 Concen: 31.704 ng/ul
 RT: 10.228 min Scan# 1179
 Delta R.T. 0.006 min
 Lab File: BN024596.D
 Acq: 28 Mar 2023 21:49




Instrument :
 BNA_N
 ClientSampleId :
 SLCS703


Tgt Ion: 93 Resp: 1241514
 Ion Ratio Lower Upper
 93 100
 95 32.6 26.4 39.6
 123 10.9 8.9 13.3



#29
 2,4-Dichlorophenol
 Concen: 32.476 ng/ul
 RT: 10.463 min Scan# 1220
 Delta R.T. -0.000 min
 Lab File: BN024596.D
 Acq: 28 Mar 2023 21:49

Tgt Ion:162 Resp: 818472
 Ion Ratio Lower Upper
 162 100
 164 65.9 50.5 75.7
 98 39.4 28.6 42.8









#30
Naphthalene
Concen: 31.586 ng/ul
RT: 10.869 min Scan# 1289
Delta R.T. -0.000 min
Lab File: BN024596.D
Acq: 28 Mar 2023 21:49




Instrument :
BNA_N
ClientSampleId :
SLCS703


Tgt Ion:128 Resp: 2802893
Ion Ratio Lower Upper
128 100
129 11.0 8.5 12.7
127 13.0 10.4 15.6



#32
4-Chloroaniline
Concen: 30.279 ng/ul
RT: 10.981 min Scan# 1308
Delta R.T. -0.000 min
Lab File: BN024596.D
Acq: 28 Mar 2023 21:49

Tgt Ion:127 Resp: 1162475
Ion Ratio Lower Upper
127 100
129 32.0 25.5 38.3









#33
Hexachlorobutadiene
Concen: 31.695 ng/ul
RT: 11.151 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN024596.D
Acq: 28 Mar 2023 21:49




Instrument :
BNA_N
ClientSampleId :
SLCS703


Tgt Ion:225 Resp: 460030
Ion Ratio Lower Upper
225 100
223 61.2 50.2 75.4
227 63.5 49.6 74.4



#34
Caprolactam
Concen: 18.139 ng/ul
RT: 11.763 min Scan# 1441
Delta R.T. 0.006 min
Lab File: BN024596.D
Acq: 28 Mar 2023 21:49

Tgt Ion:113 Resp: 158102
Ion Ratio Lower Upper
113 100
55 210.6 162.2 243.2
56 150.4 121.6 182.4









#35
4-Chloro-3-methylphenol
Concen: 33.424 ng/ul
RT: 12.098 min Scan# 1498
Delta R.T. -0.000 min
Lab File: BN024596.D
Acq: 28 Mar 2023 21:49




Instrument :
BNA_N
ClientSampleId :
SLCS703


Tgt Ion:107 Resp: 963044
Ion Ratio Lower Upper
107 100
144 24.8 21.1 31.7
142 77.0 61.0 91.4



#36
2-Methylnaphthalene
Concen: 31.705 ng/ul
RT: 12.475 min Scan# 1562
Delta R.T. -0.000 min
Lab File: BN024596.D
Acq: 28 Mar 2023 21:49

Tgt Ion:142 Resp: 1883536
Ion Ratio Lower Upper
142 100
141 88.9 70.2 105.4









#37
1-Methylnaphthalene
Concen: 33.243 ng/ul
RT: 12.692 min Scan# 1599
Delta R.T. -0.000 min
Lab File: BN024596.D
Acq: 28 Mar 2023 21:49




Instrument :
BNA_N
ClientSampleId :
SLCS703


Tgt Ion:142 Resp: 1956791
Ion Ratio Lower Upper
142 100
141 92.1 74.2 111.4
116 4.1 3.0 4.6



#39
1,2,4,5-Tetrachlorobenzene
Concen: 31.229 ng/ul
RT: 12.839 min Scan# 1624
Delta R.T. -0.000 min
Lab File: BN024596.D
Acq: 28 Mar 2023 21:49

Tgt Ion:216 Resp: 948483
Ion Ratio Lower Upper
216 100
214 74.5 61.0 91.4
179 21.0 15.7 23.5
108 18.2 14.0 21.0





#40

Hexachlorocyclopentadiene

Concen: 29.577 ng/ul

RT: 12.822 min Scan# 1621

Delta R.T. 0.006 min

Lab File: BN024596.D

Acq: 28 Mar 2023 21:49

Instrument :

BNA_N

ClientSampleId :

SLCS703





Tgt Ion:237 Resp: 565297

Ion Ratio Lower Upper

237 100

235 64.1 50.0 75.0

272 12.4 9.5 14.3



#41

2,4,6-Trichlorophenol

Concen: 32.213 ng/ul

RT: 13.075 min Scan# 1664

Delta R.T. -0.000 min

Lab File: BN024596.D

Acq: 28 Mar 2023 21:49




Tgt Ion:196 Resp: 651832


Ion Ratio Lower Upper

196 100

198 95.1 75.5 113.3

200 30.4 24.6 37.0





#42

2,4,5-Trichlorophenol

Concen: 33.258 ng/ul

RT: 13.151 min Scan# 1

Delta R.T. 0.006 min

Lab File: BN024596.D

Acq: 28 Mar 2023 21:49

Instrument :

BNA_N

ClientSampleId :

SLCS703





Tgt Ion:196 Resp: 739639

Ion Ratio Lower Upper

196 100

198 93.5 75.4 113.0

200 30.4 24.6 37.0



#43

1,1'-Biphenyl

Concen: 30.858 ng/ul

RT: 13.480 min Scan# 1733

Delta R.T. 0.006 min

Lab File: BN024596.D

Acq: 28 Mar 2023 21:49




Tgt Ion:154 Resp: 2511960


Ion Ratio Lower Upper

154 100

153 40.1 32.2 48.4

76 14.1 11.7 17.5





#44

2-Chloronaphthalene

Concen: 31.340 ng/ul

RT: 13.522 min Scan# 1740

Delta R.T. -0.000 min

Lab File: BN024596.D

Acq: 28 Mar 2023 21:49

Instrument :

BNA_N

ClientSampleId :

SLCS703





Tgt Ion:162 Resp: 1991643

Ion Ratio Lower Upper

162 100

164 33.0 26.1 39.1

127 39.5 31.1 46.7



#45

2-Nitroaniline

Concen: 34.858 ng/ul

RT: 13.728 min Scan# 1775

Delta R.T. -0.000 min

Lab File: BN024596.D

Acq: 28 Mar 2023 21:49




Tgt Ion: 65 Resp: 688262


Ion Ratio Lower Upper

65 100

92 63.8 52.6 78.8

138 96.1 81.4 122.2









#47
Dimethylphthalate
Concen: 31.907 ng/ul
RT: 14.098 min Scan# 1838
Delta R.T. -0.000 min
Lab File: BN024596.D
Acq: 28 Mar 2023 21:49




Instrument :
BNA_N
ClientSampleId :
SLCS703


Tgt Ion:163 Resp: 2552942
Ion Ratio Lower Upper
163 100
194 4.0 3.2 4.8
164 9.7 8.2 12.2



#48
2,6-Dinitrotoluene
Concen: 35.215 ng/ul
RT: 14.222 min Scan# 1859
Delta R.T. 0.006 min
Lab File: BN024596.D
Acq: 28 Mar 2023 21:49

Tgt Ion:165 Resp: 533704
Ion Ratio Lower Upper
165 100
89 57.8 47.8 71.6
121 21.1 17.3 25.9









#50
Acenaphthylene
Concen: 32.038 ng/ul
RT: 14.369 min Scan# 1884
Delta R.T. 0.006 min
Lab File: BN024596.D
Acq: 28 Mar 2023 21:49




Instrument :
BNA_N
ClientSampleId :
SLCS703


Tgt Ion:152 Resp: 3146951
Ion Ratio Lower Upper
152 100
151 19.4 15.5 23.3
153 13.0 10.3 15.5



#51
3-Nitroaniline
Concen: 35.277 ng/ul
RT: 14.551 min Scan# 1915
Delta R.T. 0.006 min
Lab File: BN024596.D
Acq: 28 Mar 2023 21:49

Tgt Ion:138 Resp: 568291
Ion Ratio Lower Upper
138 100
108 10.3 8.8 13.2
92 114.6 96.1 144.1









#52
 Acenaphthene
 Concen: 31.510 ng/ul
 RT: 14.704 min Scan# 1941
 Delta R.T. -0.000 min
 Lab File: BN024596.D
 Acq: 28 Mar 2023 21:49

Instrument : BNA_N
 ClientSampleId : SLCS703

Tgt Ion:153 Resp: 2156853




Ion	Ratio	Lower	Upper
153	100		
152	47.4	36.6	55.0
154	89.8	67.7	101.5




#53
 2,4-Dinitrophenol
 Concen: 31.975 ng/ul
 RT: 14.751 min Scan# 1949
 Delta R.T. -0.000 min
 Lab File: BN024596.D
 Acq: 28 Mar 2023 21:49

Tgt Ion:184 Resp: 329173

Ion	Ratio	Lower	Upper
184	100		
63	75.4	59.1	88.7
154	60.8	49.2	73.8









#55
4-Nitrophenol
Concen: 32.784 ng/ul
RT: 14.851 min Scan# 1966
Delta R.T. 0.006 min
Lab File: BN024596.D
Acq: 28 Mar 2023 21:49




Instrument :
BNA_N
ClientSampleId :
SLCS703


Tgt Ion:109 Resp: 413634
Ion Ratio Lower Upper
109 100
139 136.0 108.4 162.6
65 146.9 121.1 181.7



#56
Dibenzofuran
Concen: 31.614 ng/ul
RT: 15.039 min Scan# 1998
Delta R.T. -0.000 min
Lab File: BN024596.D
Acq: 28 Mar 2023 21:49

Tgt Ion:168 Resp: 3011201
Ion Ratio Lower Upper
168 100
139 39.1 30.0 45.0





#57

2,4-Dinitrotoluene

Concen: 34.663 ng/ul

RT: 15.004 min Scan# 1992

Delta R.T. 0.006 min

Lab File: BN024596.D

Acq: 28 Mar 2023 21:49

Instrument :

BNA_N

ClientSampleId :

SLCS703





Tgt Ion:165 Resp: 769633

Ion Ratio Lower Upper

165 100

63 49.9 42.2 63.4

182 2.9 2.3 3.5



#58

2,3,4,6-Tetrachlorophenol

Concen: 33.463 ng/ul

RT: 15.263 min Scan# 2036

Delta R.T. -0.000 min

Lab File: BN024596.D

Acq: 28 Mar 2023 21:49

Tgt Ion:232 Resp: 620165




Ion Ratio Lower Upper


232 100

131 47.4 36.5 54.7

130 2.4 1.8 2.8

166 31.6 24.4 36.6





#59

Diethylphthalate

Concen: 32.357 ng/ul

RT: 15.457 min Scan# 2069

Delta R.T. -0.000 min

Lab File: BN024596.D

Acq: 28 Mar 2023 21:49

Instrument :

BNA_N

ClientSampleId :

SLCS703





Tgt Ion:149 Resp: 2533423

Ion Ratio Lower Upper

149 100

177 21.4 17.5 26.3

150 12.9 9.8 14.6



#61

Fluorene

Concen: 31.778 ng/ul

RT: 15.692 min Scan# 2109

Delta R.T. 0.006 min

Lab File: BN024596.D

Acq: 28 Mar 2023 21:49




Tgt Ion:166 Resp: 2426915


Ion Ratio Lower Upper

166 100

165 94.9 78.6 118.0

167 12.9 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 31.615 ng/ul
RT: 15.680 min Scan# 2107
Delta R.T. -0.000 min
Lab File: BN024596.D
Acq: 28 Mar 2023 21:49




Instrument :
BNA_N
ClientSampleId :
SLCS703


Tgt Ion:204 Resp: 1145569
Ion Ratio Lower Upper
204 100
206 32.0 25.9 38.9
141 61.1 49.9 74.9



#63
4-Nitroaniline
Concen: 38.408 ng/ul
RT: 15.716 min Scan# 2113
Delta R.T. 0.006 min
Lab File: BN024596.D
Acq: 28 Mar 2023 21:49

Tgt Ion:138 Resp: 609535
Ion Ratio Lower Upper
138 100
92 52.2 41.5 62.3
108 77.4 63.0 94.4









#66
 4,6-Dinitro-2-methylphenol
 Concen: 32.685 ng/ul
 RT: 15.769 min Scan# 2121
 Delta R.T. 0.006 min
 Lab File: BN024596.D
 Acq: 28 Mar 2023 21:49




Instrument :
 BNA_N
 ClientSampleId :
 SLCS703


Tgt Ion:198 Resp: 463548
 Ion Ratio Lower Upper
 198 100
 182 3.7 3.9 5.9#
 77 25.0 22.2 33.2



#67
 N-Nitrosodiphenylamine
 Concen: 32.707 ng/ul
 RT: 15.892 min Scan# 2143
 Delta R.T. -0.000 min
 Lab File: BN024596.D
 Acq: 28 Mar 2023 21:49

Tgt Ion:169 Resp: 2129613
 Ion Ratio Lower Upper
 169 100
 168 62.6 54.3 81.5
 167 34.5 29.4 44.0









#68
4-Bromophenyl-phenylether
Concen: 32.470 ng/ul
RT: 16.574 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN024596.D
Acq: 28 Mar 2023 21:49




Instrument :
BNA_N
ClientSampleId :
SLCS703


Tgt Ion:248 Resp: 705650
Ion Ratio Lower Upper
248 100
250 95.1 74.5 111.7
141 75.6 60.7 91.1



#69
Hexachlorobenzene
Concen: 31.952 ng/ul
RT: 16.692 min Scan# 2279
Delta R.T. -0.000 min
Lab File: BN024596.D
Acq: 28 Mar 2023 21:49

Tgt Ion:284 Resp: 817213
Ion Ratio Lower Upper
284 100
142 33.7 28.5 42.7
249 29.7 25.5 38.3





#70

Atrazine

Concen: 25.659 ng/ul

RT: 16.845 min Scan# 2305

Delta R.T. -0.000 min

Lab File: BN024596.D

Acq: 28 Mar 2023 21:49

Instrument :

BNA_N

ClientSampleId :

SLCS703





Tgt Ion:200 Resp: 570796

Ion Ratio Lower Upper

200 100

173 29.8 21.8 32.6

215 50.0 38.2 57.2



#71

Pentachlorophenol

Concen: 32.463 ng/ul

RT: 17.033 min Scan# 2337

Delta R.T. -0.000 min

Lab File: BN024596.D

Acq: 28 Mar 2023 21:49




Tgt Ion:266 Resp: 507554


Ion Ratio Lower Upper

266 100

264 65.5 52.5 78.7

268 62.0 54.4 81.6





#72

Phenanthrene

Concen: 32.334 ng/ul

RT: 17.433 min Scan# 2405

Delta R.T. 0.006 min

Lab File: BN024596.D

Acq: 28 Mar 2023 21:49

Instrument :

BNA_N

ClientSampleId :

SLCS703





Tgt Ion:178 Resp: 3945144

Ion Ratio Lower Upper

178 100

179 14.9 12.7 19.1

176 18.9 15.5 23.3



#74

Anthracene

Concen: 31.756 ng/ul

RT: 17.521 min Scan# 2420

Delta R.T. -0.000 min

Lab File: BN024596.D

Acq: 28 Mar 2023 21:49




Tgt Ion:178 Resp: 3892583


Ion Ratio Lower Upper

178 100

179 16.1 11.9 17.9

176 19.7 14.1 21.1





#75

1,2,3,4-Tetrachlorobenzene

Concen: 30.299 ng/uL

RT: 13.445 min Scan# 1727

Delta R.T. -0.000 min

Lab File: BN024596.D

Acq: 28 Mar 2023 21:49

Instrument :

BNA_N

ClientSampleId :

SLCS703





Tgt Ion:216 Resp: 943186

Ion Ratio Lower Upper

216 100

214 80.6 62.7 94.1

218 49.2 38.0 57.0



#76

Pentachlorobenzene

Concen: 31.782 ng/uL

RT: 14.957 min Scan# 1984

Delta R.T. -0.000 min

Lab File: BN024596.D

Acq: 28 Mar 2023 21:49




Tgt Ion:250 Resp: 961040


Ion Ratio Lower Upper

250 100

248 59.3 50.5 75.7

252 65.6 50.9 76.3









#77
 Carbazole
 Concen: 33.638 ng/ul
 RT: 17.792 min Scan# 2466
 Delta R.T. -0.000 min
 Lab File: BN024596.D
 Acq: 28 Mar 2023 21:49

Instrument :
 BNA_N
 ClientSampleId :
 SLCS703

Tgt Ion:167 Resp: 3707335




Ion	Ratio	Lower	Upper
167	100		
166	21.2	17.4	26.2
139	13.4	10.5	15.7




#78
 Di-n-butylphthalate
 Concen: 34.439 ng/ul
 RT: 18.351 min Scan# 2561
 Delta R.T. -0.000 min
 Lab File: BN024596.D
 Acq: 28 Mar 2023 21:49

Tgt Ion:149 Resp: 4301107

Ion	Ratio	Lower	Upper
149	100		
150	9.2	7.4	11.2
104	5.1	4.1	6.1





#80

Fluoranthene

Concen: 33.116 ng/ul

RT: 19.445 min Scan# 21

Delta R.T. 0.006 min

Lab File: BN024596.D

Acq: 28 Mar 2023 21:49

Instrument :

BNA_N

ClientSampleId :

SLCS703





Tgt Ion:202 Resp: 4517201

Ion Ratio Lower Upper

202 100

101 12.5 10.3 15.5

100 9.6 8.2 12.2



#82

Pyrene

Concen: 32.627 ng/ul

RT: 19.810 min Scan# 2809

Delta R.T. 0.006 min

Lab File: BN024596.D

Acq: 28 Mar 2023 21:49




Tgt Ion:202 Resp: 4691666


Ion Ratio Lower Upper

202 100

101 14.2 12.5 18.7

100 11.7 10.1 15.1





#83

Butylbenzylphthalate

Concen: 35.085 ng/ul

RT: 20.698 min Scan# 2959

Delta R.T. 0.006 min

Lab File: BN024596.D

Acq: 28 Mar 2023 21:49

Instrument :

BNA_N

ClientSampleId :

SLCS703





Tgt Ion:149 Resp: 1920988

Ion Ratio Lower Upper

149 100

91 76.7 59.7 89.5

206 20.3 15.4 23.2



#84

3,3'-Dichlorobenzidine

Concen: 32.520 ng/ul

RT: 21.492 min Scan# 3095

Delta R.T. 0.006 min

Lab File: BN024596.D

Acq: 28 Mar 2023 21:49




Tgt Ion:252 Resp: 1512255


Ion Ratio Lower Upper

252 100

254 65.1 50.0 75.0

126 13.4 9.8 14.8









#85
Benzo(a)anthracene
Concen: 31.864 ng/ul
RT: 21.562 min Scan# 3106
Delta R.T. 0.006 min
Lab File: BN024596.D
Acq: 28 Mar 2023 21:49




Instrument :
BNA_N
ClientSampleId :
SLCS703


Tgt Ion:228 Resp: 4477698
Ion Ratio Lower Upper
228 100
229 19.8 15.8 23.8
226 28.3 22.7 34.1



#86
Bis(2-ethylhexyl)phthalate
Concen: 35.491 ng/ul
RT: 21.480 min Scan# 3093
Delta R.T. 0.011 min
Lab File: BN024596.D
Acq: 28 Mar 2023 21:49

Tgt Ion:149 Resp: 2709391
Ion Ratio Lower Upper
149 100
167 27.0 21.6 32.4
279 4.4 2.7 4.1#









#87
Chrysene
Concen: 32.528 ng/ul
RT: 21.615 min Scan# 3115
Delta R.T. 0.006 min
Lab File: BN024596.D
Acq: 28 Mar 2023 21:49




Instrument :
BNA_N
ClientSampleId :
SLCS703


Tgt Ion:228 Resp: 4325305
Ion Ratio Lower Upper
228 100
226 30.0 24.8 37.2
229 19.4 16.4 24.6



#89
Di-n-octyl phthalate
Concen: 36.547 ng/ul
RT: 22.439 min Scan# 3256
Delta R.T. 0.017 min
Lab File: BN024596.D
Acq: 28 Mar 2023 21:49

Tgt Ion:149 Resp: 4699073





#90

Benzo(b)fluoranthene

Concen: 33.896 ng/ul

RT: 23.304 min Scan# 3411

Delta R.T. 0.017 min

Lab File: BN024596.D

Acq: 28 Mar 2023 21:49

Instrument :

BNA_N

ClientSampleId :

SLCS703





Tgt Ion:252 Resp: 4710992

Ion Ratio Lower Upper

252 100

253 21.2 15.8 23.8

125 11.4 8.6 13.0



#91

Benzo(k)fluoranthene

Concen: 34.367 ng/ul

RT: 23.356 min Scan# 3412

Delta R.T. 0.017 min

Lab File: BN024596.D

Acq: 28 Mar 2023 21:49




Tgt Ion:252 Resp: 4559795


Ion Ratio Lower Upper

252 100

253 23.6 17.5 26.3

125 11.9 9.2 13.8





#93

Benzo(a)pyrene

Concen: 33.916 ng/ul

RT: 23.962 min Scan# 3

Delta R.T. 0.023 min

Lab File: BN024596.D

Acq: 28 Mar 2023 21:49

Instrument :

BNA_N

ClientSampleId :

SLCS703





Tgt Ion:252 Resp: 4127688

Ion Ratio Lower Upper

252 100

253 20.8 18.2 27.4

125 12.1 10.2 15.4



#94

Indeno(1,2,3-cd)pyrene

Concen: 33.465 ng/ul

RT: 26.680 min Scan# 3977

Delta R.T. 0.053 min

Lab File: BN024596.D

Acq: 28 Mar 2023 21:49




Tgt Ion:276 Resp: 5190034


Ion Ratio Lower Upper

276 100

138 27.8 22.2 33.2

277 25.0 20.3 30.5









#95
Dibenzo(a,h)anthracene
Concen: 32.258 ng/ul
RT: 26.697 min Scan# 3971
Delta R.T. 0.053 min
Lab File: BN024596.D
Acq: 28 Mar 2023 21:49

Instrument :
BNA_N
ClientSampleId :
SLCS703

Tgt Ion:278 Resp: 4149118
Ion Ratio Lower Upper
278 100
139 17.9 14.2 21.4
279 25.4 19.6 29.4



#96
Benzo(g,h,i)perylene
Concen: 31.930 ng/ul
RT: 27.480 min Scan# 4113
Delta R.T. 0.059 min
Lab File: BN024596.D
Acq: 28 Mar 2023 21:49

Tgt Ion:276 Resp: 4114548
Ion Ratio Lower Upper
276 100
138 24.2 19.5 29.3
277 24.5 18.4 27.6

