

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032923\
 Data File : BN024600.D
 Acq On : 29 Mar 2023 08:34
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020265

Quant Time: Mar 30 02:28:50 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032923.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 30 02:26:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.993	152	346099	20.000	ng/ul	0.00
20) Naphthalene-d8	10.810	136	1534367	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.640	164	972606	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.387	188	2076544	20.000	ng/ul	0.00
79) Chrysene-d12	21.581	240	1949449	20.000	ng/ul	0.00
88) Perylene-d12	24.063	264	2182063	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.364	96	76348	8.553	ng/uL	0.00
4) Pyridine-d5	3.793	84	546221	21.191	ng/ul	0.00
7) Phenol-d5	7.146	99	671420	20.499	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.317	67	453820	21.456	ng/ul	0.00
11) 2-Chlorophenol-d4	7.522	132	513579	21.150	ng/ul	0.00
15) 4-Methylphenol-d8	8.705	113	534707	20.413	ng/ul	0.00
21) Nitrobenzene-d5	9.164	128	258106	21.255	ng/ul	0.00
24) 2-Nitrophenol-d4	9.893	143	277077	21.080	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.428	165	514067	21.534	ng/ul	0.00
31) 4-Chloroaniline-d4	10.952	131	772299	21.202	ng/ul	0.00
46) Dimethylphthalate-d6	14.046	166	1582084	21.153	ng/ul	0.00
49) Acenaphthylene-d8	14.334	160	1848128	21.700	ng/ul	0.00
54) 4-Nitrophenol-d4	14.834	143	325140	20.771	ng/ul	0.00
60) Fluorene-d10	15.628	176	1341708	20.907	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.746	200	271132	19.824	ng/ul	0.00
73) Anthracene-d10	17.487	188	2097083	21.472	ng/ul	0.00
81) Pyrene-d10	19.781	212	2461110	21.834	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.898	264	2411959	21.815	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.399	88	81036	8.425	ng/uL	97
5) Pyridine	3.811	79	562689	20.972	ng/ul	100
6) Benzaldehyde	7.128	77	387763	25.938	ng/ul	97
8) Phenol	7.175	94	706660	20.784	ng/ul	99
10) Bis(2-Chloroethyl)ether	7.411	93	571095	21.301	ng/ul	96
12) 2-Chlorophenol	7.552	128	528438	20.859	ng/ul	99
13) 2-Methylphenol	8.440	108	523433	20.616	ng/ul	100
14) 2,2'-oxybis(1-Chloropr...	8.528	45	1024966	21.521	ng/ul	100
16) Acetophenone	8.822	105	878257	21.478	ng/ul	100
17) N-Nitroso-di-n-propyla...	8.811	70	461744	21.457	ng/ul	99
18) 4-Methylphenol	8.769	108	582264	20.656	ng/ul	97
19) Hexachloroethane	9.081	117	206331	20.621	ng/ul	98
22) Nitrobenzene	9.205	77	676735	21.455	ng/ul	99
23) Isophorone	9.734	82	1252409	21.362	ng/ul	99
25) 2-Nitrophenol	9.922	139	308432	21.585	ng/ul	97
26) 2,4-Dimethylphenol	9.975	107	641595	21.616	ng/ul	96
27) Bis(2-Chloroethoxy)met...	10.216	93	793280	21.085	ng/ul	99
29) 2,4-Dichlorophenol	10.458	162	513489	21.206	ng/ul	97
30) Naphthalene	10.863	128	1795691	21.062	ng/ul	100
32) 4-Chloroaniline	10.975	127	792898	21.495	ng/ul	98
33) Hexachlorobutadiene	11.146	225	299914	21.507	ng/ul	97
34) Caprolactam	11.752	113	92346	11.027	ng/ul	97
35) 4-Chloro-3-methylphenol	12.093	107	599152	21.643	ng/ul	97
36) 2-Methylnaphthalene	12.469	142	1228587	21.524	ng/ul	98


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032923\
 Data File : BN024600.D
 Acq On : 29 Mar 2023 08:34
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020265

Quant Time: Mar 30 02:28:50 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032923.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 30 02:26:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.687	142	1213169	21.451	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.834	216	606154	21.025	ng/ul	98
40) Hexachlorocyclopentadiene	12.816	237	360153	19.851	ng/ul	99
41) 2,4,6-Trichlorophenol	13.075	196	411348	21.415	ng/ul	94
42) 2,4,5-Trichlorophenol	13.146	196	449973	21.314	ng/ul	99
43) 1,1'-Biphenyl	13.475	154	1653101	21.393	ng/ul	99
44) 2-Chloronaphthalene	13.516	162	1273421	21.109	ng/ul	99
45) 2-Nitroaniline	13.722	65	407085	21.719	ng/ul	95
47) Dimethylphthalate	14.093	163	1619080	21.317	ng/ul	100
48) 2,6-Dinitrotoluene	14.216	165	314313	21.847	ng/ul	98
50) Acenaphthylene	14.363	152	2049505	21.980	ng/ul	99
51) 3-Nitroaniline	14.546	138	331716	21.692	ng/ul	97
52) Acenaphthene	14.704	153	1381243	21.258	ng/ul	95
53) 2,4-Dinitrophenol	14.746	184	176609	18.072	ng/ul	90
55) 4-Nitrophenol	14.846	109	259724	21.686	ng/ul	95
56) Dibenzofuran	15.040	168	1891015	20.915	ng/ul	100
57) 2,4-Dinitrotoluene	14.999	165	457269	21.696	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	15.263	232	374644	21.295	ng/ul	98
59) Diethylphthalate	15.457	149	1558409	20.968	ng/ul	99
61) Fluorene	15.687	166	1576544	21.747	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.675	204	743124	21.604	ng/ul	98
63) 4-Nitroaniline	15.704	138	338374	22.461	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.763	198	276844	20.468	ng/ul	96
67) N-Nitrosodiphenylamine	15.893	169	1328561	21.395	ng/ul	96
68) 4-Bromophenyl-phenylether	16.575	248	443292	21.388	ng/ul	94
69) Hexachlorobenzene	16.693	284	512780	21.023	ng/ul	98
70) Atrazine	16.840	200	462566	21.803	ng/ul	98
71) Pentachlorophenol	17.034	266	301690	20.233	ng/ul	95
72) Phenanthrene	17.428	178	2534504	21.781	ng/ul	99
74) Anthracene	17.522	178	2567731	21.965	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.440	216	635488	21.406	ng/ul	97
76) Pentachlorobenzene	14.957	250	610749	21.179	ng/ul	99
77) Carbazole	17.792	167	2339277	22.256	ng/ul	98
78) Di-n-butylphthalate	18.351	149	2593047	21.771	ng/ul	99
80) Fluoranthene	19.445	202	2906408	21.743	ng/ul	98
82) Pyrene	19.804	202	3005644	21.330	ng/ul	100
83) Butylbenzylphthalate	20.704	149	1136068	21.174	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.498	252	979162	21.487	ng/ul	99
85) Benzo(a)anthracene	21.563	228	2953375	21.446	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.480	149	1666101	22.271	ng/ul	100
87) Chrysene	21.622	228	2881884	22.116	ng/ul	98
89) Di-n-octyl phthalate	22.433	149	2810671	21.263	ng/ul	100
90) Benzo(b)fluoranthene	23.298	252	3120931	21.843	ng/ul	99
91) Benzo(k)fluoranthene	23.351	252	2909753	21.332	ng/ul	99
93) Benzo(a)pyrene	23.951	252	2708461	21.647	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.645	276	3451605	21.648	ng/ul	100
95) Dibenzo(a,h)anthracene	26.663	278	2816749	21.302	ng/ul	98
96) Benzo(g,h,i)perylene	27.439	276	2827345	21.342	ng/ul	99





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 8.425 ng/uL
RT: 3.399 min Scan# 19
Delta R.T. 0.000 min
Lab File: BN024600.D
Acq: 29 Mar 2023 08:34




Instrument :
BNA_N
ClientSampleId :
SSTD020265


Tgt Ion: 88 Resp: 81036
Ion Ratio Lower Upper
88 100
43 39.4 29.0 43.6
58 86.0 67.5 101.3



#5
Pyridine
Concen: 20.972 ng/uL
RT: 3.811 min Scan# 89
Delta R.T. 0.000 min
Lab File: BN024600.D
Acq: 29 Mar 2023 08:34

Tgt Ion: 79 Resp: 562689
Ion Ratio Lower Upper
79 100
52 94.7 76.0 114.0
51 43.3 35.0 52.4









#6
Benzaldehyde
Concen: 25.938 ng/ul
RT: 7.128 min Scan# 653
Delta R.T. 0.000 min
Lab File: BN024600.D
Acq: 29 Mar 2023 08:34




Instrument :
BNA_N
ClientSampleId :
SSTD020265


Tgt Ion: 77 Resp: 387763
Ion Ratio Lower Upper
77 100
105 95.9 74.0 111.0
106 91.8 71.7 107.5



#8
Phenol
Concen: 20.784 ng/ul
RT: 7.175 min Scan# 661
Delta R.T. 0.000 min
Lab File: BN024600.D
Acq: 29 Mar 2023 08:34

Tgt Ion: 94 Resp: 706660
Ion Ratio Lower Upper
94 100
65 29.9 23.7 35.5
66 38.7 29.9 44.9









#10
Bis(2-Chloroethyl)ether
Concen: 21.301 ng/ul
RT: 7.411 min Scan# 701
Delta R.T. 0.000 min
Lab File: BN024600.D
Acq: 29 Mar 2023 08:34




Instrument :
BNA_N
ClientSampleId :
SSTD020265


Tgt Ion: 93 Resp: 571095
Ion Ratio Lower Upper
93 100
63 85.4 64.9 97.3
95 32.8 25.2 37.8



#12
2-Chlorophenol
Concen: 20.859 ng/ul
RT: 7.552 min Scan# 725
Delta R.T. 0.000 min
Lab File: BN024600.D
Acq: 29 Mar 2023 08:34

Tgt Ion:128 Resp: 528438
Ion Ratio Lower Upper
128 100
64 53.2 41.6 62.4
130 31.9 25.6 38.4









#13
2-Methylphenol
Concen: 20.616 ng/ul
RT: 8.440 min Scan# 876
Delta R.T. 0.000 min
Lab File: BN024600.D
Acq: 29 Mar 2023 08:34




Instrument :
BNA_N
ClientSampleId :
SSTD020265


Tgt Ion:108 Resp: 523433
Ion Ratio Lower Upper
108 100
107 89.0 70.9 106.3



#14
2,2'-oxybis(1-Chloropropane)
Concen: 21.521 ng/ul
RT: 8.528 min Scan# 891
Delta R.T. 0.000 min
Lab File: BN024600.D
Acq: 29 Mar 2023 08:34

Tgt Ion: 45 Resp: 1024966
Ion Ratio Lower Upper
45 100
77 11.1 9.0 13.4
79 8.5 6.9 10.3









#16
Acetophenone
Concen: 21.478 ng/ul
RT: 8.822 min Scan# 941
Delta R.T. 0.000 min
Lab File: BN024600.D
Acq: 29 Mar 2023 08:34




Instrument :
BNA_N
ClientSampleId :
SSTD020265


Tgt Ion: 105 Resp: 878257
Ion Ratio Lower Upper
105 100
77 80.2 64.3 96.5
51 33.9 27.3 40.9



#17
N-Nitroso-di-n-propylamine
Concen: 21.457 ng/ul
RT: 8.811 min Scan# 939
Delta R.T. 0.000 min
Lab File: BN024600.D
Acq: 29 Mar 2023 08:34

Tgt Ion: 70 Resp: 461744
Ion Ratio Lower Upper
70 100
42 73.3 57.5 86.3
101 9.6 7.8 11.8
130 19.6 15.7 23.5









#18
4-Methylphenol
Concen: 20.656 ng/ul
RT: 8.769 min Scan# 911
Delta R.T. 0.000 min
Lab File: BN024600.D
Acq: 29 Mar 2023 08:34




Instrument :
BNA_N
ClientSampleId :
SSTD020265


Tgt Ion:108 Resp: 582264
Ion Ratio Lower Upper
108 100
107 106.9 87.7 131.5



#19
Hexachloroethane
Concen: 20.621 ng/ul
RT: 9.081 min Scan# 985
Delta R.T. 0.000 min
Lab File: BN024600.D
Acq: 29 Mar 2023 08:34

Tgt Ion:117 Resp: 206331
Ion Ratio Lower Upper
117 100
201 91.9 71.9 107.9
199 58.5 45.2 67.8









#22
Nitrobenzene
Concen: 21.455 ng/ul
RT: 9.205 min Scan# 1006
Delta R.T. 0.000 min
Lab File: BN024600.D
Acq: 29 Mar 2023 08:34




Instrument :
BNA_N
ClientSampleId :
SSTD020265


Tgt Ion: 77 Resp: 676735
Ion Ratio Lower Upper
77 100
123 42.4 34.3 51.5
65 13.6 10.8 16.2



#23
Isophorone
Concen: 21.362 ng/ul
RT: 9.734 min Scan# 1096
Delta R.T. 0.000 min
Lab File: BN024600.D
Acq: 29 Mar 2023 08:34

Tgt Ion: 82 Resp: 1252409
Ion Ratio Lower Upper
82 100
95 7.1 5.4 8.0
138 16.3 12.6 18.8









#25
2-Nitrophenol
Concen: 21.585 ng/ul
RT: 9.922 min Scan# 1128
Delta R.T. 0.000 min
Lab File: BN024600.D
Acq: 29 Mar 2023 08:34




Instrument :
BNA_N
ClientSampleId :
SSTD020265


Tgt Ion:139 Resp: 308432
Ion Ratio Lower Upper
139 100
65 55.4 46.3 69.5
109 29.7 24.6 37.0



#26
2,4-Dimethylphenol
Concen: 21.616 ng/ul
RT: 9.975 min Scan# 1137
Delta R.T. 0.000 min
Lab File: BN024600.D
Acq: 29 Mar 2023 08:34

Tgt Ion:107 Resp: 641595
Ion Ratio Lower Upper
107 100
121 50.3 40.7 61.1
122 84.0 71.2 106.8









#27
 Bis(2-Chloroethoxy)methane
 Concen: 21.085 ng/ul
 RT: 10.216 min Scan# 1178
 Delta R.T. 0.000 min
 Lab File: BN024600.D
 Acq: 29 Mar 2023 08:34




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020265


Tgt Ion: 93 Resp: 793280
 Ion Ratio Lower Upper
 93 100
 95 32.3 26.4 39.6
 123 11.3 8.9 13.3



#29
 2,4-Dichlorophenol
 Concen: 21.206 ng/ul
 RT: 10.458 min Scan# 1219
 Delta R.T. 0.000 min
 Lab File: BN024600.D
 Acq: 29 Mar 2023 08:34

Tgt Ion: 162 Resp: 513489
 Ion Ratio Lower Upper
 162 100
 164 65.4 50.5 75.7
 98 37.9 28.6 42.8









#30
Naphthalene
Concen: 21.062 ng/ul
RT: 10.863 min Scan# 1288
Delta R.T. 0.000 min
Lab File: BN024600.D
Acq: 29 Mar 2023 08:34




Instrument :
BNA_N
ClientSampleId :
SSTD020265


Tgt Ion:128 Resp: 1795691
Ion Ratio Lower Upper
128 100
129 10.8 8.5 12.7
127 13.1 10.4 15.6



#32
4-Chloroaniline
Concen: 21.495 ng/ul
RT: 10.975 min Scan# 1307
Delta R.T. 0.000 min
Lab File: BN024600.D
Acq: 29 Mar 2023 08:34

Tgt Ion:127 Resp: 792898
Ion Ratio Lower Upper
127 100
129 30.9 25.5 38.3





#33

Hexachlorobutadiene

Concen: 21.507 ng/uI

RT: 11.146 min Scan# 1336

Delta R.T. 0.000 min

Lab File: BN024600.D

Acq: 29 Mar 2023 08:34

Instrument :

BNA_N

ClientSampleId :

SSTD020265





Tgt Ion:225 Resp: 299914

Ion Ratio Lower Upper

225 100

223 59.9 50.2 75.4

227 63.8 49.6 74.4



#34

Caprolactam

Concen: 11.027 ng/uI

RT: 11.752 min Scan# 1439

Delta R.T. 0.000 min

Lab File: BN024600.D

Acq: 29 Mar 2023 08:34




Tgt Ion:113 Resp: 92346


Ion Ratio Lower Upper

113 100

55 204.5 162.2 243.2

56 145.4 121.6 182.4









#35
4-Chloro-3-methylphenol
Concen: 21.643 ng/ul
RT: 12.093 min Scan# 1497
Delta R.T. 0.000 min
Lab File: BN024600.D
Acq: 29 Mar 2023 08:34




Instrument :
BNA_N
ClientSampleId :
SSTD020265


Tgt Ion:107 Resp: 599152
Ion Ratio Lower Upper
107 100
144 25.2 21.1 31.7
142 78.8 61.0 91.4



#36
2-Methylnaphthalene
Concen: 21.524 ng/ul
RT: 12.469 min Scan# 1561
Delta R.T. 0.000 min
Lab File: BN024600.D
Acq: 29 Mar 2023 08:34

Tgt Ion:142 Resp: 1228587
Ion Ratio Lower Upper
142 100
141 89.7 70.2 105.4









#37
1-Methylnaphthalene
Concen: 21.451 ng/ul
RT: 12.687 min Scan# 1598
Delta R.T. 0.000 min
Lab File: BN024600.D
Acq: 29 Mar 2023 08:34




Instrument :
BNA_N
ClientSampleId :
SSTD020265


Tgt Ion:142 Resp: 1213169
Ion Ratio Lower Upper
142 100
141 91.8 74.2 111.4
116 3.9 3.0 4.6



#39
1,2,4,5-Tetrachlorobenzene
Concen: 21.025 ng/ul
RT: 12.834 min Scan# 1623
Delta R.T. 0.000 min
Lab File: BN024600.D
Acq: 29 Mar 2023 08:34

Tgt Ion:216 Resp: 606154
Ion Ratio Lower Upper
216 100
214 77.1 61.0 91.4
179 21.2 15.7 23.5
108 19.6 14.0 21.0









#40
Hexachlorocyclopentadiene
Concen: 19.851 ng/ul
RT: 12.816 min Scan# 1620
Delta R.T. 0.000 min
Lab File: BN024600.D
Acq: 29 Mar 2023 08:34




Instrument :
BNA_N
ClientSampleId :
SSTD020265


Tgt Ion:237 Resp: 360153
Ion Ratio Lower Upper
237 100
235 62.5 50.0 75.0
272 13.2 9.5 14.3



#41
2,4,6-Trichlorophenol
Concen: 21.415 ng/ul
RT: 13.075 min Scan# 1664
Delta R.T. 0.000 min
Lab File: BN024600.D
Acq: 29 Mar 2023 08:34

Tgt Ion:196 Resp: 411348
Ion Ratio Lower Upper
196 100
198 100.9 75.5 113.3
200 31.8 24.6 37.0





#42

2,4,5-Trichlorophenol

Concen: 21.314 ng/ul

RT: 13.146 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN024600.D

Acq: 29 Mar 2023 08:34

Instrument :

BNA_N

ClientSampleId :

SSTD020265





Tgt Ion:196 Resp: 449973

Ion Ratio Lower Upper

196 100

198 95.2 75.4 113.0

200 30.0 24.6 37.0



#43

1,1'-Biphenyl

Concen: 21.393 ng/ul

RT: 13.475 min Scan# 1732

Delta R.T. 0.000 min

Lab File: BN024600.D

Acq: 29 Mar 2023 08:34




Tgt Ion:154 Resp: 1653101


Ion Ratio Lower Upper

154 100

153 40.5 32.2 48.4

76 14.1 11.7 17.5









#44
2-Chloronaphthalene
Concen: 21.109 ng/ul
RT: 13.516 min Scan# 1739
Delta R.T. 0.000 min
Lab File: BN024600.D
Acq: 29 Mar 2023 08:34




Instrument :
BNA_N
ClientSampleId :
SSTD020265


Tgt Ion:162 Resp: 1273421
Ion Ratio Lower Upper
162 100
164 33.0 26.1 39.1
127 39.8 31.1 46.7



#45
2-Nitroaniline
Concen: 21.719 ng/ul
RT: 13.722 min Scan# 1774
Delta R.T. 0.000 min
Lab File: BN024600.D
Acq: 29 Mar 2023 08:34

Tgt Ion: 65 Resp: 407085
Ion Ratio Lower Upper
65 100
92 65.0 52.6 78.8
138 94.2 81.4 122.2









#47
Dimethylphthalate
Concen: 21.317 ng/ul
RT: 14.093 min Scan# 1837
Delta R.T. 0.000 min
Lab File: BN024600.D
Acq: 29 Mar 2023 08:34




Instrument :
BNA_N
ClientSampleId :
SSTD020265


Tgt Ion:163 Resp: 1619080
Ion Ratio Lower Upper
163 100
194 3.9 3.2 4.8
164 10.1 8.2 12.2



#48
2,6-Dinitrotoluene
Concen: 21.847 ng/ul
RT: 14.216 min Scan# 1858
Delta R.T. 0.000 min
Lab File: BN024600.D
Acq: 29 Mar 2023 08:34

Tgt Ion:165 Resp: 314313
Ion Ratio Lower Upper
165 100
89 62.2 47.8 71.6
121 21.7 17.3 25.9









#50
Acenaphthylene
Concen: 21.980 ng/ul
RT: 14.363 min Scan# 1883
Delta R.T. 0.000 min
Lab File: BN024600.D
Acq: 29 Mar 2023 08:34




Instrument :
BNA_N
ClientSampleId :
SSTD020265


Tgt Ion:152 Resp: 2049505
Ion Ratio Lower Upper
152 100
151 18.9 15.5 23.3
153 13.0 10.3 15.5



#51
3-Nitroaniline
Concen: 21.692 ng/ul
RT: 14.546 min Scan# 1914
Delta R.T. 0.000 min
Lab File: BN024600.D
Acq: 29 Mar 2023 08:34

Tgt Ion:138 Resp: 331716
Ion Ratio Lower Upper
138 100
108 10.9 8.8 13.2
92 116.2 96.1 144.1









#52
Acenaphthene
Concen: 21.258 ng/ul
RT: 14.704 min Scan# 1941
Delta R.T. 0.000 min
Lab File: BN024600.D
Acq: 29 Mar 2023 08:34




Instrument :
BNA_N
ClientSampleId :
SSTD020265


Tgt Ion:153 Resp: 1381243
Ion Ratio Lower Upper
153 100
152 47.8 36.6 55.0
154 89.5 67.7 101.5



#53
2,4-Dinitrophenol
Concen: 18.072 ng/ul
RT: 14.746 min Scan# 1948
Delta R.T. 0.000 min
Lab File: BN024600.D
Acq: 29 Mar 2023 08:34

Tgt Ion:184 Resp: 176609
Ion Ratio Lower Upper
184 100
63 85.1 59.1 88.7
154 65.5 49.2 73.8









#55
4-Nitrophenol
Concen: 21.686 ng/ul
RT: 14.846 min Scan# 1965
Delta R.T. 0.000 min
Lab File: BN024600.D
Acq: 29 Mar 2023 08:34




Instrument :
BNA_N
ClientSampleId :
SSTD020265


Tgt Ion:109 Resp: 259724
Ion Ratio Lower Upper
109 100
139 129.5 108.4 162.6
65 144.7 121.1 181.7



#56
Dibenzofuran
Concen: 20.915 ng/ul
RT: 15.040 min Scan# 1998
Delta R.T. 0.000 min
Lab File: BN024600.D
Acq: 29 Mar 2023 08:34

Tgt Ion:168 Resp: 1891015
Ion Ratio Lower Upper
168 100
139 37.3 30.0 45.0





#57

2,4-Dinitrotoluene

Concen: 21.696 ng/ul

RT: 14.999 min Scan# 1991

Delta R.T. 0.000 min

Lab File: BN024600.D

Acq: 29 Mar 2023 08:34

Instrument :

BNA_N

ClientSampleId :

SSTD020265





Tgt Ion:165 Resp: 457269

Ion Ratio Lower Upper

165 100

63 52.1 42.2 63.4

182 2.6 2.3 3.5



#58

2,3,4,6-Tetrachlorophenol

Concen: 21.295 ng/ul

RT: 15.263 min Scan# 2036

Delta R.T. 0.000 min

Lab File: BN024600.D

Acq: 29 Mar 2023 08:34

Tgt Ion:232 Resp: 374644




Ion Ratio Lower Upper


232 100

131 46.9 36.5 54.7

130 2.5 1.8 2.8

166 31.6 24.4 36.6





#59

Diethylphthalate

Concen: 20.968 ng/ul

RT: 15.457 min Scan# 2069

Delta R.T. 0.000 min

Lab File: BN024600.D

Acq: 29 Mar 2023 08:34

Instrument :

BNA_N

ClientSampleId :

SSTD020265





Tgt Ion:149 Resp: 1558409

Ion Ratio Lower Upper

149 100

177 22.6 17.5 26.3

150 12.2 9.8 14.6



#61

Fluorene

Concen: 21.747 ng/ul

RT: 15.687 min Scan# 2108

Delta R.T. 0.000 min

Lab File: BN024600.D

Acq: 29 Mar 2023 08:34




Tgt Ion:166 Resp: 1576544


Ion Ratio Lower Upper

166 100

165 95.5 78.6 118.0

167 13.7 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 21.604 ng/ul
RT: 15.675 min Scan# 2106
Delta R.T. 0.000 min
Lab File: BN024600.D
Acq: 29 Mar 2023 08:34




Instrument :
BNA_N
ClientSampleId :
SSTD020265


Tgt Ion:204 Resp: 743124
Ion Ratio Lower Upper
204 100
206 34.0 25.9 38.9
141 63.4 49.9 74.9



#63
4-Nitroaniline
Concen: 22.461 ng/ul
RT: 15.704 min Scan# 2111
Delta R.T. 0.000 min
Lab File: BN024600.D
Acq: 29 Mar 2023 08:34

Tgt Ion:138 Resp: 338374
Ion Ratio Lower Upper
138 100
92 53.6 41.5 62.3
108 80.2 63.0 94.4









#66
 4,6-Dinitro-2-methylphenol
 Concen: 20.468 ng/ul
 RT: 15.763 min Scan# 2121
 Delta R.T. 0.000 min
 Lab File: BN024600.D
 Acq: 29 Mar 2023 08:34




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020265


Tgt Ion:198 Resp: 276844
 Ion Ratio Lower Upper
 198 100
 182 4.2 3.9 5.9
 77 25.4 22.2 33.2



#67
 N-Nitrosodiphenylamine
 Concen: 21.395 ng/ul
 RT: 15.893 min Scan# 2143
 Delta R.T. 0.000 min
 Lab File: BN024600.D
 Acq: 29 Mar 2023 08:34

Tgt Ion:169 Resp: 1328561
 Ion Ratio Lower Upper
 169 100
 168 63.7 54.3 81.5
 167 36.1 29.4 44.0









#68
4-Bromophenyl-phenylether
Concen: 21.388 ng/ul
RT: 16.575 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN024600.D
Acq: 29 Mar 2023 08:34




Instrument :
BNA_N
ClientSampleId :
SSTD020265


Tgt Ion:248 Resp: 443292
Ion Ratio Lower Upper
248 100
250 100.8 74.5 111.7
141 78.2 60.7 91.1



#69
Hexachlorobenzene
Concen: 21.023 ng/ul
RT: 16.693 min Scan# 2279
Delta R.T. 0.000 min
Lab File: BN024600.D
Acq: 29 Mar 2023 08:34

Tgt Ion:284 Resp: 512780
Ion Ratio Lower Upper
284 100
142 35.3 28.5 42.7
249 30.2 25.5 38.3









#70
 Atrazine
 Concen: 21.803 ng/ul
 RT: 16.840 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BN024600.D
 Acq: 29 Mar 2023 08:34




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020265


Tgt Ion:200 Resp: 462566
 Ion Ratio Lower Upper
 200 100
 173 27.3 21.8 32.6
 215 45.9 38.2 57.2



#71
 Pentachlorophenol
 Concen: 20.233 ng/ul
 RT: 17.034 min Scan# 2337
 Delta R.T. 0.000 min
 Lab File: BN024600.D
 Acq: 29 Mar 2023 08:34

Tgt Ion:266 Resp: 301690
 Ion Ratio Lower Upper
 266 100
 264 70.6 52.5 78.7
 268 64.5 54.4 81.6









#72
Phenanthrene
Concen: 21.781 ng/ul
RT: 17.428 min Scan# 2404
Delta R.T. 0.000 min
Lab File: BN024600.D
Acq: 29 Mar 2023 08:34




Instrument :
BNA_N
ClientSampleId :
SSTD020265


Tgt Ion:178 Resp: 2534504
Ion Ratio Lower Upper
178 100
179 15.6 12.7 19.1
176 18.5 15.5 23.3



#74
Anthracene
Concen: 21.965 ng/ul
RT: 17.522 min Scan# 2420
Delta R.T. 0.000 min
Lab File: BN024600.D
Acq: 29 Mar 2023 08:34

Tgt Ion:178 Resp: 2567731
Ion Ratio Lower Upper
178 100
179 15.3 11.9 17.9
176 18.7 14.1 21.1





#75

1,2,3,4-Tetrachlorobenzene

Concen: 21.406 ng/uL

RT: 13.440 min Scan# 1726

Delta R.T. 0.000 min

Lab File: BN024600.D

Acq: 29 Mar 2023 08:34

Instrument :

BNA_N

ClientSampleId :

SSTD020265





Tgt Ion:216 Resp: 635488

Ion Ratio Lower Upper

216 100

214 74.3 62.7 94.1

218 47.6 38.0 57.0



#76

Pentachlorobenzene

Concen: 21.179 ng/uL

RT: 14.957 min Scan# 1984

Delta R.T. 0.000 min

Lab File: BN024600.D

Acq: 29 Mar 2023 08:34




Tgt Ion:250 Resp: 610749


Ion Ratio Lower Upper

250 100

248 63.8 50.5 75.7

252 64.7 50.9 76.3





#77

Carbazole

Concen: 22.256 ng/ul

RT: 17.792 min Scan# 2466

Delta R.T. 0.000 min

Lab File: BN024600.D

Acq: 29 Mar 2023 08:34

Instrument :





BNA_N

ClientSampleId :

SSTD020265

Tgt Ion:167 Resp: 2339277

Ion	Ratio	Lower	Upper
167	100		
166	20.4	17.4	26.2
139	13.0	10.5	15.7



#78

Di-n-butylphthalate

Concen: 21.771 ng/ul

RT: 18.351 min Scan# 2561




Delta R.T. 0.000 min


Lab File: BN024600.D

Acq: 29 Mar 2023 08:34

Tgt Ion:149 Resp: 2593047

Ion	Ratio	Lower	Upper
149	100		
150	8.6	7.4	11.2
104	5.1	4.1	6.1









#80
Fluoranthene
Concen: 21.743 ng/ul
RT: 19.445 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN024600.D
Acq: 29 Mar 2023 08:34




Instrument :
BNA_N
ClientSampleId :
SSTD020265


Tgt Ion:202 Resp: 2906408
Ion Ratio Lower Upper
202 100
101 12.3 10.3 15.5
100 9.6 8.2 12.2



#82
Pyrene
Concen: 21.330 ng/ul
RT: 19.804 min Scan# 2808
Delta R.T. 0.000 min
Lab File: BN024600.D
Acq: 29 Mar 2023 08:34

Tgt Ion:202 Resp: 3005644
Ion Ratio Lower Upper
202 100
101 15.4 12.5 18.7
100 12.8 10.1 15.1





#83

Butylbenzylphthalate

Concen: 21.174 ng/ul

RT: 20.704 min Scan# 2961

Delta R.T. 0.000 min

Lab File: BN024600.D

Acq: 29 Mar 2023 08:34

Instrument :

BNA_N

ClientSampleId :

SSTD020265





Tgt Ion:149 Resp: 1136068

Ion Ratio Lower Upper

149 100

91 74.8 59.7 89.5

206 19.6 15.4 23.2



#84

3,3'-Dichlorobenzidine

Concen: 21.487 ng/ul

RT: 21.498 min Scan# 3096

Delta R.T. 0.000 min

Lab File: BN024600.D

Acq: 29 Mar 2023 08:34




Tgt Ion:252 Resp: 979162


Ion Ratio Lower Upper

252 100

254 62.1 50.0 75.0

126 11.5 9.8 14.8





#85

Benzo(a)anthracene

Concen: 21.446 ng/ul

RT: 21.563 min Scan# 3107

Delta R.T. 0.000 min

Lab File: BN024600.D

Acq: 29 Mar 2023 08:34

Instrument :

BNA_N

ClientSampleId :

SSTD020265





Tgt Ion:228 Resp: 2953375

Ion Ratio Lower Upper

228 100

229 20.6 15.8 23.8

226 27.2 22.7 34.1



#86

Bis(2-ethylhexyl)phthalate

Concen: 22.271 ng/ul

RT: 21.480 min Scan# 3093

Delta R.T. 0.000 min

Lab File: BN024600.D

Acq: 29 Mar 2023 08:34




Tgt Ion:149 Resp: 1666101


Ion Ratio Lower Upper

149 100

167 27.1 21.6 32.4

279 3.4 2.7 4.1





#87

Chrysene

Concen: 22.116 ng/ul

RT: 21.622 min Scan# 3117

Delta R.T. 0.000 min

Lab File: BN024600.D

Acq: 29 Mar 2023 08:34

Instrument :

BNA_N

ClientSampleId :

SSTD020265





Tgt Ion:228 Resp: 2881884

Ion Ratio Lower Upper

228 100

226 29.9 24.8 37.2

229 19.6 16.4 24.6



#89

Di-n-octyl phthalate

Concen: 21.263 ng/ul




RT: 22.433 min Scan# 3255


Delta R.T. 0.000 min

Lab File: BN024600.D

Acq: 29 Mar 2023 08:34

Tgt Ion:149 Resp: 2810671





#90

Benzo(b)fluoranthene

Concen: 21.843 ng/ul

RT: 23.298 min Scan# 3402

Delta R.T. 0.000 min

Lab File: BN024600.D

Acq: 29 Mar 2023 08:34

Instrument :

BNA_N

ClientSampleId :

SSTD020265





Tgt Ion:252 Resp: 3120931

Ion Ratio Lower Upper

252 100

253 19.7 15.8 23.8

125 11.3 8.6 13.0



#91

Benzo(k)fluoranthene

Concen: 21.332 ng/ul

RT: 23.351 min Scan# 3411

Delta R.T. 0.000 min

Lab File: BN024600.D

Acq: 29 Mar 2023 08:34




Tgt Ion:252 Resp: 2909753


Ion Ratio Lower Upper

252 100

253 22.3 17.5 26.3

125 11.6 9.2 13.8





#93

Benzo(a)pyrene

Concen: 21.647 ng/ul

RT: 23.951 min Scan# 3513

Delta R.T. 0.000 min

Lab File: BN024600.D

Acq: 29 Mar 2023 08:34

Instrument :

BNA_N

ClientSampleId :

SSTD020265





Tgt Ion:252 Resp: 2708461

Ion Ratio Lower Upper

252 100

253 22.6 18.2 27.4

125 12.7 10.2 15.4



#94

Indeno(1,2,3-cd)pyrene

Concen: 21.648 ng/ul

RT: 26.645 min Scan# 3971

Delta R.T. 0.000 min

Lab File: BN024600.D

Acq: 29 Mar 2023 08:34




Tgt Ion:276 Resp: 3451605


Ion Ratio Lower Upper

276 100

138 27.9 22.2 33.2

277 25.4 20.3 30.5





#95

Dibenzo(a,h)anthracene

Concen: 21.302 ng/ul

RT: 26.663 min Scan# 3974

Delta R.T. 0.000 min

Lab File: BN024600.D

Acq: 29 Mar 2023 08:34

Instrument :

BNA_N

ClientSampleId :

SSTD020265





Tgt Ion:278 Resp: 2816749

Ion Ratio Lower Upper

278 100

139 19.5 14.2 21.4

279 25.0 19.6 29.4



#96

Benzo(g,h,i)perylene

Concen: 21.342 ng/ul

RT: 27.439 min Scan# 4106

Delta R.T. 0.000 min

Lab File: BN024600.D

Acq: 29 Mar 2023 08:34

Tgt Ion:276 Resp: 2827345

Ion Ratio Lower Upper

276 100

138 24.1 19.5 29.3

277 23.8 18.4 27.6

