

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032923\
 Data File : BN024613.D
 Acq On : 29 Mar 2023 16:42
 Operator : CG/JU
 Sample : 01982-16
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 EYGC5

Quant Time: Mar 30 02:30:49 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032923.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 30 02:26:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.999	152	357383	20.000	ng/ul	0.00
20) Naphthalene-d8	10.816	136	1562517	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.639	164	973207	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.386	188	2143923	20.000	ng/ul	0.00
79) Chrysene-d12	21.574	240	1919326	20.000	ng/ul	0.00
88) Perylene-d12	24.063	264	2107764	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.369	96	73825	8.009	ng/uL	0.00
4) Pyridine-d5	3.793	84	1082101	40.655	ng/ul	0.00
7) Phenol-d5	7.158	99	1406425	41.584	ng/ul	0.01
9) Bis-(2-Chloroethyl)eth...	7.328	67	935659	42.840	ng/ul	0.01
11) 2-Chlorophenol-d4	7.528	132	1067576	42.576	ng/ul	0.00
15) 4-Methylphenol-d8	8.710	113	1107246	40.935	ng/ul	0.00
21) Nitrobenzene-d5	9.169	128	550842	44.545	ng/ul	0.00
24) 2-Nitrophenol-d4	9.893	143	611137	45.658	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.434	165	1040287	42.791	ng/ul	0.00
31) 4-Chloroaniline-d4	10.957	131	1551466	41.825	ng/ul	0.00
46) Dimethylphthalate-d6	14.051	166	3337098	44.590	ng/ul	0.00
49) Acenaphthylene-d8	14.334	160	3888303	45.628	ng/ul	0.00
54) 4-Nitrophenol-d4	14.828	143	607796	38.803	ng/ul	0.00
60) Fluorene-d10	15.634	176	2971365	46.273	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.745	200	460798	32.632	ng/ul	0.00
73) Anthracene-d10	17.486	188	4545338	45.076	ng/ul	0.00
81) Pyrene-d10	19.780	212	5161413	46.509	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.915	264	4928386	46.146	ng/ul	0.02

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032923\
Data File : BN024613.D
Acq On : 29 Mar 2023 16:42
Operator : CG/JU
Sample : 01982-16
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
EYGC5

Quant Time: Mar 30 02:30:49 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032923.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Mar 30 02:26:44 2023
Response via : Initial Calibration

