

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032923\
 Data File : BN024614.D
 Acq On : 29 Mar 2023 17:55
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020266

Quant Time: Mar 30 02:30:57 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032923.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 30 02:26:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.005	152	363605	20.000	ng/u1	0.01
20) Naphthalene-d8	10.816	136	1635913	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.639	164	1014076	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.386	188	2167101	20.000	ng/u1	0.00
79) Chrysene-d12	21.574	240	1993833	20.000	ng/u1	0.00
88) Perylene-d12	24.074	264	2105897	20.000	ng/u1	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.369	96	79005	8.424	ng/uL	0.00
4) Pyridine-d5	3.799	84	579803	21.411	ng/u1	0.00
7) Phenol-d5	7.157	99	717928	20.864	ng/u1	0.01
9) Bis-(2-Chloroethyl)eth...	7.328	67	472102	21.246	ng/u1	0.01
11) 2-Chlorophenol-d4	7.528	132	551247	21.608	ng/u1	0.00
15) 4-Methylphenol-d8	8.710	113	573897	20.854	ng/u1	0.00
21) Nitrobenzene-d5	9.169	128	275027	21.243	ng/u1	0.00
24) 2-Nitrophenol-d4	9.899	143	303794	21.678	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.434	165	550324	21.621	ng/u1	0.00
31) 4-Chloroaniline-d4	10.951	131	826960	21.293	ng/u1	0.00
46) Dimethylphthalate-d6	14.045	166	1669314	21.406	ng/u1	0.00
49) Acenaphthylene-d8	14.339	160	1957977	22.050	ng/u1	0.00
54) 4-Nitrophenol-d4	14.834	143	340041	20.834	ng/u1	0.00
60) Fluorene-d10	15.634	176	1415744	21.159	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.751	200	290388	20.344	ng/u1	0.00
73) Anthracene-d10	17.486	188	2202345	21.607	ng/u1	0.00
81) Pyrene-d10	19.774	212	2460785	21.345	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.904	264	2326057	21.799	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.405	88	87287	8.638	ng/uL	98
5) Pyridine	3.816	79	594522	21.092	ng/u1	98
6) Benzaldehyde	7.134	77	415614	26.462	ng/u1	99
8) Phenol	7.181	94	747323	20.922	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.422	93	607837	21.580	ng/u1	99
12) 2-Chlorophenol	7.557	128	571624	21.477	ng/u1	98
13) 2-Methylphenol	8.446	108	561412	21.047	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.534	45	1084691	21.679	ng/u1	99
16) Acetophenone	8.828	105	931183	21.676	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.816	70	490042	21.676	ng/u1	99
18) 4-Methylphenol	8.775	108	618425	20.883	ng/u1	100
19) Hexachloroethane	9.087	117	225551	21.457	ng/u1	98
22) Nitrobenzene	9.210	77	721673	21.460	ng/u1	99
23) Isophorone	9.740	82	1347053	21.550	ng/u1	99
25) 2-Nitrophenol	9.928	139	330822	21.714	ng/u1	100
26) 2,4-Dimethylphenol	9.987	107	680181	21.493	ng/u1	96
27) Bis(2-Chloroethoxy)met...	10.222	93	840015	20.941	ng/u1	98
29) 2,4-Dichlorophenol	10.457	162	551490	21.362	ng/u1	96
30) Naphthalene	10.869	128	1923831	21.164	ng/u1	99
32) 4-Chloroaniline	10.981	127	831267	21.137	ng/u1	100
33) Hexachlorobutadiene	11.151	225	321295	21.610	ng/u1	97
34) Caprolactam	11.757	113	110769	12.406	ng/u1	98
35) 4-Chloro-3-methylphenol	12.093	107	640668	21.706	ng/u1	99
36) 2-Methylnaphthalene	12.475	142	1303515	21.419	ng/u1	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032923\
 Data File : BN024614.D
 Acq On : 29 Mar 2023 17:55
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020266

Quant Time: Mar 30 02:30:57 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032923.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 30 02:26:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.693	142	1273757	21.124	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.840	216	631289	21.001	ng/ul	92
40) Hexachlorocyclopentadiene	12.816	237	382052	20.196	ng/ul	97
41) 2,4,6-Trichlorophenol	13.075	196	439049	21.923	ng/ul	98
42) 2,4,5-Trichlorophenol	13.145	196	480600	21.834	ng/ul	97
43) 1,1'-Biphenyl	13.475	154	1705400	21.167	ng/ul	100
44) 2-Chloronaphthalene	13.522	162	1347487	21.424	ng/ul	99
45) 2-Nitroaniline	13.722	65	441657	22.600	ng/ul	96
47) Dimethylphthalate	14.098	163	1656797	20.922	ng/ul	99
48) 2,6-Dinitrotoluene	14.216	165	334861	22.324	ng/ul	99
50) Acenaphthylene	14.363	152	2137030	21.982	ng/ul	100
51) 3-Nitroaniline	14.545	138	350239	21.967	ng/ul	97
52) Acenaphthene	14.704	153	1465757	21.636	ng/ul	98
53) 2,4-Dinitrophenol	14.751	184	185723	18.228	ng/ul	99
55) 4-Nitrophenol	14.845	109	266392	21.333	ng/ul	98
56) Dibenzofuran	15.039	168	2022618	21.456	ng/ul	99
57) 2,4-Dinitrotoluene	15.004	165	486549	22.141	ng/ul	93
58) 2,3,4,6-Tetrachlorophenol	15.263	232	401536	21.891	ng/ul	95
59) Diethylphthalate	15.457	149	1652186	21.321	ng/ul	100
61) Fluorene	15.686	166	1636151	21.646	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.681	204	785960	21.915	ng/ul	97
63) 4-Nitroaniline	15.710	138	364530	23.208	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.763	198	282877	20.040	ng/ul	99
67) N-Nitrosodiphenylamine	15.892	169	1391917	21.479	ng/ul	100
68) 4-Bromophenyl-phenylether	16.575	248	473086	21.872	ng/ul	98
69) Hexachlorobenzene	16.692	284	533310	20.951	ng/ul	97
70) Atrazine	16.845	200	472652	21.348	ng/ul	100
71) Pentachlorophenol	17.033	266	328564	21.115	ng/ul	95
72) Phenanthrene	17.428	178	2575671	21.210	ng/ul	98
74) Anthracene	17.522	178	2643468	21.668	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.445	216	653288	21.086	ng/ul	95
76) Pentachlorobenzene	14.957	250	636564	21.151	ng/ul	97
77) Carbazole	17.792	167	2440514	22.249	ng/ul	98
78) Di-n-butylphthalate	18.351	149	2712461	21.822	ng/ul	100
80) Fluoranthene	19.439	202	2984780	21.832	ng/ul	99
82) Pyrene	19.804	202	3090375	21.443	ng/ul	99
83) Butylbenzylphthalate	20.698	149	1175899	21.428	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.492	252	1000909	21.475	ng/ul	99
85) Benzo(a)anthracene	21.563	228	2939843	20.873	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.474	149	1659938	21.695	ng/ul	98
87) Chrysene	21.616	228	2860228	21.461	ng/ul	95
89) Di-n-octyl phthalate	22.433	149	2817877	22.089	ng/ul	100
90) Benzo(b)fluoranthene	23.310	252	2982158	21.626	ng/ul	99
91) Benzo(k)fluoranthene	23.357	252	2879405	21.873	ng/ul	99
93) Benzo(a)pyrene	23.957	252	2572954	21.308	ng/ul	96
94) Indeno(1,2,3-cd)pyrene	26.674	276	3195718	20.768	ng/ul	96
95) Dibenzo(a,h)anthracene	26.686	278	2613173	20.477	ng/ul	98
96) Benzo(g,h,i)perylene	27.474	276	2479551	19.394	ng/ul	99





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 8.638 ng/uL
RT: 3.405 min Scan# 20
Delta R.T. 0.006 min
Lab File: BN024614.D
Acq: 29 Mar 2023 17:55




Instrument :
BNA_N
ClientSampleId :
SSTD020266


Tgt Ion: 88 Resp: 87287
Ion Ratio Lower Upper
88 100
43 36.8 29.0 43.6
58 82.4 67.5 101.3



#5
Pyridine
Concen: 21.092 ng/uL
RT: 3.816 min Scan# 90
Delta R.T. 0.006 min
Lab File: BN024614.D
Acq: 29 Mar 2023 17:55

Tgt Ion: 79 Resp: 594522
Ion Ratio Lower Upper
79 100
52 92.7 76.0 114.0
51 42.7 35.0 52.4









#6
Benzaldehyde
Concen: 26.462 ng/ul
RT: 7.134 min Scan# 61
Delta R.T. 0.006 min
Lab File: BN024614.D
Acq: 29 Mar 2023 17:55




Instrument :
BNA_N
ClientSampleId :
SSTD020266


Tgt Ion: 77 Resp: 415614
Ion Ratio Lower Upper
77 100
105 93.6 74.0 111.0
106 90.0 71.7 107.5



#8
Phenol
Concen: 20.922 ng/ul
RT: 7.181 min Scan# 662
Delta R.T. 0.006 min
Lab File: BN024614.D
Acq: 29 Mar 2023 17:55

Tgt Ion: 94 Resp: 747323
Ion Ratio Lower Upper
94 100
65 30.3 23.7 35.5
66 38.4 29.9 44.9









#10
Bis(2-Chloroethyl)ether
Concen: 21.580 ng/ul
RT: 7.422 min Scan# 701
Delta R.T. 0.012 min
Lab File: BN024614.D
Acq: 29 Mar 2023 17:55




Instrument :
BNA_N
ClientSampleId :
SSTD020266


Tgt Ion: 93 Resp: 607837
Ion Ratio Lower Upper
93 100
63 80.4 64.9 97.3
95 31.7 25.2 37.8



#12
2-Chlorophenol
Concen: 21.477 ng/ul
RT: 7.557 min Scan# 726
Delta R.T. 0.006 min
Lab File: BN024614.D
Acq: 29 Mar 2023 17:55

Tgt Ion:128 Resp: 571624
Ion Ratio Lower Upper
128 100
64 54.0 41.6 62.4
130 32.0 25.6 38.4





#13

2-Methylphenol

Concen: 21.047 ng/ul

RT: 8.446 min Scan# 81

Delta R.T. 0.006 min

Lab File: BN024614.D

Acq: 29 Mar 2023 17:55

Instrument :

BNA_N

ClientSampleId :





SSTD020266

Tgt Ion:108 Resp: 561412

Ion Ratio Lower Upper

108 100

107 90.0 70.9 106.3



#14

2,2'-oxybis(1-Chloropropane)

Concen: 21.679 ng/ul

RT: 8.534 min Scan# 892

Delta R.T. 0.006 min

Lab File: BN024614.D

Acq: 29 Mar 2023 17:55




Tgt Ion: 45 Resp: 1084691


Ion Ratio Lower Upper

45 100

77 10.9 9.0 13.4

79 8.2 6.9 10.3









#16
Acetophenone
Concen: 21.676 ng/ul
RT: 8.828 min Scan# 941
Delta R.T. 0.006 min
Lab File: BN024614.D
Acq: 29 Mar 2023 17:55




Instrument :
BNA_N
ClientSampleId :
SSTD020266


Tgt Ion:105 Resp: 931183
Ion Ratio Lower Upper
105 100
77 79.6 64.3 96.5
51 34.0 27.3 40.9



#17
N-Nitroso-di-n-propylamine
Concen: 21.676 ng/ul
RT: 8.816 min Scan# 940
Delta R.T. 0.006 min
Lab File: BN024614.D
Acq: 29 Mar 2023 17:55

Tgt Ion: 70 Resp: 490042
Ion Ratio Lower Upper
70 100
42 73.2 57.5 86.3
101 9.5 7.8 11.8
130 19.4 15.7 23.5









#18
4-Methylphenol
Concen: 20.883 ng/ul
RT: 8.775 min Scan# 911
Delta R.T. 0.006 min
Lab File: BN024614.D
Acq: 29 Mar 2023 17:55




Instrument :
BNA_N
ClientSampleId :
SSTD020266


Tgt Ion:108 Resp: 618425
Ion Ratio Lower Upper
108 100
107 109.3 87.7 131.5



#19
Hexachloroethane
Concen: 21.457 ng/ul
RT: 9.087 min Scan# 986
Delta R.T. 0.006 min
Lab File: BN024614.D
Acq: 29 Mar 2023 17:55

Tgt Ion:117 Resp: 225551
Ion Ratio Lower Upper
117 100
201 89.9 71.9 107.9
199 60.3 45.2 67.8









#22
Nitrobenzene
Concen: 21.460 ng/ul
RT: 9.210 min Scan# 1007
Delta R.T. 0.006 min
Lab File: BN024614.D
Acq: 29 Mar 2023 17:55




Instrument :
BNA_N
ClientSampleId :
SSTD020266


Tgt Ion: 77 Resp: 721673
Ion Ratio Lower Upper
77 100
123 42.4 34.3 51.5
65 13.7 10.8 16.2



#23
Isophorone
Concen: 21.550 ng/ul
RT: 9.740 min Scan# 1097
Delta R.T. 0.006 min
Lab File: BN024614.D
Acq: 29 Mar 2023 17:55

Tgt Ion: 82 Resp: 1347053
Ion Ratio Lower Upper
82 100
95 6.6 5.4 8.0
138 16.3 12.6 18.8





#25

2-Nitrophenol

Concen: 21.714 ng/ul

RT: 9.928 min Scan# 1128

Delta R.T. 0.006 min

Lab File: BN024614.D

Acq: 29 Mar 2023 17:55

Instrument :

BNA_N

ClientSampleId :

SSTD020266





Tgt Ion:139 Resp: 330822

Ion Ratio Lower Upper

139 100

65 58.1 46.3 69.5

109 30.6 24.6 37.0



#26

2,4-Dimethylphenol

Concen: 21.493 ng/ul

RT: 9.987 min Scan# 1139

Delta R.T. 0.012 min

Lab File: BN024614.D

Acq: 29 Mar 2023 17:55




Tgt Ion:107 Resp: 680181


Ion Ratio Lower Upper

107 100

121 50.1 40.7 61.1

122 83.3 71.2 106.8









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.941 ng/ul
 RT: 10.222 min Scan# 1179
 Delta R.T. 0.006 min
 Lab File: BN024614.D
 Acq: 29 Mar 2023 17:55




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020266


Tgt Ion: 93 Resp: 840015
 Ion Ratio Lower Upper
 93 100
 95 31.8 26.4 39.6
 123 10.9 8.9 13.3



#29
 2,4-Dichlorophenol
 Concen: 21.362 ng/ul
 RT: 10.457 min Scan# 1219
 Delta R.T. -0.000 min
 Lab File: BN024614.D
 Acq: 29 Mar 2023 17:55

Tgt Ion:162 Resp: 551490
 Ion Ratio Lower Upper
 162 100
 164 65.9 50.5 75.7
 98 38.4 28.6 42.8





#30

Naphthalene

Concen: 21.164 ng/ul

RT: 10.869 min Scan# 1288

Delta R.T. 0.006 min

Lab File: BN024614.D

Acq: 29 Mar 2023 17:55

Instrument :

BNA_N

ClientSampleId :

SSTD020266





Tgt Ion:128 Resp: 1923831

Ion Ratio Lower Upper

128 100

129 11.0 8.5 12.7

127 13.3 10.4 15.6



#32

4-Chloroaniline

Concen: 21.137 ng/ul

RT: 10.981 min Scan# 1308

Delta R.T. 0.006 min

Lab File: BN024614.D




Acq: 29 Mar 2023 17:55


Tgt Ion:127 Resp: 831267

Ion Ratio Lower Upper

127 100

129 32.0 25.5 38.3









#33
Hexachlorobutadiene
Concen: 21.610 ng/ul
RT: 11.151 min Scan# 111
Delta R.T. 0.006 min
Lab File: BN024614.D
Acq: 29 Mar 2023 17:55




Instrument :
BNA_N
ClientSampleId :
SSTD020266


Tgt Ion:225 Resp: 321295
Ion Ratio Lower Upper
225 100
223 62.9 50.2 75.4
227 66.0 49.6 74.4



#34
Caprolactam
Concen: 12.406 ng/ul
RT: 11.757 min Scan# 1440
Delta R.T. 0.006 min
Lab File: BN024614.D
Acq: 29 Mar 2023 17:55

Tgt Ion:113 Resp: 110769
Ion Ratio Lower Upper
113 100
55 204.3 162.2 243.2
56 147.3 121.6 182.4









#35
4-Chloro-3-methylphenol
Concen: 21.706 ng/ul
RT: 12.093 min Scan# 1497
Delta R.T. -0.000 min
Lab File: BN024614.D
Acq: 29 Mar 2023 17:55




Instrument :
BNA_N
ClientSampleId :
SSTD020266


Tgt Ion:107 Resp: 640668
Ion Ratio Lower Upper
107 100
144 25.0 21.1 31.7
142 75.7 61.0 91.4



#36
2-Methylnaphthalene
Concen: 21.419 ng/ul
RT: 12.475 min Scan# 1562
Delta R.T. 0.006 min
Lab File: BN024614.D
Acq: 29 Mar 2023 17:55

Tgt Ion:142 Resp: 1303515
Ion Ratio Lower Upper
142 100
141 86.8 70.2 105.4









#37
1-Methylnaphthalene
Concen: 21.124 ng/ul
RT: 12.693 min Scan# 1599
Delta R.T. 0.006 min
Lab File: BN024614.D
Acq: 29 Mar 2023 17:55




Instrument :
BNA_N
ClientSampleId :
SSTD020266


Tgt Ion:142 Resp: 1273757
Ion Ratio Lower Upper
142 100
141 93.7 74.2 111.4
116 3.7 3.0 4.6



#39
1,2,4,5-Tetrachlorobenzene
Concen: 21.001 ng/ul
RT: 12.840 min Scan# 1624
Delta R.T. 0.006 min
Lab File: BN024614.D
Acq: 29 Mar 2023 17:55

Tgt Ion:216 Resp: 631289
Ion Ratio Lower Upper
216 100
214 82.7 61.0 91.4
179 23.4 15.7 23.5
108 20.5 14.0 21.0









#40
Hexachlorocyclopentadiene
Concen: 20.196 ng/ul
RT: 12.816 min Scan# 1620
Delta R.T. -0.000 min
Lab File: BN024614.D
Acq: 29 Mar 2023 17:55




Instrument :
BNA_N
ClientSampleId :
SSTD020266


Tgt Ion:237 Resp: 382052
Ion Ratio Lower Upper
237 100
235 64.7 50.0 75.0
272 13.5 9.5 14.3



#41
2,4,6-Trichlorophenol
Concen: 21.923 ng/ul
RT: 13.075 min Scan# 1664
Delta R.T. -0.000 min
Lab File: BN024614.D
Acq: 29 Mar 2023 17:55

Tgt Ion:196 Resp: 439049
Ion Ratio Lower Upper
196 100
198 92.4 75.5 113.3
200 29.8 24.6 37.0





#42

2,4,5-Trichlorophenol

Concen: 21.834 ng/ul

RT: 13.145 min Scan# 1

Delta R.T. -0.000 min

Lab File: BN024614.D

Acq: 29 Mar 2023 17:55

Instrument :

BNA_N

ClientSampleId :

SSTD020266





Tgt Ion:196 Resp: 480600

Ion Ratio Lower Upper

196 100

198 91.3 75.4 113.0

200 29.4 24.6 37.0



#43

1,1'-Biphenyl

Concen: 21.167 ng/ul

RT: 13.475 min Scan# 1732

Delta R.T. -0.000 min

Lab File: BN024614.D

Acq: 29 Mar 2023 17:55




Tgt Ion:154 Resp: 1705400


Ion Ratio Lower Upper

154 100

153 40.3 32.2 48.4

76 14.9 11.7 17.5





#44

2-Chloronaphthalene

Concen: 21.424 ng/ul

RT: 13.522 min Scan# 1740

Delta R.T. 0.006 min

Lab File: BN024614.D

Acq: 29 Mar 2023 17:55

Instrument :

BNA_N

ClientSampleId :

SSTD020266





Tgt Ion:162 Resp: 1347487

Ion Ratio Lower Upper

162 100

164 33.1 26.1 39.1

127 39.9 31.1 46.7



#45

2-Nitroaniline

Concen: 22.600 ng/ul

RT: 13.722 min Scan# 1774

Delta R.T. -0.000 min

Lab File: BN024614.D

Acq: 29 Mar 2023 17:55




Tgt Ion: 65 Resp: 441657


Ion Ratio Lower Upper

65 100

92 64.2 52.6 78.8

138 96.2 81.4 122.2









#47
Dimethylphthalate
Concen: 20.922 ng/ul
RT: 14.098 min Scan# 1838
Delta R.T. 0.006 min
Lab File: BN024614.D
Acq: 29 Mar 2023 17:55




Instrument :
BNA_N
ClientSampleId :
SSTD020266


Tgt Ion:163 Resp: 1656797
Ion Ratio Lower Upper
163 100
194 4.2 3.2 4.8
164 9.9 8.2 12.2



#48
2,6-Dinitrotoluene
Concen: 22.324 ng/ul
RT: 14.216 min Scan# 1858
Delta R.T. -0.000 min
Lab File: BN024614.D
Acq: 29 Mar 2023 17:55

Tgt Ion:165 Resp: 334861
Ion Ratio Lower Upper
165 100
89 60.7 47.8 71.6
121 21.9 17.3 25.9









#50
Acenaphthylene
Concen: 21.982 ng/ul
RT: 14.363 min Scan# 1883
Delta R.T. -0.000 min
Lab File: BN024614.D
Acq: 29 Mar 2023 17:55




Instrument :
BNA_N
ClientSampleId :
SSTD020266


Tgt Ion:152 Resp: 2137030
Ion Ratio Lower Upper
152 100
151 19.5 15.5 23.3
153 13.1 10.3 15.5



#51
3-Nitroaniline
Concen: 21.967 ng/ul
RT: 14.545 min Scan# 1914
Delta R.T. -0.000 min
Lab File: BN024614.D
Acq: 29 Mar 2023 17:55

Tgt Ion:138 Resp: 350239
Ion Ratio Lower Upper
138 100
108 11.0 8.8 13.2
92 123.6 96.1 144.1





#52

Acenaphthene

Concen: 21.636 ng/ul

RT: 14.704 min Scan# 1941

Delta R.T. -0.000 min

Lab File: BN024614.D

Acq: 29 Mar 2023 17:55

Instrument :

BNA_N

ClientSampleId :

SSTD020266





Tgt Ion:153 Resp: 1465757

Ion Ratio Lower Upper

153 100

152 47.6 36.6 55.0

154 85.5 67.7 101.5



#53

2,4-Dinitrophenol

Concen: 18.228 ng/ul

RT: 14.751 min Scan# 1949

Delta R.T. 0.006 min

Lab File: BN024614.D

Acq: 29 Mar 2023 17:55




Tgt Ion:184 Resp: 185723


Ion Ratio Lower Upper

184 100

63 72.5 59.1 88.7

154 60.9 49.2 73.8









#55
4-Nitrophenol
Concen: 21.333 ng/ul
RT: 14.845 min Scan# 1965
Delta R.T. -0.000 min
Lab File: BN024614.D
Acq: 29 Mar 2023 17:55




Instrument :
BNA_N
ClientSampleId :
SSTD020266


Tgt Ion:109 Resp: 266392
Ion Ratio Lower Upper
109 100
139 133.9 108.4 162.6
65 148.4 121.1 181.7



#56
Dibenzofuran
Concen: 21.456 ng/ul
RT: 15.039 min Scan# 1998
Delta R.T. -0.000 min
Lab File: BN024614.D
Acq: 29 Mar 2023 17:55

Tgt Ion:168 Resp: 2022618
Ion Ratio Lower Upper
168 100
139 36.9 30.0 45.0





#57

2,4-Dinitrotoluene

Concen: 22.141 ng/ul

RT: 15.004 min Scan# 1992

Delta R.T. 0.006 min

Lab File: BN024614.D

Acq: 29 Mar 2023 17:55

Instrument :

BNA_N

ClientSampleId :

SSTD020266





Tgt Ion:165 Resp: 486549

Ion Ratio Lower Upper

165 100

63 47.7 42.2 63.4

182 2.8 2.3 3.5



#58

2,3,4,6-Tetrachlorophenol

Concen: 21.891 ng/ul

RT: 15.263 min Scan# 2036

Delta R.T. -0.000 min

Lab File: BN024614.D

Acq: 29 Mar 2023 17:55

Tgt Ion:232 Resp: 401536




Ion Ratio Lower Upper


232 100

131 49.8 36.5 54.7

130 2.6 1.8 2.8

166 32.6 24.4 36.6





#59

Diethylphthalate

Concen: 21.321 ng/ul

RT: 15.457 min Scan# 2069

Delta R.T. -0.000 min

Lab File: BN024614.D

Acq: 29 Mar 2023 17:55

Instrument :

BNA_N

ClientSampleId :

SSTD020266





Tgt Ion:149 Resp: 1652186

Ion Ratio Lower Upper

149 100

177 22.0 17.5 26.3

150 12.0 9.8 14.6



#61

Fluorene

Concen: 21.646 ng/ul

RT: 15.686 min Scan# 2108

Delta R.T. -0.000 min

Lab File: BN024614.D

Acq: 29 Mar 2023 17:55




Tgt Ion:166 Resp: 1636151


Ion Ratio Lower Upper

166 100

165 96.9 78.6 118.0

167 13.2 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 21.915 ng/ul
RT: 15.681 min Scan# 2107
Delta R.T. 0.006 min
Lab File: BN024614.D
Acq: 29 Mar 2023 17:55




Instrument :
BNA_N
ClientSampleId :
SSTD020266


Tgt Ion:204 Resp: 785960
Ion Ratio Lower Upper
204 100
206 33.3 25.9 38.9
141 59.0 49.9 74.9



#63
4-Nitroaniline
Concen: 23.208 ng/ul
RT: 15.710 min Scan# 2112
Delta R.T. 0.006 min
Lab File: BN024614.D
Acq: 29 Mar 2023 17:55

Tgt Ion:138 Resp: 364530
Ion Ratio Lower Upper
138 100
92 51.0 41.5 62.3
108 76.7 63.0 94.4









#66
4,6-Dinitro-2-methylphenol
Concen: 20.040 ng/ul
RT: 15.763 min Scan# 2121
Delta R.T. -0.000 min
Lab File: BN024614.D
Acq: 29 Mar 2023 17:55




Instrument :
BNA_N
ClientSampleId :
SSTD020266


Tgt Ion:198 Resp: 282877
Ion Ratio Lower Upper
198 100
182 4.6 3.9 5.9
77 28.1 22.2 33.2



#67
N-Nitrosodiphenylamine
Concen: 21.479 ng/ul
RT: 15.892 min Scan# 2143
Delta R.T. -0.000 min
Lab File: BN024614.D
Acq: 29 Mar 2023 17:55

Tgt Ion:169 Resp: 1391917
Ion Ratio Lower Upper
169 100
168 68.0 54.3 81.5
167 36.7 29.4 44.0









#68
4-Bromophenyl-phenylether
Concen: 21.872 ng/ul
RT: 16.575 min Scan# 2259
Delta R.T. -0.000 min
Lab File: BN024614.D
Acq: 29 Mar 2023 17:55




Instrument :
BNA_N
ClientSampleId :
SSTD020266


Tgt Ion:248 Resp: 473086
Ion Ratio Lower Upper
248 100
250 92.5 74.5 111.7
141 73.4 60.7 91.1



#69
Hexachlorobenzene
Concen: 20.951 ng/ul
RT: 16.692 min Scan# 2279
Delta R.T. -0.000 min
Lab File: BN024614.D
Acq: 29 Mar 2023 17:55

Tgt Ion:284 Resp: 533310
Ion Ratio Lower Upper
284 100
142 36.9 28.5 42.7
249 29.4 25.5 38.3





#70

Atrazine

Concen: 21.348 ng/ul

RT: 16.845 min Scan# 21

Delta R.T. 0.006 min

Lab File: BN024614.D

Acq: 29 Mar 2023 17:55

Instrument :

BNA_N

ClientSampleId :

SSTD020266





Tgt Ion:200 Resp: 472652

Ion Ratio Lower Upper

200 100

173 27.6 21.8 32.6

215 47.9 38.2 57.2



#71

Pentachlorophenol

Concen: 21.115 ng/ul

RT: 17.033 min Scan# 2337

Delta R.T. -0.000 min

Lab File: BN024614.D

Acq: 29 Mar 2023 17:55




Tgt Ion:266 Resp: 328564


Ion Ratio Lower Upper

266 100

264 60.4 52.5 78.7

268 65.6 54.4 81.6









#72
Phenanthrene
Concen: 21.210 ng/ul
RT: 17.428 min Scan# 2404
Delta R.T. -0.000 min
Lab File: BN024614.D
Acq: 29 Mar 2023 17:55




Instrument :
BNA_N
ClientSampleId :
SSTD020266


Tgt Ion:178 Resp: 2575671
Ion Ratio Lower Upper
178 100
179 15.0 12.7 19.1
176 18.6 15.5 23.3



#74
Anthracene
Concen: 21.668 ng/ul
RT: 17.522 min Scan# 2420
Delta R.T. -0.000 min
Lab File: BN024614.D
Acq: 29 Mar 2023 17:55

Tgt Ion:178 Resp: 2643468
Ion Ratio Lower Upper
178 100
179 15.1 11.9 17.9
176 17.6 14.1 21.1





#75

1,2,3,4-Tetrachlorobenzene

Concen: 21.086 ng/uL

RT: 13.445 min Scan# 1727

Delta R.T. 0.006 min

Lab File: BN024614.D

Acq: 29 Mar 2023 17:55

Instrument :

BNA_N

ClientSampleId :

SSTD020266





Tgt Ion:216 Resp: 653288

Ion Ratio Lower Upper

216 100

214 83.0 62.7 94.1

218 51.2 38.0 57.0



#76

Pentachlorobenzene

Concen: 21.151 ng/uL

RT: 14.957 min Scan# 1984

Delta R.T. -0.000 min

Lab File: BN024614.D

Acq: 29 Mar 2023 17:55




Tgt Ion:250 Resp: 636564


Ion Ratio Lower Upper

250 100

248 67.7 50.5 75.7

252 62.9 50.9 76.3









#77
 Carbazole
 Concen: 22.249 ng/ul
 RT: 17.792 min Scan# 2466
 Delta R.T. -0.000 min
 Lab File: BN024614.D
 Acq: 29 Mar 2023 17:55

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020266

Tgt Ion:167 Resp: 2440514




Ion	Ratio	Lower	Upper
167	100		
166	20.7	17.4	26.2
139	12.6	10.5	15.7




#78
 Di-n-butylphthalate
 Concen: 21.822 ng/ul
 RT: 18.351 min Scan# 2561
 Delta R.T. -0.000 min
 Lab File: BN024614.D
 Acq: 29 Mar 2023 17:55

Tgt Ion:149 Resp: 2712461

Ion	Ratio	Lower	Upper
149	100		
150	9.1	7.4	11.2
104	5.1	4.1	6.1





#80

Fluoranthene

Concen: 21.832 ng/ul

RT: 19.439 min Scan# 21

Delta R.T. -0.006 min

Lab File: BN024614.D

Acq: 29 Mar 2023 17:55

Instrument :

BNA_N

ClientSampleId :

SSTD020266





Tgt Ion:202 Resp: 2984780

Ion Ratio Lower Upper

202 100

101 12.9 10.3 15.5

100 9.7 8.2 12.2



#82

Pyrene

Concen: 21.443 ng/ul

RT: 19.804 min Scan# 2808

Delta R.T. -0.000 min

Lab File: BN024614.D

Acq: 29 Mar 2023 17:55




Tgt Ion:202 Resp: 3090375


Ion Ratio Lower Upper

202 100

101 14.9 12.5 18.7

100 12.3 10.1 15.1





#83

Butylbenzylphthalate

Concen: 21.428 ng/ul

RT: 20.698 min Scan# 2960

Delta R.T. -0.006 min

Lab File: BN024614.D

Acq: 29 Mar 2023 17:55

Instrument :

BNA_N

ClientSampleId :

SSTD020266





Tgt Ion:149 Resp: 1175899

Ion Ratio Lower Upper

149 100

91 73.8 59.7 89.5

206 19.5 15.4 23.2



#84

3,3'-Dichlorobenzidine

Concen: 21.475 ng/ul

RT: 21.492 min Scan# 3095

Delta R.T. -0.006 min

Lab File: BN024614.D

Acq: 29 Mar 2023 17:55




Tgt Ion:252 Resp: 1000909


Ion Ratio Lower Upper

252 100

254 63.1 50.0 75.0

126 11.9 9.8 14.8





#85

Benzo(a)anthracene

Concen: 20.873 ng/ul

RT: 21.563 min Scan# 3107

Delta R.T. -0.000 min

Lab File: BN024614.D

Acq: 29 Mar 2023 17:55

Instrument :

BNA_N

ClientSampleId :

SSTD020266





Tgt Ion:228 Resp: 2939843

Ion Ratio Lower Upper

228 100

229 19.5 15.8 23.8

226 27.0 22.7 34.1



#86

Bis(2-ethylhexyl)phthalate

Concen: 21.695 ng/ul

RT: 21.474 min Scan# 3092

Delta R.T. -0.006 min

Lab File: BN024614.D

Acq: 29 Mar 2023 17:55




Tgt Ion:149 Resp: 1659938


Ion Ratio Lower Upper

149 100

167 28.1 21.6 32.4

279 3.7 2.7 4.1





#87

Chrysene

Concen: 21.461 ng/ul

RT: 21.616 min Scan# 3116

Delta R.T. -0.006 min

Lab File: BN024614.D

Acq: 29 Mar 2023 17:55

Instrument :

BNA_N

ClientSampleId :

SSTD020266





Tgt Ion:228 Resp: 2860228

Ion Ratio Lower Upper

228 100

226 28.9 24.8 37.2

229 17.5 16.4 24.6



#89

Di-n-octyl phthalate

Concen: 22.089 ng/ul




RT: 22.433 min Scan# 3255


Delta R.T. -0.000 min

Lab File: BN024614.D

Acq: 29 Mar 2023 17:55

Tgt Ion:149 Resp: 2817877









#90
Benzo(b)fluoranthene
Concen: 21.626 ng/ul
RT: 23.310 min Scan# 3404
Delta R.T. 0.012 min
Lab File: BN024614.D
Acq: 29 Mar 2023 17:55




Instrument :
BNA_N
ClientSampleId :
SSTD020266


Tgt Ion:252 Resp: 2982158
Ion Ratio Lower Upper
252 100
253 18.8 15.8 23.8
125 10.7 8.6 13.0



#91
Benzo(k)fluoranthene
Concen: 21.873 ng/ul
RT: 23.357 min Scan# 3412
Delta R.T. 0.006 min
Lab File: BN024614.D
Acq: 29 Mar 2023 17:55

Tgt Ion:252 Resp: 2879405
Ion Ratio Lower Upper
252 100
253 21.2 17.5 26.3
125 11.4 9.2 13.8





#93

Benzo(a)pyrene

Concen: 21.308 ng/ul

RT: 23.957 min Scan# 31

Delta R.T. 0.006 min

Lab File: BN024614.D

Acq: 29 Mar 2023 17:55

Instrument :

BNA_N

ClientSampleId :

SSTD020266





Tgt Ion:252 Resp: 2572954

Ion Ratio Lower Upper

252 100

253 20.9 18.2 27.4

125 11.4 10.2 15.4



#94

Indeno(1,2,3-cd)pyrene

Concen: 20.768 ng/ul

RT: 26.674 min Scan# 3976

Delta R.T. 0.029 min

Lab File: BN024614.D

Acq: 29 Mar 2023 17:55




Tgt Ion:276 Resp: 3195718


Ion Ratio Lower Upper

276 100

138 26.0 22.2 33.2

277 23.4 20.3 30.5









#95
 Dibenzo(a,h)anthracene
 Concen: 20.477 ng/ul
 RT: 26.686 min Scan# 3978
 Delta R.T. 0.023 min
 Lab File: BN024614.D
 Acq: 29 Mar 2023 17:55

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020266

Tgt Ion:278 Resp: 2613173
 Ion Ratio Lower Upper
 278 100
 139 17.7 14.2 21.4
 279 23.1 19.6 29.4



#96
 Benzo(g,h,i)perylene
 Concen: 19.394 ng/ul
 RT: 27.474 min Scan# 4112
 Delta R.T. 0.035 min
 Lab File: BN024614.D
 Acq: 29 Mar 2023 17:55

Tgt Ion:276 Resp: 2479551
 Ion Ratio Lower Upper
 276 100
 138 24.0 19.5 29.3
 277 22.5 18.4 27.6

