

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032923\
 Data File : BN024621.D
 Acq On : 29 Mar 2023 22:11
 Operator : CG/JU
 Sample : 02085-01
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 COMBO

Quant Time: Mar 30 03:34:33 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032923.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 30 03:30:09 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.999	152	321303	20.000	ng/u1	0.00
20) Naphthalene-d8	10.816	136	1408893	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.639	164	890995	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.386	188	1907851	20.000	ng/u1	0.00
79) Chrysene-d12	21.568	240	1745374	20.000	ng/u1	0.00
88) Perylene-d12	24.051	264	1840746	20.000	ng/u1	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.369	96	39111	4.719	ng/uL	0.00
4) Pyridine-d5	3.805	84	196520	8.212	ng/u1	0.00
7) Phenol-d5	7.152	99	174820	5.749	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.322	67	584953	29.790	ng/u1	0.00
11) 2-Chlorophenol-d4	7.528	132	499443	22.155	ng/u1	0.00
15) 4-Methylphenol-d8	8.704	113	333269	13.705	ng/u1	0.00
21) Nitrobenzene-d5	9.169	128	340095	30.502	ng/u1	0.00
24) 2-Nitrophenol-d4	9.893	143	345912	28.661	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.434	165	573063	26.143	ng/u1	0.00
31) 4-Chloroaniline-d4	10.951	131	777454	23.244	ng/u1	0.00
46) Dimethylphthalate-d6	14.045	166	2299305	33.558	ng/u1	0.00
49) Acenaphthylene-d8	14.334	160	2465161	31.597	ng/u1	0.00
54) 4-Nitrophenol-d4	14.828	143	74123	5.169	ng/u1	0.00
60) Fluorene-d10	15.628	176	1924513	32.736	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.745	200	343080	27.302	ng/u1	0.00
73) Anthracene-d10	17.486	188	3227147	35.964	ng/u1	0.00
81) Pyrene-d10	19.774	212	3684736	36.512	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.898	264	3526531	37.810	ng/u1	0.00
Target Compounds						
					Qvalue	
12) 2-Chlorophenol	7.557	128	26127	1.111	ng/u1	88
22) Nitrobenzene	9.210	77	1074012	37.083	ng/u1	99
25) 2-Nitrophenol	9.928	139	20773	1.583	ng/u1	99
39) 1,2,4,5-Tetrachloroben...	12.834	216	46649	1.766	ng/u1	98
75) 1,2,3,4-Tetrachloroben...	13.440	216	227427	8.338	ng/uL	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032923\
Data File : BN024621.D
Acq On : 29 Mar 2023 22:11
Operator : CG/JU
Sample : 02085-01
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
COMBO

Quant Time: Mar 30 03:34:33 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032923.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Mar 30 03:30:09 2023
Response via : Initial Calibration

