

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN040119\
 Data File : BN005040.D
 Acq On : 01 Apr 2019 09:58
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 01 19:13:50 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 29 03:42:17 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	123	0.00
2	1,4-Dioxane	0.430	0.419	2.6	123	0.00
3	n-Nitrosodimethylamine	0.262	0.239	8.8	138	0.00
4 S	2-Fluorophenol	0.999	1.060	-6.1	136	0.00
5 S	Phenol-d6	1.183	1.271	-7.4	141	0.00
6	bis(2-Chloroethyl)ether	1.117	1.229	-10.0	139	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	127	0.00
8 S	Nitrobenzene-d5	0.232	0.239	-3.0	140	-0.01
9	Naphthalene	1.006	1.100	-9.3	142	0.00
10	Hexachlorobutadiene	0.187	0.202	-8.0	138	0.00
11 SURR	2-Methylnaphthalene-d10	0.620	0.667	-7.6	140	0.00
12	2-Methylnaphthalene	0.725	0.784	-8.1	140	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	122	0.00
14 S	2,4,6-Tribromophenol	0.224	0.213	4.9	131	0.00
15 S	2-Fluorobiphenyl	1.510	1.666	-10.3	136	0.00
16	Acenaphthylene	1.888	1.978	-4.8	138	0.00
17	Acenaphthene	1.199	1.312	-9.4	138	0.00
18	Fluorene	1.609	1.744	-8.4	136	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	120	0.00
20	4-Bromophenyl-phenylether	0.229	0.240	-4.8	132	0.00
21	Hexachlorobenzene	0.253	0.275	-8.7	136	0.00
22	Pentachlorophenol	0.056	0.035	37.5#	107	0.00
23	Phenanthrene	1.119	1.205	-7.7	134	0.00
24	Anthracene	0.997	1.043	-4.6	135	0.00
25 SURR	Fluoranthene-d10	1.142	1.218	-6.7	135	0.00
26	Fluoranthene	1.371	1.461	-6.6	134	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	122	0.00
28	Pyrene	1.136	1.229	-8.2	135	0.00
29 S	Terphenyl-d14	0.822	0.838	-1.9	128	0.00
30	Benzo(a)anthracene	1.128	1.179	-4.5	136	-0.01
31	Chrysene	1.216	1.304	-7.2	132	-0.01
32	Bis(2-ethylhexyl)phthalate	0.472	0.438	7.2	132	0.00
33	Indeno(1,2,3-cd)pyrene	1.367	1.633	-19.5	148	0.00
34 I	Perylene-d12	1.000	1.000	0.0	128	0.00
35	Benzo(b)fluoranthene	1.282	1.369	-6.8	143	0.00
36	Benzo(k)fluoranthene	1.239	1.311	-5.8	137	0.00
37 C	Benzo(a)pyrene	1.140	1.212	-6.3	142	0.00
38	Dibenzo(a,h)anthracene	1.141	1.254	-9.9	147	0.00
39	Benzo(g,h,i)perylene	1.149	1.299	-13.1	150#	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN040119\
Data File : BN005040.D
Acq On : 01 Apr 2019 09:58
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 01 19:13:50 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 29 03:42:17 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			