

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN040119\
 Data File : BN005051.D
 Acq On : 01 Apr 2019 16:52
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 02 01:37:01 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 29 03:42:17 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
2	1,4-Dioxane	0.430	0.431	-0.2	105	0.00
3	n-Nitrosodimethylamine	0.262	0.208	20.6	99	0.00
4 S	2-Fluorophenol	0.999	1.013	-1.4	108	0.00
5 S	Phenol-d6	1.183	1.210	-2.3	111	0.00
6	bis(2-Chloroethyl)ether	1.117	1.111	0.5	104	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	107	0.00
8 S	Nitrobenzene-d5	0.232	0.223	3.9	110	0.00
9	Naphthalene	1.006	1.087	-8.1	118	0.00
10	Hexachlorobutadiene	0.187	0.206	-10.2	119	0.00
11 SURR	2-Methylnaphthalene-d10	0.620	0.666	-7.4	118	0.00
12	2-Methylnaphthalene	0.725	0.771	-6.3	116	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
14 S	2,4,6-Tribromophenol	0.224	0.205	8.5	107	0.00
15 S	2-Fluorobiphenyl	1.510	1.649	-9.2	113	0.00
16	Acenaphthylene	1.888	1.914	-1.4	112	0.00
17	Acenaphthene	1.199	1.293	-7.8	115	0.00
18	Fluorene	1.609	1.725	-7.2	114	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	100	0.00
20	4-Bromophenyl-phenylether	0.229	0.239	-4.4	109	0.00
21	Hexachlorobenzene	0.253	0.281	-11.1	116	0.00
22	Pentachlorophenol	0.056	0.029	48.2#	74	0.00
23	Phenanthrene	1.119	1.182	-5.6	109	0.00
24	Anthracene	0.997	1.005	-0.8	108	0.00
25 SURR	Fluoranthene-d10	1.142	1.167	-2.2	107	0.00
26	Fluoranthene	1.371	1.384	-0.9	106	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	89	0.00
28	Pyrene	1.136	1.317	-15.9	106	0.00
29 S	Terphenyl-d14	0.822	0.903	-9.9	101	0.00
30	Benzo(a)anthracene	1.128	1.200	-6.4	101	0.00
31	Chrysene	1.216	1.279	-5.2	95	0.00
32	Bis(2-ethylhexyl)phthalate	0.472	0.426	9.7	94	0.00
33	Indeno(1,2,3-cd)pyrene	1.367	1.362	0.4	90	0.01
34 I	Perylene-d12	1.000	1.000	0.0	84	0.00
35	Benzo(b)fluoranthene	1.282	1.373	-7.1	94	0.00
36	Benzo(k)fluoranthene	1.239	1.360	-9.8	93	0.00
37 C	Benzo(a)pyrene	1.140	1.224	-7.4	94	0.00
38	Dibenzo(a,h)anthracene	1.141	1.172	-2.7	90	0.00
39	Benzo(g,h,i)perylene	1.149	1.194	-3.9	90	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN040119\
Data File : BN005051.D
Acq On : 01 Apr 2019 16:52
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 02 01:37:01 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 29 03:42:17 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			