

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN040119\
 Data File : BN005051.D
 Acq On : 01 Apr 2019 16:52
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 02 01:37:01 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 29 03:42:17 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	102	0.00
2	1,4-Dioxane	0.400	0.401	-0.3	105	0.00
3	n-Nitrosodimethylamine	0.400	0.317	20.8	99	0.00
4 S	2-Fluorophenol	0.400	0.406	-1.5	108	0.00
5 S	Phenol-d6	0.400	0.409	-2.2	111	0.00
6	bis(2-Chloroethyl)ether	0.400	0.398	0.5	104	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	107	0.00
8 S	Nitrobenzene-d5	0.400	0.384	4.0	110	0.00
9	Naphthalene	0.400	0.433	-8.2	118	0.00
10	Hexachlorobutadiene	0.400	0.442	-10.5	119	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.430	-7.5	118	0.00
12	2-Methylnaphthalene	0.400	0.425	-6.2	116	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	103	0.00
14 S	2,4,6-Tribromophenol	0.400	0.365	8.8	107	0.00
15 S	2-Fluorobiphenyl	0.400	0.437	-9.2	113	0.00
16	Acenaphthylene	0.400	0.405	-1.3	112	0.00
17	Acenaphthene	0.400	0.431	-7.7	115	0.00
18	Fluorene	0.400	0.429	-7.2	114	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	100	0.00
20	4-Bromophenyl-phenylether	0.400	0.417	-4.2	109	0.00
21	Hexachlorobenzene	0.400	0.445	-11.2	116	0.00
22	Pentachlorophenol	0.400	0.317	20.8	74	0.00
23	Phenanthrene	0.400	0.422	-5.5	109	0.00
24	Anthracene	0.400	0.403	-0.8	108	0.00
25 SURR	Fluoranthene-d10	0.400	0.409	-2.2	107	0.00
26	Fluoranthene	0.400	0.404	-1.0	106	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	89	0.00
28	Pyrene	0.400	0.464	-16.0	106	0.00
29 S	Terphenyl-d14	0.400	0.439	-9.7	101	0.00
30	Benzo(a)anthracene	0.400	0.426	-6.5	101	0.00
31	Chrysene	0.400	0.421	-5.2	95	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.361	9.8	94	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.399	0.3	90	0.01
34 I	Perylene-d12	0.400	0.400	0.0	84	0.00
35	Benzo(b)fluoranthene	0.400	0.428	-7.0	94	0.00
36	Benzo(k)fluoranthene	0.400	0.439	-9.7	93	0.00
37 C	Benzo(a)pyrene	0.400	0.430	-7.5	94	0.00
38	Dibenzo(a,h)anthracene	0.400	0.411	-2.7	90	0.00
39	Benzo(g,h,i)perylene	0.400	0.416	-4.0	90	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN040119\
Data File : BN005051.D
Acq On : 01 Apr 2019 16:52
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 02 01:37:01 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 29 03:42:17 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				