

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN040219\
 Data File : BN005063.D
 Acq On : 02 Apr 2019 16:58
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 03 01:58:08 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 29 03:42:17 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	117	0.00
2	1,4-Dioxane	0.430	0.452	-5.1	126	0.00
3	n-Nitrosodimethylamine	0.262	0.236	9.9	128	0.00
4 S	2-Fluorophenol	0.999	1.052	-5.3	127	0.00
5 S	Phenol-d6	1.183	1.244	-5.2	131	-0.02
6	bis(2-Chloroethyl)ether	1.117	1.195	-7.0	128	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	119	0.00
8 S	Nitrobenzene-d5	0.232	0.231	0.4	127	-0.01
9	Naphthalene	1.006	1.082	-7.6	131	-0.01
10	Hexachlorobutadiene	0.187	0.208	-11.2	134	-0.01
11 SURR	2-Methylnaphthalene-d10	0.620	0.659	-6.3	130	0.00
12	2-Methylnaphthalene	0.725	0.769	-6.1	129	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	113	0.00
14 S	2,4,6-Tribromophenol	0.224	0.219	2.2	125	-0.01
15 S	2-Fluorobiphenyl	1.510	1.670	-10.6	126	0.00
16	Acenaphthylene	1.888	1.942	-2.9	125	0.00
17	Acenaphthene	1.199	1.304	-8.8	127	0.00
18	Fluorene	1.609	1.729	-7.5	125	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	110	0.00
20	4-Bromophenyl-phenylether	0.229	0.236	-3.1	118	0.00
21	Hexachlorobenzene	0.253	0.281	-11.1	128	0.00
22	Pentachlorophenol	0.056	0.045	19.6	124	-0.01
23	Phenanthrene	1.119	1.193	-6.6	122	0.00
24	Anthracene	0.997	1.019	-2.2	121	0.00
25 SURR	Fluoranthene-d10	1.142	1.190	-4.2	121	0.00
26	Fluoranthene	1.371	1.414	-3.1	119	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	103	0.01
28	Pvrene	1.136	1.286	-13.2	119	0.00
29 S	Terphenyl-d14	0.822	0.891	-8.4	115	0.00
30	Benzo(a)anthracene	1.128	1.167	-3.5	114	0.00
31	Chrysene	1.216	1.320	-8.6	113	0.00
32	Bis(2-ethylhexyl)phthalate	0.472	0.440	6.8	112	0.01
33	Indeno(1,2,3-cd)pyrene	1.367	1.361	0.4	104	0.02
34 I	Perylene-d12	1.000	1.000	0.0	97	0.01
35	Benzo(b)fluoranthene	1.282	1.451	-13.2	115	0.01
36	Benzo(k)fluoranthene	1.239	1.377	-11.1	109	0.01
37 C	Benzo(a)pyrene	1.140	1.219	-6.9	108	0.01
38	Dibenzo(a,h)anthracene	1.141	1.170	-2.5	104	0.01
39	Benzo(a,h,i)perylene	1.149	1.190	-3.6	104	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN040219\
Data File : BN005063.D
Acq On : 02 Apr 2019 16:58
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 03 01:58:08 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 29 03:42:17 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			