

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN040219\
 Data File : BN005063.D
 Acq On : 02 Apr 2019 16:58
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 03 01:58:08 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 29 03:42:17 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	117	0.00
2	1,4-Dioxane	0.400	0.420	-5.0	126	0.00
3	n-Nitrosodimethylamine	0.400	0.360	10.0	128	0.00
4 S	2-Fluorophenol	0.400	0.422	-5.5	127	0.00
5 S	Phenol-d6	0.400	0.421	-5.2	131	-0.02
6	bis(2-Chloroethyl)ether	0.400	0.428	-7.0	128	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	119	0.00
8 S	Nitrobenzene-d5	0.400	0.398	0.5	127	-0.01
9	Naphthalene	0.400	0.431	-7.7	131	-0.01
10	Hexachlorobutadiene	0.400	0.445	-11.2	134	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.426	-6.5	130	0.00
12	2-Methylnaphthalene	0.400	0.424	-6.0	129	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	113	0.00
14 S	2,4,6-Tribromophenol	0.400	0.391	2.3	125	-0.01
15 S	2-Fluorobiphenyl	0.400	0.442	-10.5	126	0.00
16	Acenaphthylene	0.400	0.411	-2.7	125	0.00
17	Acenaphthene	0.400	0.435	-8.7	127	0.00
18	Fluorene	0.400	0.430	-7.5	125	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	110	0.00
20	4-Bromophenyl-phenylether	0.400	0.412	-3.0	118	0.00
21	Hexachlorobenzene	0.400	0.444	-11.0	128	0.00
22	Pentachlorophenol	0.400	0.429	-7.2	124	-0.01
23	Phenanthrene	0.400	0.426	-6.5	122	0.00
24	Anthracene	0.400	0.409	-2.2	121	0.00
25 SURR	Fluoranthene-d10	0.400	0.417	-4.2	121	0.00
26	Fluoranthene	0.400	0.412	-3.0	119	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	103	0.01
28	Pvrene	0.400	0.453	-13.2	119	0.00
29 S	Terphenyl-d14	0.400	0.434	-8.5	115	0.00
30	Benzo(a)anthracene	0.400	0.414	-3.5	114	0.00
31	Chrysene	0.400	0.434	-8.5	113	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.373	6.8	112	0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.398	0.5	104	0.02
34 I	Perylene-d12	0.400	0.400	0.0	97	0.01
35	Benzo(b)fluoranthene	0.400	0.453	-13.2	115	0.01
36	Benzo(k)fluoranthene	0.400	0.444	-11.0	109	0.01
37 C	Benzo(a)pyrene	0.400	0.427	-6.7	108	0.01
38	Dibenzo(a,h)anthracene	0.400	0.410	-2.5	104	0.01
39	Benzo(a,h,i)perylene	0.400	0.414	-3.5	104	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN040219\
Data File : BN005063.D
Acq On : 02 Apr 2019 16:58
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 03 01:58:08 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 29 03:42:17 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				