


Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN040320\
 Data File : BN010401.D
 Acq On : 03 Apr 2020 08:16
 Operator : CG/JU
 Sample : DFTPP62
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP62

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN040120MA.M
 Title : SVOA CALIBRATION
 Last Update : Fri Apr 03 10:35:52 2020



AutoFind: Scans 2493, 2494, 2495; Background Corrected with Scan 2487

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	18.4	218965	PASS
68	69	0.00	2	1.1	4497	PASS
69	198	0.00	100	33.6	400618	PASS
70	69	0.00	2	0.7	2804	PASS
127	198	10	80	42.3	504192	PASS
197	198	0.00	2	0.4	5234	PASS
198	198	100	100	100.0	1192597	PASS
199	198	5	9	6.7	80272	PASS
275	198	10	60	27.4	326826	PASS
365	198	1	100	4.0	48026	PASS
441	198	0.01	100	11.9	141824	PASS
442	198	50	100	79.7	950293	PASS
443	442	15	24	18.9	179776	PASS

m/z	Abundance
36.10	348.0
37.00	708.0
38.00	2447.0
39.00	11951.0
39.90	1344.0
41.10	640.0
42.00	388.0
43.90	1455.0
44.90	775.0
49.10	966.0
50.00	37144.0
51.00	140480.0
52.00	6370.0
53.00	504.0
55.00	1631.0
56.00	5542.0
57.00	10472.0
57.80	395.0
61.00	2360.0
62.00	2503.0
63.00	8502.0
63.90	835.0
65.00	4174.0
67.20	612.0
68.10	5165.0
68.90	267584.0
70.00	1609.0
73.00	3962.0
74.00	30840.0
75.00	42080.0
76.00	16960.0
77.00	299072.0
78.00	20984.0
79.00	22728.0
80.00	16648.0
81.00	22912.0
82.00	6031.0
83.10	4761.0
84.10	468.0
85.00	4266.0
86.00	6487.0
87.00	2394.0
87.90	1772.0
89.10	604.0
91.00	4753.0
92.00	6449.0
93.00	42088.0
93.90	2382.0
95.00	730.0
96.10	1982.0
97.00	856.0
98.00	33992.0
99.00	26376.0
100.00	2486.0
101.00	14210.0
102.00	697.0
102.90	4587.0
104.00	8730.0
105.00	9373.0
106.00	2981.0
107.00	95448.0
108.00	14430.0
109.00	3450.0
110.00	186496.0
111.00	29784.0
112.00	4258.0
113.10	1381.0
113.80	508.0
114.90	984.0
116.00	6702.0
117.00	99544.0
118.00	6500.0
119.00	809.0
120.10	1353.0
121.00	773.0
122.00	6684.0
123.00	12021.0
123.90	5925.0
125.00	5386.0
127.00	357568.0
128.00	25968.0
129.00	154176.0
130.00	12178.0
131.00	3436.0
132.10	1306.0
133.00	642.0
134.00	4590.0
135.00	11239.0
135.90	4722.0
136.90	7217.0
138.10	1571.0
139.10	907.0

Instrument :
BNA_N
ClientSampleId :
DFTPP62

140.10	1470.0
141.00	20168.0
142.00	6468.0
143.00	4910.0
143.90	939.0
144.90	1646.0
145.90	3937.0
147.00	10832.0
148.00	26392.0
149.00	4665.0
149.90	1119.0
151.20	2739.0
152.00	1615.0
153.00	7600.0
153.90	4861.0
155.00	11533.0
156.00	16160.0
156.90	4316.0
158.00	4271.0
159.00	2805.0
160.00	7735.0
161.00	11203.0
161.90	2915.0
163.10	1135.0
164.20	1338.0
165.00	9104.0
166.00	7040.0
167.00	50040.0
168.00	27312.0
169.10	4403.0
170.10	1904.0
170.90	1837.0
172.00	3076.0
173.00	5241.0
174.00	8291.0
175.00	17800.0
176.00	4786.0
177.00	9310.0
177.90	3367.0
178.90	33016.0
180.00	23776.0
181.00	10947.0
182.00	1831.0
183.20	1435.0
184.00	2569.0
185.00	15839.0
186.00	131040.0
187.00	36128.0
188.00	3925.0
189.00	7903.0
189.90	1042.0
191.10	4859.0
192.00	10603.0
193.00	12418.0
194.00	2808.0
194.90	1511.0
196.00	25176.0
197.10	7092.0
198.00	957376.0
199.00	61024.0
199.90	6076.0
201.40	4253.0
203.00	8466.0
204.00	38016.0
205.00	60160.0
206.00	260544.0
207.00	35176.0
208.00	9444.0
209.00	3011.0
210.10	5229.0
211.00	11383.0
211.90	2242.0
212.90	792.0
215.10	3841.0
216.00	6484.0
217.00	73296.0
218.00	9650.0
219.10	938.0
221.10	36296.0
221.90	8819.0
223.00	17408.0
224.00	149312.0
225.00	34888.0
226.10	5668.0
227.00	63312.0
228.00	10155.0
229.00	11307.0
229.90	1309.0
231.10	5065.0
232.10	877.0
233.00	2547.0
234.00	4467.0
235.00	4077.0
235.90	3616.0
237.00	6377.0
237.90	1112.0

Instrument :
BNA_N
ClientSampleId :
DFTPP62

239.00	2319.0
240.00	1712.0
241.00	4014.0
242.00	8236.0
243.00	8679.0
244.00	112072.0
245.10	15392.0
246.00	22352.0
247.00	5130.0
247.90	806.0
249.00	4342.0
250.00	434.0
251.00	1068.0
252.10	1090.0
253.00	3408.0
255.00	631616.0
256.00	90952.0
257.00	7432.0
258.00	34888.0
259.00	4646.0
259.90	770.0
261.00	1357.0
264.10	1134.0
265.00	15828.0
266.00	2346.0
267.00	595.0
269.90	923.0
271.10	2255.0
272.10	2281.0
273.00	18040.0
274.00	50736.0
275.00	276928.0
276.00	38024.0
277.00	23528.0
278.00	4416.0
278.90	1087.0
281.10	1562.0
282.20	888.0
283.00	3617.0
284.10	2167.0
285.00	3863.0
286.10	890.0
287.70	440.0
289.00	1022.0
289.80	643.0
290.20	857.0
291.00	664.0
292.00	1162.0
292.90	5244.0
294.00	1656.0
295.00	1930.0
296.00	91984.0
297.00	12227.0
298.00	805.0
301.00	1123.0
301.90	1505.0
303.00	9487.0
304.10	3024.0
305.10	557.0
308.00	1598.0
309.00	911.0
310.00	1609.0
311.80	416.0
313.00	913.0
314.00	5340.0
315.00	10532.0
316.00	6819.0
316.90	728.0
319.00	321.0
320.00	433.0
320.90	2891.0
322.20	1991.0
323.00	27096.0
324.00	4339.0
325.10	813.0
325.80	498.0
327.00	5910.0
328.00	2686.0
329.00	689.0
332.10	2501.0
333.10	2719.0
334.00	19696.0
335.00	4266.0
336.10	808.0
339.10	415.0
340.20	404.0
341.00	3508.0
342.20	1218.0
346.00	5919.0
347.00	1071.0
347.90	309.0
351.10	352.0
352.00	10493.0
353.10	5603.0
354.00	10455.0
355.00	2809.0

Instrument :
BNA_N
ClientSampleId :
DFTPP62

357.20	306.0
358.80	438.0
359.10	506.0
361.10	431.0
365.00	45712.0
366.00	6260.0
370.10	764.0
370.90	1910.0
372.00	16480.0
373.10	4815.0
373.80	574.0
377.10	403.0
378.00	344.0
383.00	5236.0
384.00	1338.0
385.10	338.0
390.00	1899.0
390.90	1197.0
392.00	1160.0
400.90	1095.0
402.00	7355.0
403.00	9290.0
404.00	4129.0
404.90	893.0
410.00	487.0
414.70	341.0
421.00	9182.0
422.00	7376.0
423.00	63472.0
424.00	14237.0
425.00	2429.0
425.90	413.0
439.30	444.0
441.10	135936.0
442.10	945856.0
443.10	191936.0
444.00	17648.0
445.00	808.0
445.30	653.0

Instrument :
BNA_N
ClientSampleId :
DFTPP62

m/z	Abundance
35.90	496.0
36.90	1402.0
38.00	3686.0
39.00	23448.0
39.90	1731.0
41.00	887.0
43.10	600.0
44.00	1113.0
44.90	595.0
48.40	344.0
49.10	2060.0
50.00	71560.0
51.00	258624.0
52.00	12799.0
53.00	657.0
55.00	1861.0
55.90	9056.0
57.00	19976.0
58.00	1150.0
61.00	4898.0
62.00	5003.0
63.00	14575.0
63.90	2120.0
65.00	7837.0
65.90	557.0
67.10	1266.0
68.10	9287.0
69.00	478272.0
69.90	3559.0
71.00	642.0
73.00	6961.0
74.00	51752.0
75.00	73656.0
76.00	27696.0
77.00	551360.0
78.00	37480.0
79.00	38872.0
80.00	30064.0
81.00	39288.0
82.00	10560.0
83.00	7186.0
84.10	1492.0
85.00	7728.0
86.00	15013.0
87.00	5094.0
88.00	1368.0
89.10	1217.0
90.10	550.0
91.00	8217.0
92.00	10477.0
93.00	67208.0
94.00	4464.0
94.90	1287.0
96.10	2953.0
96.80	1593.0
98.00	58920.0
99.00	41568.0
100.00	3741.0
101.00	22752.0
102.10	1321.0
103.00	6808.0
104.00	15010.0
104.90	15340.0
106.00	4757.0
107.00	180224.0
108.00	25976.0
109.00	6425.0
110.00	324608.0
111.00	45312.0
112.00	6163.0
112.90	2805.0
114.70	1043.0
116.10	9795.0
117.00	166656.0
118.00	11387.0
119.00	1882.0
120.00	2351.0
122.00	12713.0
123.00	18304.0
124.00	7630.0
125.00	7890.0
127.00	620032.0
128.00	46712.0
129.00	241664.0
130.00	20728.0
131.00	4740.0
132.00	2165.0
133.30	1155.0
134.00	6822.0
135.00	18192.0
136.00	8579.0
137.10	11723.0

Instrument :
BNA_N
ClientSampleId :
DFTPP62

138.10	2459.0
139.00	2361.0
140.10	3026.0
141.00	34840.0
142.00	10154.0
143.00	6582.0
144.00	1807.0
145.10	1570.0
145.90	6498.0
147.00	17544.0
148.00	42176.0
149.00	8900.0
150.00	2196.0
151.00	4001.0
151.90	2896.0
153.00	9846.0
154.00	9657.0
155.00	17712.0
156.00	27128.0
157.00	5263.0
157.90	7174.0
159.00	4158.0
160.00	12362.0
161.00	16472.0
161.90	4516.0
163.00	1651.0
164.00	1409.0
165.00	13420.0
166.00	10971.0
167.00	78560.0
168.00	42768.0
169.10	5561.0
170.00	2344.0
170.90	3411.0
172.00	6825.0
173.00	7882.0
174.00	17048.0
175.10	29952.0
176.00	7324.0
177.00	14833.0
178.10	4419.0
179.00	56408.0
180.00	37328.0
181.00	16480.0
182.00	2841.0
183.10	972.0
184.00	6339.0
185.10	22712.0
186.00	211264.0
187.00	59744.0
188.00	7463.0
189.00	13328.0
189.90	2086.0
191.00	6638.0
192.00	18112.0
193.00	21320.0
194.00	4939.0
194.90	3727.0
196.00	41088.0
198.00	1449984.0
199.00	96624.0
200.00	8238.0
201.40	5303.0
203.00	12579.0
204.10	59272.0
205.00	95984.0
206.00	400384.0
207.00	58000.0
208.00	14204.0
209.00	4225.0
210.00	5561.0
211.00	14789.0
213.00	1494.0
213.90	567.0
215.00	5283.0
216.00	9618.0
217.00	111104.0
218.00	14128.0
219.20	1735.0
221.00	52160.0
221.70	15647.0
223.00	24776.0
224.00	234624.0
225.10	54496.0
226.10	6850.0
227.00	100016.0
228.00	15411.0
229.00	20352.0
229.90	2469.0
231.00	7845.0
232.10	1726.0
233.00	1803.0
234.00	5923.0
235.00	6933.0
235.90	4371.0
237.00	9273.0

Instrument :
BNA_N
ClientSampleId :
DFTPP62

238.10	1358.0
239.00	3037.0
239.90	2057.0
241.00	4185.0
242.00	12396.0
243.10	13314.0
244.00	164736.0
245.00	23056.0
246.00	38688.0
247.00	6700.0
248.00	1934.0
249.00	6582.0
249.90	1460.0
250.90	1709.0
252.00	2190.0
253.10	4303.0
255.00	936768.0
256.00	135488.0
257.10	11887.0
258.00	54584.0
259.00	7731.0
259.90	1803.0
261.00	1493.0
263.10	497.0
263.80	1287.0
265.00	21704.0
266.00	3295.0
267.10	639.0
268.30	312.0
269.90	939.0
271.00	2487.0
272.00	2835.0
273.00	27056.0
274.00	72656.0
275.00	408896.0
276.00	52120.0
277.00	36704.0
278.00	6970.0
278.80	1258.0
281.00	1737.0
282.10	1184.0
283.10	3985.0
284.00	2078.0
285.00	6560.0
286.10	1520.0
287.90	769.0
288.90	1837.0
290.10	1472.0
291.10	1325.0
292.00	1664.0
293.00	8346.0
294.00	2967.0
295.10	2904.0
296.00	123032.0
297.00	17400.0
297.90	748.0
301.00	2085.0
302.00	2574.0
303.00	14504.0
304.00	4204.0
305.00	651.0
308.00	1722.0
309.00	1143.0
310.00	2444.0
311.00	396.0
312.00	510.0
313.10	1094.0
314.00	6260.0
315.00	14010.0
316.00	7013.0
317.00	994.0
318.80	444.0
320.00	1026.0
321.00	4161.0
322.10	2402.0
323.00	36528.0
324.00	6314.0
324.90	483.0
325.90	735.0
326.90	7633.0
328.00	2973.0
329.20	702.0
332.10	3052.0
333.00	3707.0
334.00	26344.0
335.00	7390.0
336.00	420.0
339.00	563.0
340.20	650.0
341.00	4841.0
342.10	1415.0
342.80	406.0
346.00	7127.0
346.80	1533.0
347.20	1533.0
351.10	667.0

Instrument :
BNA_N
ClientSampleId :
DFTPP62

352.00	11139.0
353.00	7725.0
354.00	12076.0
355.00	2422.0
357.10	435.0
359.00	1197.0
363.40	430.0
365.00	60536.0
366.00	8602.0
367.10	852.0
369.90	1302.0
370.90	3184.0
372.00	21248.0
373.00	4973.0
374.00	785.0
377.00	393.0
383.00	5602.0
384.00	2007.0
385.00	702.0
390.00	3168.0
391.00	1521.0
392.00	1539.0
396.10	502.0
400.80	1292.0
402.00	8197.0
403.00	13023.0
404.00	5121.0
405.00	859.0
415.10	460.0
421.00	10388.0
422.00	10110.0
423.00	72760.0
424.10	17360.0
425.00	2945.0
439.00	974.0
439.60	369.0
441.00	181440.0
442.00	1186816.0
443.00	219840.0
444.10	19576.0
444.90	1270.0

Instrument :
BNA_N
ClientSampleId :
DFTPP62

m/z	Abundance
36.00	359.0
37.00	1037.0
37.90	3461.0
39.00	21608.0
40.00	2026.0
41.10	1099.0
44.00	1883.0
45.00	592.0
49.00	2677.0
50.00	67224.0
51.00	257792.0
52.00	12726.0
52.90	468.0
55.10	2538.0
56.00	8750.0
57.00	17568.0
58.00	1075.0
59.00	521.0
61.00	4018.0
62.00	5154.0
63.00	14757.0
64.00	2246.0
65.00	7424.0
66.00	470.0
67.00	1128.0
69.00	456000.0
70.00	3246.0
73.00	7136.0
74.00	52912.0
75.00	74920.0
76.00	26576.0
77.00	498304.0
78.00	37360.0
79.00	35792.0
80.00	28952.0
81.00	36336.0
82.00	10282.0
83.10	8079.0
84.00	1042.0
85.00	7298.0
86.00	11237.0
87.00	5432.0
88.00	2293.0
88.90	836.0
90.10	499.0
91.00	8924.0
92.00	10970.0
93.00	62464.0
93.90	4465.0
95.00	827.0
96.10	2760.0
97.10	1373.0
98.00	52632.0
99.00	42264.0
100.00	3639.0
101.00	21216.0
101.80	1032.0
103.00	7246.0
103.90	15043.0
105.00	14944.0
106.00	5017.0
107.00	160000.0
108.00	22096.0
109.10	5171.0
110.00	289536.0
111.00	44064.0
111.90	4782.0
113.00	1450.0
113.60	478.0
114.20	561.0
114.70	548.0
115.10	952.0
116.00	10345.0
117.00	147712.0
118.00	11687.0
118.90	1153.0
120.00	1754.0
121.00	977.0
121.90	10977.0
123.00	15714.0
124.00	8254.0
125.00	8361.0
127.00	534976.0
128.00	50000.0
129.00	208768.0
130.00	21768.0
131.00	3915.0
132.00	1709.0
133.10	1193.0
134.00	7144.0
135.00	18088.0
136.00	7642.0

Instrument :
BNA_N
ClientSampleId :
DFTPP62

137.10	9162.0
138.00	2738.0
138.90	1636.0
140.00	3279.0
141.00	32040.0
141.90	10331.0
143.00	7146.0
144.10	1442.0
145.00	1958.0
145.90	6038.0
147.00	15447.0
148.00	37064.0
149.00	6933.0
150.00	1972.0
150.90	2996.0
151.90	1988.0
153.00	9572.0
154.00	6438.0
155.00	14572.0
156.00	24904.0
157.00	4536.0
158.00	3841.0
159.00	4779.0
160.00	9752.0
161.00	14716.0
162.00	3832.0
163.10	1033.0
163.90	1198.0
165.00	12205.0
166.00	9068.0
167.00	61112.0
168.00	38800.0
169.00	6235.0
170.00	2443.0
170.80	3110.0
172.00	5452.0
173.00	7021.0
174.00	13154.0
175.10	23320.0
176.10	5630.0
177.00	10136.0
178.10	4544.0
179.00	46832.0
180.00	33240.0
181.00	12002.0
182.10	2357.0
183.00	960.0
184.00	4039.0
185.00	22368.0
186.00	170752.0
187.00	45048.0
188.00	6101.0
189.00	10601.0
190.20	1603.0
191.10	4895.0
192.00	15454.0
193.00	15944.0
194.10	3663.0
195.20	1672.0
196.00	34576.0
197.00	8611.0
198.00	1170432.0
199.00	83168.0
200.00	5179.0
201.30	4194.0
201.80	3405.0
203.00	9059.0
204.00	45472.0
205.00	74424.0
206.00	319488.0
207.00	46976.0
208.10	12659.0
209.00	4558.0
210.10	5330.0
211.00	11355.0
212.20	2382.0
212.90	874.0
214.00	624.0
215.00	3770.0
216.00	6324.0
217.00	85208.0
218.00	11187.0
218.80	879.0
221.00	48952.0
221.90	11175.0
223.00	21200.0
224.10	179392.0
225.00	46008.0
226.00	5799.0
227.00	83024.0
228.00	11528.0
229.00	16400.0
230.00	1921.0
230.90	5024.0
231.90	1042.0
233.00	1834.0

Instrument :
BNA_N
ClientSampleId :
DFTPP62

234.10	4157.0
235.00	4936.0
236.00	4293.0
237.00	9615.0
238.00	1343.0
239.00	2378.0
240.00	3249.0
241.10	4309.0
242.00	8244.0
243.00	9421.0
244.00	137088.0
245.10	17480.0
246.00	30344.0
246.90	6166.0
248.00	1349.0
249.00	5917.0
250.00	1008.0
250.80	1329.0
252.10	750.0
253.00	2844.0
255.00	715456.0
256.00	108384.0
257.00	11263.0
258.00	39328.0
259.00	6365.0
260.00	1178.0
260.80	1034.0
262.90	420.0
264.20	1139.0
265.00	14940.0
265.90	2846.0
267.00	780.0
268.00	467.0
270.10	858.0
271.00	1853.0
272.10	1697.0
273.10	20288.0
274.00	53456.0
275.00	294656.0
276.10	40872.0
277.00	30256.0
278.00	5819.0
278.80	1104.0
281.00	995.0
282.10	810.0
283.00	3437.0
283.90	1622.0
285.00	5335.0
286.00	991.0
288.00	371.0
288.90	1030.0
289.90	995.0
290.90	644.0
291.90	1108.0
293.00	6980.0
294.00	2713.0
295.10	3276.0
295.90	81304.0
297.00	11384.0
298.10	1601.0
301.00	1110.0
302.00	1184.0
303.10	9198.0
304.00	3506.0
305.10	644.0
308.00	1293.0
309.00	1128.0
310.00	1197.0
310.80	469.0
312.80	716.0
314.00	4746.0
314.90	10197.0
316.00	6593.0
317.10	1330.0
319.70	345.0
321.10	2309.0
322.00	1557.0
323.00	25168.0
324.10	4273.0
325.20	708.0
325.90	741.0
327.00	4579.0
327.90	2840.0
328.90	466.0
332.00	1938.0
332.80	2469.0
334.00	17536.0
335.00	4839.0
336.20	719.0
339.20	512.0
340.10	342.0
341.10	3606.0
342.00	1004.0
346.00	5177.0
347.00	1420.0
350.70	335.0

Instrument :
BNA_N
ClientSampleId :
DFTPP62

352.00	8640.0
353.00	6479.0
354.00	10141.0
355.00	1156.0
358.90	625.0
365.00	37832.0
365.90	4683.0
366.90	894.0
369.70	423.0
370.10	596.0
371.00	1687.0
372.00	14929.0
373.00	3539.0
374.10	727.0
377.00	355.0
383.00	4016.0
384.00	1146.0
384.80	470.0
390.00	1490.0
390.90	1027.0
392.10	607.0
400.80	818.0
401.90	5630.0
403.00	8099.0
404.10	2636.0
404.90	980.0
414.90	656.0
421.00	7454.0
422.00	7016.0
423.00	45328.0
424.00	14808.0
425.00	2413.0
441.00	108096.0
442.10	718208.0
443.00	127552.0
444.10	12623.0
445.00	1092.0

Instrument :
BNA_N
ClientSampleId :
DFTPP62

m/z	Abundance
40.00	1438.0
40.90	371.0
43.00	334.0
43.90	943.0
49.00	332.0
54.80	401.0
68.10	320.0
73.00	1210.0
77.00	696.0
78.80	310.0
82.40	376.0
90.90	754.0
104.90	606.0
133.00	710.0
135.10	337.0
147.10	653.0
191.10	710.0
192.90	575.0
207.10	3601.0
208.00	950.0
209.10	623.0
221.90	396.0
266.90	591.0
281.10	1208.0
354.90	368.0

Instrument :
BNA_N
ClientSampleId :
DFTPP62