

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040424\
 Data File : BN030624.D
 Acq On : 03 Apr 2024 12:12
 Operator : MA/JU
 Sample : P1934-05
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0R40

Quant Time: Apr 03 12:49:55 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032824.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Mar 31 12:24:07 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.693	152	247233	20.000	ng/u1	0.00
20) Naphthalene-d8	10.475	136	1092625	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.334	164	720393	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.086	188	1509914	20.000	ng/u1	0.00
79) Chrysene-d12	21.321	240	1377589	20.000	ng/u1	0.00
88) Perylene-d12	24.268	264	1340651	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.193	96	26794	5.081	ng/uL	0.00
4) Pyridine-d5	3.605	84	93885	5.938	ng/u1	0.00
7) Phenol-d5	6.857	99	118777	5.939	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.034	67	413125	33.444	ng/u1	0.00
11) 2-Chlorophenol-d4	7.222	132	389014	25.640	ng/u1	0.00
15) 4-Methylphenol-d8	8.393	113	238740	14.243	ng/u1	0.00
21) Nitrobenzene-d5	8.846	128	267142	35.246	ng/u1	0.00
24) 2-Nitrophenol-d4	9.563	143	225268	32.488	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.093	165	438649	28.897	ng/u1	0.00
31) 4-Chloroaniline-d4	10.616	131	420749	16.836	ng/u1	0.00
46) Dimethylphthalate-d6	13.751	166	1737541	36.125	ng/u1	0.00
49) Acenaphthylene-d8	14.028	160	1954815	35.349	ng/u1	0.00
54) 4-Nitrophenol-d4	14.534	143	36208	3.936	ng/u1	0.00
60) Fluorene-d10	15.334	176	1532293	36.406	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.445	200	179830	26.983	ng/u1	0.00
73) Anthracene-d10	17.186	188	2689560	41.342	ng/u1	0.00
81) Pyrene-d10	19.486	212	3116364	42.076	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.068	264	2749493	44.509	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040424\
Data File : BN030624.D
Acq On : 03 Apr 2024 12:12
Operator : MA/JU
Sample : P1934-05
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0R40

Quant Time: Apr 03 12:49:55 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032824.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sun Mar 31 12:24:07 2024
Response via : Initial Calibration

